

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB92575

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP60917
calibration std. phosphate 0.5 ppm	WP60915
calibration std. phosphate 0.3 ppm	WP60913
calibration std. 0 ppm	WP60908
phosphate ICV std.	WP60922
phosphate CCV std.	WP60919
calibration std. phosphate 0.1 ppm	WP60912
calibration std. phosphate 0.05 ppm	WP60909
Combined reagent	WP60927
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036
Ascorbic Acid	WP60924

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/12/2018	11:09
2	CAL2	0.05	1	50	50	0.033	01/12/2018	11:09
3	CAL3	0.10	1	50	50	0.065	01/12/2018	11:10
4	CAL4	0.30	1	50	50	0.186	01/12/2018	11:10
5	CAL5	0.50	1	50	50	0.330	01/12/2018	11:11
6	CAL6	1.00	1	50	50	0.649	01/12/2018	11:11



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB92575

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.319	0.492	01/12/2018	11:12
2	ICB		1	50	50	0.000	0.001	01/12/2018	11:12
3	CCV1	0.50	1	50	50	0.327	0.504	01/12/2018	11:13
4	CCB1		1	50	50	0.000	0.001	01/12/2018	11:13
5	PB105602BL		1	50	50	0.000	0.001	01/12/2018	11:14
6	PB105602BS	0.50	1	50	50	0.291	0.449	01/12/2018	11:14
7	J1054-03		1	50	50	1.106	1.702	01/12/2018	11:15
8	J1054-03DUP		1	50	50	1.226	1.886	01/12/2018	11:15
9	J1054-03MS	0.50	1	50	50	1.239	1.906	01/12/2018	11:16
10	J1054-03MSD	0.50	1	50	50	1.231	1.894	01/12/2018	11:16
11	J1054-03		10	50	50	0.299	0.461	01/12/2018	11:17
12	J1054-03DUP		10	50	50	0.343	0.529	01/12/2018	11:17
13	J1054-03MS	0.50	10	50	50	0.334	0.515	01/12/2018	11:30
14	J1054-03MSD	0.50	10	50	50	0.330	0.509	01/12/2018	11:30
15	CCV2	0.50	1	50	50	0.328	0.506	01/12/2018	11:31
16	CCB2		1	50	50	0.000	0.001	01/12/2018	11:31