

7470A


LB92577

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35717



A= 0.0000e+000

B= 2.8133e-004

C= 3.3529e-002

Rho= 0.9998948

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.048	0.048	51	0.000	51				
0.2	0.200	0.208	0.008	620	0.0 %	620				
2.5	2.500	2.479	-0.021	8693	0.0 %	8693				
5.0	5.000	4.963	-0.037	17524	0.0 %	17524				
7.5	7.500	7.420	-0.080	26256	0.0 %	26256				
10.0	10.000	10.082	0.082	35717	0.0 %	35717				

LB92577 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jan 2018 10:42:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.0000	7470A	PPB	12 Jan 2018 10:47:46	S	Std
0.2 - 1	50.2	620	-	0.2000	7470A	PPB	12 Jan 2018 10:49:54	S	Std
2.5 - 1	52.5	8693	-	2.5000	7470A	PPB	12 Jan 2018 10:52:02	S	Std
5.0 - 1	55	17524	-	5.0000	7470A	PPB	12 Jan 2018 10:54:10	S	Std
7.5 - 1	57.5	26256	-	7.5000	7470A	PPB	12 Jan 2018 10:56:19	S	Std
10.0 - 1	510	35717	-	10.0000	7470A	PPB	12 Jan 2018 11:00:56	S	Std
ICV44 - 1	ICV44	13965	3.9622	-	7470A	PPB	12 Jan 2018 11:03:56	U	SMPL
ICB44 - 1	ICB44	-67	0.0147	-	7470A	PPB	12 Jan 2018 11:06:02	U	SMPL
CCV39 - 1	CCV39	17720	5.0186	-	7470A	PPB	12 Jan 2018 11:08:09	U	SMPL
CCB39 - 1	CCB39	-74	0.0127	-	7470A	PPB	12 Jan 2018 11:10:15	U	SMPL
CRA - 1	CRA	573	0.1947	-	7470A	PPB	12 Jan 2018 11:12:23	U	SMPL
HighStd - 1	HighStd	35169	9.9274	-	7470A	PPB	12 Jan 2018 11:14:29	U	SMPL
ChkStd - 1	ChkStd	24946	7.0515	-	7470A	PPB	12 Jan 2018 11:16:36	U	SMPL
PB105587BL - 1	PBW	-164	-0.0126	-	7470A	PPB	12 Jan 2018 11:18:45	U	SMPL
PB105587BS - 1	LCSW	13091	3.7164	-	7470A	PPB	12 Jan 2018 11:20:53	U	SMPL
PB105587TB - 1	PB105564TB	-69	0.0141	-	7470A	PPB	12 Jan 2018 11:23:01	U	SMPL
J1078-05 - 1	PB-6-OW(8-32)	18	0.0386	-	7470A	PPB	12 Jan 2018 11:25:09	U	SMPL
J1078-05DUP - 1	PB-6-OW(8-32)DUP	9	0.0361	-	7470A	PPB	12 Jan 2018 11:27:17	U	SMPL
J1078-05MS - 1	PB-6-OW(8-32)MS	13505	3.8328	-	7470A	PPB	12 Jan 2018 11:29:26	U	SMPL
J1078-05MSD - 1	PB-6-OW(8-32)MSD	13965	3.9622	-	7470A	PPB	12 Jan 2018 11:31:33	U	SMPL
CCV40 - 1	CCV40	17792	5.0389	-	7470A	PPB	12 Jan 2018 11:33:39	U	SMPL
CCB40 - 1	CCB40	-75	0.0124	-	7470A	PPB	12 Jan 2018 11:35:46	U	SMPL
J1088-04 - 1	TP-COMP	11	0.0366	-	7470A	PPB	12 Jan 2018 11:37:53	U	SMPL
PB105588BL - 1	PBW	28	0.0414	-	7470A	PPB	12 Jan 2018 11:39:59	U	SMPL
PB105588BS - 1	LCSW	13390	3.8005	-	7470A	PPB	12 Jan 2018 11:42:06	U	SMPL
J1082-03 - 1	FB-4	-144	-0.0070	-	7470A	PPB	12 Jan 2018 11:44:12	U	SMPL
J1078-05LX5 - 1		17	0.0383	-	7470A	PPB	12 Jan 2018 11:46:20	U	SMPL
J1078-05A - 1		14063	3.9898	-	7470A	PPB	12 Jan 2018 11:48:28	U	SMPL
CCV41 - 1	CCV41	17731	5.0217	-	7470A	PPB	12 Jan 2018 11:50:35	U	SMPL
CCB41 - 1	CCB41	-73	0.0130	-	7470A	PPB	12 Jan 2018 11:52:43	U	SMPL