


7471B

LB92589
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42800



A= 0.0000e+000

B= 2.3173e-004

C= 1.6657e-002

Rho= 0.9999145

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	-22	0.000	-22				
0.2	0.200	0.170	-0.030	663	0.0 %	663	0			
2.5	2.500	2.463	-0.037	10556	0.0 %	10556				
5.0	5.000	5.067	0.067	21793	0.0 %	21793				
7.5	7.500	7.554	0.054	32524	0.0 %	32524				
10.0	10.000	9.935	-0.065	42800	0.0 %	42800				

LB92589 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Jan 2018 14:31:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-22	-	0.00007471B	PPB		12 Jan 2018 14:56:29	S	Std
0.2 - 1	50.2	663	-	0.20007471B	PPB		12 Jan 2018 14:58:46	S	Std
2.5 - 1	52.5	10556	-	2.50007471B	PPB		12 Jan 2018 15:01:04	S	Std
5.0 - 1	55	21793	-	5.00007471B	PPB		12 Jan 2018 15:03:22	S	Std
7.5 - 1	57.5	32524	-	7.50007471B	PPB		12 Jan 2018 15:05:39	S	Std
10.0 - 1	510	42800	-	10.00007471B	PPB		12 Jan 2018 15:07:56	S	Std
ICV45 - 1	ICV45	16724	3.8922	-7471B	PPB		12 Jan 2018 15:10:37	U	SMPL
ICB45 - 1	ICB45	-162	-0.0209	-7471B	PPB		12 Jan 2018 15:12:52	U	SMPL
CCV42 - 1	CCV42	21765	5.0604	-7471B	PPB		12 Jan 2018 15:15:08	U	SMPL
CCB42 - 1	CCB42	-161	-0.0207	-7471B	PPB		12 Jan 2018 15:17:24	U	SMPL
CRA - 1	CRA	647	0.1666	-7471B	PPB		12 Jan 2018 15:19:40	U	SMPL
HighStd - 1	HighStd	42537	9.8739	-7471B	PPB		12 Jan 2018 15:21:56	U	SMPL
ChkStd - 1	ChkStd	28980	6.7323	-7471B	PPB		12 Jan 2018 15:24:12	U	SMPL
PB105597BL - 1	PBS	-208	-0.0315	-7471B	PPB		12 Jan 2018 15:26:31	U	SMPL
PB105597BS - 1	LCSS	16357	3.8071	-7471B	PPB		12 Jan 2018 15:28:48	U	SMPL
J1010-01 - 1	BU-02-011118	2097	0.5026	-7471B	PPB		12 Jan 2018 15:31:06	U	SMPL
J1011-10 - 1	HR-05-011118-A	21799	5.0682	-7471B	PPB		12 Jan 2018 15:33:23	U	SMPL
J1011-12 - 1	HR-05-011118-B	1238	0.3035	-7471B	PPB		12 Jan 2018 15:35:40	U	SMPL
J1011-14 - 1	HR-05-011118-C	4458	1.0497	-7471B	PPB		12 Jan 2018 15:37:58	U	SMPL
J1015-01 - 1	CL-02-011118-A	2394	0.5714	-7471B	PPB		12 Jan 2018 15:40:14	U	SMPL
CCV43 - 1	CCV43	21724	5.0509	-7471B	PPB		12 Jan 2018 15:42:30	U	SMPL
CCB43 - 1	CCB43	-177	-0.0244	-7471B	PPB		12 Jan 2018 15:44:45	U	SMPL
J1015-03 - 1	CL-02-011118-B	10995	2.5646	-7471B	PPB		12 Jan 2018 15:47:01	U	SMPL
J1015-05 - 1	CL-02-011118-C	1096	0.2706	-7471B	PPB		12 Jan 2018 15:49:17	U	SMPL
J1015-07 - 1	CL-02-011118-D	1	0.0169	-7471B	PPB		12 Jan 2018 15:51:33	U	SMPL
J1018-03 - 1	OR-02-011118	648	0.1668	-7471B	PPB		12 Jan 2018 15:53:50	U	SMPL
J1094-01 - 1	TP-03	629	0.1624	-7471B	PPB		12 Jan 2018 15:56:06	U	SMPL
J1094-02 - 1	DBSW-2(0-4)	9582	2.2371	-7471B	PPB		12 Jan 2018 15:58:23	U	SMPL
J1094-03 - 1	DBSW-2(4-10)	29341	6.8160	-7471B	PPB		12 Jan 2018 16:00:40	U	SMPL
J1094-04 - 1	DBWW-4(5-8)	6504	1.5239	-7471B	PPB		12 Jan 2018 16:02:58	U	SMPL
J1096-01 - 1	TP-01-BOTTOM	5073	1.1922	-7471B	PPB		12 Jan 2018 16:05:17	U	SMPL
J1096-02 - 1	TP-02-BOTTOM	609	0.1578	-7471B	PPB		12 Jan 2018 16:07:33	U	SMPL
CCV44 - 1	CCV44	21942	5.1014	-7471B	PPB		12 Jan 2018 16:09:49	U	SMPL
CCB44 - 1	CCB44	-169	-0.0225	-7471B	PPB		12 Jan 2018 16:12:06	U	SMPL
J1106-01 - 1	CLEAN-SAND-13	366	0.1015	-7471B	PPB		12 Jan 2018 16:14:22	U	SMPL
J1106-01DUP - 1	CLEAN-SAND-13DUP	367	0.1017	-7471B	PPB		12 Jan 2018 16:16:38	U	SMPL
J1106-01MS - 1	CLEAN-SAND-13MS	17035	3.9642	-7471B	PPB		12 Jan 2018 16:18:54	U	SMPL
J1106-01MSD - 1	CLEAN-SAND-13MSD	16982	3.9520	-7471B	PPB		12 Jan 2018 16:21:11	U	SMPL
J1106-02 - 1	CLEAN-SAND-14	252	0.0751	-7471B	PPB		12 Jan 2018 16:23:27	U	SMPL
J1106-03 - 1	CLEAN-SAND-15	291	0.0841	-7471B	PPB		12 Jan 2018 16:25:44	U	SMPL
J1106-04 - 1	CLEAN-SAND-16	325	0.0920	-7471B	PPB		12 Jan 2018 16:28:01	U	SMPL
J1107-01 - 1	WASTE-CLASS-2	95623	22.1758	-7471B	PPB		12 Jan 2018 16:30:19	U	SMPL
J1106-01LX5 - 1		-406	-0.0774	-7471B	PPB		12 Jan 2018 16:32:36	U	SMPL
J1106-01A - 1		17589	4.0926	-7471B	PPB		12 Jan 2018 16:34:55	U	SMPL
CCV45 - 1	CCV45	21796	5.0675	-7471B	PPB		12 Jan 2018 16:37:11	U	SMPL
CCB45 - 1	CCB45	-179	-0.0248	-7471B	PPB		12 Jan 2018 16:39:28	U	SMPL
J1107-01DLX5 - 1		18655	4.3397	-7471B	PPB		12 Jan 2018 16:44:47	U	SMPL
CCV46 - 1	CCV46	21669	5.0381	-7471B	PPB		12 Jan 2018 16:47:03	U	SMPL
CCB46 - 1	CCB46	-173	-0.0234	-7471B	PPB		12 Jan 2018 16:49:20	U	SMPL