



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92591

Temp.: 20.2 °C

SLOPE.: 98.0

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/12/2018	14:00
2	CAL2	1	Water			7.01	01/12/2018	14:02
3	CAL3	1	Water			10.01	01/12/2018	14:04
4	ICV	1	Water			7.00	01/12/2018	14:06
5	CCV1	1	Water			2.01	01/12/2018	14:08
6	J1119-01	1	Water			7.89	01/12/2018	14:10
7	J1119-01DUP	1	Water			7.90	01/12/2018	14:12
8	CCV2	1	Water			12.01	01/12/2018	14:14

WORKLIST(Hardcopy Internal Chain)

LB92591

Worklist Name : PH011218

Worklist ID : 107372

Date : 1/12/2018 1:41:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/11/2018	Water	J1119-01	pH	Cool 4 deg C	STER10	K61	CELL-A-POOL-WATER(DE-1C	01/11/2018	SM 4500-PH B

Date/Time

01/12/18 1:46 PM

Received by:

SP

Relinquished by:

SP

Date/Time

01/12/18 3:24 PM

Received by:

SP

Relinquished by:

SP