



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92612

Temp.: 21.3 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/15/2018	09:30
2	CAL2	1	Water			7.00	01/15/2018	09:32
3	CAL3	1	Water			10.01	01/15/2018	09:34
4	ICV	1	Water			7.00	01/15/2018	09:36
5	CCV1	1	Water			2.01	01/15/2018	09:38
6	J1122-19	1	Solid	20	20	7.25	01/15/2018	09:40
7	J1122-19DUP	1	Solid	20	20	7.26	01/15/2018	09:42
8	J1122-23	1	Solid	20	20	7.53	01/15/2018	09:44
9	J1122-27	1	Solid	20	20	6.89	01/15/2018	09:46
10	J1123-04	1	Solid	20	20	8.59	01/15/2018	09:48
11	J1123-05	1	Solid	20	20	8.41	01/15/2018	09:50
12	J1123-06	1	Solid	20	20	9.43	01/15/2018	09:52
13	J1131-01	1	Solid	20	20	8.38	01/15/2018	09:54
14	J1131-02	1	Solid	20	20	8.88	01/15/2018	09:54
15	J1131-03	1	Solid	20	20	7.59	01/15/2018	09:56
16	CCV2	1	Water			12.01	01/15/2018	09:58
17	J1131-04	1	Solid	20	20	6.90	01/15/2018	10:00
18	CCV3	1	Water			2.02	01/15/2018	10:02

WORKLIST(Hardcopy Internal Chain)

LB 922612

Worklist Name : corrosivity011518

Worklist ID : 107406

Date : 1/15/2018 8:58:00 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/12/2018	Solid	J1122-19	Corrosivity	Cool 4 deg C	PSEG01	K52	G-1	01/12/2018	9045C
01/12/2018	Solid	J1122-23	Corrosivity	Cool 4 deg C	PSEG01	K52	G-2	01/12/2018	9045C
01/12/2018	Solid	J1122-27	Corrosivity	Cool 4 deg C	PSEG01	K52	G-3	01/12/2018	9045C
01/12/2018	Solid	J1123-04	Corrosivity	Cool 4 deg C	DIST01	K42	B-4(20)	01/12/2018	9045C
01/12/2018	Solid	J1123-05	Corrosivity	Cool 4 deg C	DIST01	K42	B-4(30)	01/12/2018	9045C
01/12/2018	Solid	J1123-06	Corrosivity	Cool 4 deg C	DIST01	K42	B-4(40)	01/12/2018	9045C
01/11/2018	Solid	J1131-01	Corrosivity	Cool 4 deg C	CLEA11	K13	OWS-2	01/11/2018	9045C
01/11/2018	Solid	J1131-02	Corrosivity	Cool 4 deg C	CLEA11	K13	OWS-3	01/11/2018	9045C
01/11/2018	Solid	J1131-03	Corrosivity	Cool 4 deg C	CLEA11	K13	OWS-1	01/11/2018	9045C
01/11/2018	Solid	J1131-04	Corrosivity	Cool 4 deg C	CLEA11	K13	OWS-4	01/11/2018	9045C

Date/Time

01/15/18 9:15 AM

Received by:

SP

Relinquished by:

SP

Date/Time

01/15/18 10:20 AM

Received by:

SP

Relinquished by:

SP