



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92613

Temp.: 19.6 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/15/2018	09:30
2	CAL2	1	Water			6.99	01/15/2018	09:32
3	CAL3	1	Water			10.01	01/15/2018	09:34
4	ICV	1	Water			7.00	01/15/2018	09:36
5	CCV1	1	Water			2.01	01/15/2018	12:30
6	J1143-01	1	Water			6.00	01/15/2018	12:32
7	J1143-01DUP	1	Water			6.01	01/15/2018	12:34
8	CCV2	1	Water			12.01	01/15/2018	12:36

WORKLIST(Hardcopy Internal Chain)

LB 92613

Worklist Name : PH011518

Worklist ID : 107436

Date : 1/15/2018 12:12:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/11/2018	Water	J1143-01	pH	NONE	STER08	L22	POOL-WATER	01/11/2018	SM 4500-PH B

Date/Time

01/15/18 12:15 AM

Received by:

SP

Relinquished by:

OP

Date/Time

01/15/18 12:45 PM

Received by:

SP

Relinquished by:

SP