


7470A

LB 92615
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36038



A= 0.0000e+000

B= 2.7873e-004

C= -9.8340e-003

Rho= 0.9999304

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	-90	0.000	-90				
0.2	0.200	0.206	0.006	775	0.0 %	775				
2.5	2.500	2.559	0.059	9218	0.0 %	9218				
5.0	5.000	5.007	0.007	18000	0.0 %	18000				
7.5	7.500	7.427	-0.073	26682	0.0 %	26682				
10.0	10.000	10.035	0.035	36038	0.0 %	36038				

LB92615 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Jan 2018 14:39:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-90	-	0.0000	7470A	PPB	15 Jan 2018 14:39:53	S	Std
0.2 - 1	50.2	775	-	0.2000	7470A	PPB	15 Jan 2018 14:42:02	S	Std
2.5 - 1	52.5	9218	-	2.5000	7470A	PPB	15 Jan 2018 14:44:10	S	Std
5.0 - 1	55	18000	-	5.0000	7470A	PPB	15 Jan 2018 14:46:18	S	Std
7.5 - 1	57.5	26682	-	7.5000	7470A	PPB	15 Jan 2018 14:48:26	S	Std
10.0 - 1	50	36038	-	10.0000	7470A	PPB	15 Jan 2018 14:50:34	S	Std
ICV47 - 1	ICV47	13732	3.8176	-7470A	PPB	15 Jan 2018 14:53:06	U	SMPL	
ICB47 - 1	ICB47	-168	-0.0567	-7470A	PPB	15 Jan 2018 14:55:12	U	SMPL	
CCV57 - 1	CCV57	18059	5.0237	-7470A	PPB	15 Jan 2018 14:57:19	U	SMPL	
CCB57 - 1	CCB57	-179	-0.0597	-7470A	PPB	15 Jan 2018 14:59:25	U	SMPL	
CRA - 1	CRA	723	0.1917	-7470A	PPB	15 Jan 2018 15:01:33	U	SMPL	
HighStd - 1	HighStd	36783	10.2426	-7470A	PPB	15 Jan 2018 15:03:39	U	SMPL	
ChkStd - 1	ChkStd	25991	7.2345	-7470A	PPB	15 Jan 2018 15:05:46	U	SMPL	
PB105616BL - 1	PBW	-217	-0.0703	-7470A	PPB	15 Jan 2018 15:07:55	U	SMPL	
PB105616BS - 1	LCSW	13806	3.8383	-7470A	PPB	15 Jan 2018 15:10:04	U	SMPL	
PB105616TB - 1	PB105590TB	-154	-0.0528	-7470A	PPB	15 Jan 2018 15:12:12	U	SMPL	
J1010-02 - 1	BU-02-011118	-58	-0.0260	-7470A	PPB	15 Jan 2018 15:14:20	U	SMPL	
J1010-02DUP - 1	BU-02-011118DUP	-48	-0.0232	-7470A	PPB	15 Jan 2018 15:16:28	U	SMPL	
J1010-02MS - 1	BU-02-011118MS	13524	3.7597	-7470A	PPB	15 Jan 2018 15:18:36	U	SMPL	
J1010-02MSD - 1	BU-02-011118MSD	13722	3.8148	-7470A	PPB	15 Jan 2018 15:20:43	U	SMPL	
CCV58 - 1	CCV58	17748	4.9370	-7470A	PPB	15 Jan 2018 15:22:49	U	SMPL	
CCB58 - 1	CCB58	-160	-0.0544	-7470A	PPB	15 Jan 2018 15:24:56	U	SMPL	
J1011-11 - 1	HR-05-011118-A	-65	-0.0280	-7470A	PPB	15 Jan 2018 15:27:03	U	SMPL	
J1011-13 - 1	HR-05-011118-B	-46	-0.0227	-7470A	PPB	15 Jan 2018 15:29:10	U	SMPL	
J1011-15 - 1	HR-05-011118-C	-46	-0.0227	-7470A	PPB	15 Jan 2018 15:31:16	U	SMPL	
J1015-02 - 1	CL-02-011118-A	-40	-0.0210	-7470A	PPB	15 Jan 2018 15:33:23	U	SMPL	
J1015-04 - 1	CL-02-011118-B	-34	-0.0193	-7470A	PPB	15 Jan 2018 15:35:31	U	SMPL	
J1015-06 - 1	CL-02-011118-C	-61	-0.0268	-7470A	PPB	15 Jan 2018 15:37:38	U	SMPL	
J1015-08 - 1	CL-02-011118-D	-40	-0.0210	-7470A	PPB	15 Jan 2018 15:39:46	U	SMPL	
J1018-04 - 1	OR-01-011118	-50	-0.0238	-7470A	PPB	15 Jan 2018 15:41:54	U	SMPL	
J1098-01 - 1	OWS-1	-16	-0.0143	-7470A	PPB	15 Jan 2018 15:44:03	U	SMPL	
J1098-02 - 1	OWS-2	-38	-0.0204	-7470A	PPB	15 Jan 2018 15:46:10	U	SMPL	
CCV59 - 1	CCV59	17937	4.9897	-7470A	PPB	15 Jan 2018 15:48:17	U	SMPL	
CCB59 - 1	CCB59	-168	-0.0567	-7470A	PPB	15 Jan 2018 15:50:24	U	SMPL	
J1098-03 - 1	OWS-3	7	-0.0079	-7470A	PPB	15 Jan 2018 15:52:31	U	SMPL	
J1098-04 - 1	OWS-4	-46	-0.0227	-7470A	PPB	15 Jan 2018 15:54:39	U	SMPL	
J1100-01 - 1	OWS-1	-16	-0.0143	-7470A	PPB	15 Jan 2018 15:56:46	U	SMPL	
J1107-02 - 1	WASTE-CLASS-2	-52	-0.0243	-7470A	PPB	15 Jan 2018 15:58:53	U	SMPL	
J1010-02LX5 - 1		-125	-0.0447	-7470A	PPB	15 Jan 2018 16:03:25	U	SMPL	
J1010-02A - 1		13823	3.8430	-7470A	PPB	15 Jan 2018 16:05:32	U	SMPL	
CCV60 - 1	CCV60	17951	4.9936	-7470A	PPB	15 Jan 2018 16:07:40	U	SMPL	
CCB60 - 1	CCB60	-236	-0.0756	-7470A	PPB	15 Jan 2018 16:09:49	U	SMPL	