

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: Aparna

Date: 01/16/2018

DATE & TIME IN: 01/15/2018 16:00

OVEN TEMP: 103 °C

Run Number: LB92617

DATE & TIME OUT: 01/16/2018 10:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB92617BL	LB92617BL	76.777	76.777	100	76.777	76.777	76.777	0	0
2	J1112-03	ENV-106-1 (MW)	75.0924	75.0924	100	75.214	75.214	75.214	0.1216	1216
3	J1112-04	ENV-116-6 (MW)	74.5378	74.5378	100	74.6293	74.6293	74.6293	0.0915	915
4	J1112-04DUP	ENV-116-6 (MW) DUP	82.1697	82.1697	100	82.2615	82.2615	82.2615	0.0918	918
5	J1112-05	ENV-125-11 (MW)	91.0125	91.0125	100	91.2005	91.2005	91.2005	0.188	1880
6	J1112-08	ENV-106-1 (MW) -DUP	93.077	93.077	100	93.1973	93.1973	93.1973	0.1203	1203

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS011518

WorkList ID : 107433

Date : 1/15/2018 11:36:20 AM

1892617

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/17/2018	Water	J1112-03	TS	Cool 4 deg C	THEP04	K63	ENV-106-1(MW)	01/10/2018	SM2540 B
01/18/2018	Water	J1112-04	TS	Cool 4 deg C	THEP04	K63	ENV-116-6(MW)	01/11/2018	SM2540 B
01/18/2018	Water	J1112-05	TS	Cool 4 deg C	THEP04	K63	ENV-125-11(MW)	01/11/2018	SM2540 B
01/17/2018	Water	J1112-08	TS	Cool 4 deg C	THEP04	K63	ENV-106-1(MW)-DUP	01/10/2018	SM2540 B

Date/Time

Received by:

Relinquished by:

01/15/18, 14:30

AS

CP

Date/Time

Received by:

Relinquished by:

01/15/18 16:15

CP

AS