

245.1


LB92621

INSTRUMENT ID: CV2

Linear

 μ Abs.:

32228



A= 0.0000e+000

B= 3.0564e-004

C= 2.2791e-002

Rho= 0.9997345

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	-27	0.000	-27				
0.05	0.050	0.059	0.009	117	0.0 %	117				
0.2	0.200	0.157	-0.043	440	0.0 %	440				
2.5	2.500	2.483	-0.017	8048	0.0 %	8048				
5.0	5.000	4.984	-0.016	16232	0.0 %	16232				
7.5	7.500	7.680	0.180	25053	0.0 %	25053				
10.0	10.000	9.873	-0.127	32228	0.0 %	32228				

LB92621 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Jan 2018 16:16:54

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-27	-	0.0000245.1	PPB	15 Jan 2018 16:18:57	S	Std
0.05 - 1	50.65	117	-	0.0500245.1	PPB	15 Jan 2018 16:23:37	S	Std
0.2 - 1	50.2	440	-	0.2000245.1	PPB	15 Jan 2018 16:25:45	S	Std
2.5 - 1	52.5	8048	-	2.5000245.1	PPB	15 Jan 2018 16:27:52	S	Std
5.0 - 1	55	16232	-	5.0000245.1	PPB	15 Jan 2018 16:30:00	S	Std
7.5 - 1	57.5	25053	-	7.5000245.1	PPB	15 Jan 2018 16:32:08	S	Std
10.0 - 1	50	32228	-	10.0000245.1	PPB	15 Jan 2018 16:34:16	S	Std
ICV48 - 1	ICV48	12918	3.9711	-245.1	PPB	15 Jan 2018 16:37:31	U	SMPL
ICB48 - 1	ICB48	-114	-0.0121	-245.1	PPB	15 Jan 2018 16:39:37	U	SMPL
CCV61 - 1	CCV61	16408	5.0378	-245.1	PPB	15 Jan 2018 16:41:45	U	SMPL
CCB61 - 1	CCB61	-177	-0.0313	-245.1	PPB	15 Jan 2018 16:43:51	U	SMPL
CRA - 1	CRA	449	0.1600	-245.1	PPB	15 Jan 2018 16:45:59	U	SMPL
HighStd - 1	HighStd	32748	10.0320	-245.1	PPB	15 Jan 2018 16:48:05	U	SMPL
ChkStd - 1	ChkStd	23219	7.1195	-245.1	PPB	15 Jan 2018 16:50:12	U	SMPL
PB105662BL - 1	PBW	-202	-0.0389	-245.1	PPB	15 Jan 2018 16:52:21	U	SMPL
PB105662BS - 1	LCSW	13431	4.1279	-245.1	PPB	15 Jan 2018 16:54:31	U	SMPL
J1112-03 - 1	ENV-106-1(MW)	-214	-0.0426	-245.1	PPB	15 Jan 2018 16:56:38	U	SMPL
J1112-04 - 1	ENV-116-6(MW)	-98	-0.0072	-245.1	PPB	15 Jan 2018 16:58:47	U	SMPL
J1112-05 - 1	ENV-125-11(MW)	670	0.2276	-245.1	PPB	15 Jan 2018 17:00:55	U	SMPL
J1112-05DUP - 1	ENV-125-11(MW)DUP	670	0.2276	-245.1	PPB	15 Jan 2018 17:08:28	U	SMPL
J1112-05MS - 1	ENV-125-11(MW)MS	15527	4.7685	-245.1	PPB	15 Jan 2018 17:10:34	U	SMPL
CCV62 - 1	CCV62	16116	4.9485	-245.1	PPB	15 Jan 2018 17:12:40	U	SMPL
CCB62 - 1	CCB62	-150	-0.0231	-245.1	PPB	15 Jan 2018 17:14:47	U	SMPL
J1112-05MSD - 1	ENV-125-11(MW)MSD	15623	4.7978	-245.1	PPB	15 Jan 2018 17:16:54	U	SMPL
J1112-08 - 1	ENV-106-1(MW)-DUP	-202	-0.0389	-245.1	PPB	15 Jan 2018 17:19:01	U	SMPL
J1132-01 - 1	001M-WILLETS-PT-BLVD (JAN)	-63	0.0035	-245.1	PPB	15 Jan 2018 17:21:07	U	SMPL
J1132-02 - 1	002M-35TH-AVE (JAN)	-80	-0.0017	-245.1	PPB	15 Jan 2018 17:23:14	U	SMPL
J1133-01 - 1	001M-WILLETS-PT-BLVD (DEC)	-74	0.0002	-245.1	PPB	15 Jan 2018 17:25:22	U	SMPL
J1112-05LX5 - 1		89	0.0500	-245.1	PPB	15 Jan 2018 17:27:30	U	SMPL
J1112-05A - 1		13890	4.2682	-245.1	PPB	15 Jan 2018 17:29:38	U	SMPL
CCV63 - 1	CCV63	16144	4.9571	-245.1	PPB	15 Jan 2018 17:31:46	U	SMPL
CCB63 - 1	CCB63	-157	-0.0252	-245.1	PPB	15 Jan 2018 17:33:55	U	SMPL