



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB92624

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	01/15/2018	09:34
2	J1112-03		1	>100.0	>212.0	01/15/2018	10:10
3	J1112-03DUP		1	>100.0	>212.0	01/15/2018	10:45
4	J1112-04		1	>100.0	>212.0	01/15/2018	11:10
5	J1112-05		1	>100.0	>212.0	01/15/2018	11:30
6	J1112-08		1	>100.0	>212.0	01/15/2018	11:50

Result = (Celsius * 1.8) + 32

LB92624

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP011518

WorkList ID : 107401

Date : 1/15/2018 8:47:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2018	Water	J1112-03	Flash Point	Cool 4 deg C	THEP04	K63	ENV-106-1(MW)	01/10/2018	1010A
01/18/2018	Water	J1112-04	Flash Point	Cool 4 deg C	THEP04	K63	ENV-116-6(MW)	01/11/2018	1010A
01/18/2018	Water	J1112-05	Flash Point	Cool 4 deg C	THEP04	K63	ENV-125-11(MW)	01/11/2018	1010A
01/18/2018	Water	J1112-08	Flash Point	Cool 4 deg C	THEP04	K63	ENV-106-1(MW)-DUP	01/10/2018	1010A

Date/Time 1/15/18 9:00 AM
 Received by: KS
 Relinquished by: CP

Date/Time 1/15/18 12:00 PM
 Received by: [Signature]
 Relinquished by: KS