



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92632

Temp.: 20.2 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/16/2018	09:45
2	CAL2	1	Water			7.00	01/16/2018	09:47
3	CAL3	1	Water			10.01	01/16/2018	09:49
4	ICV	1	Water			7.00	01/16/2018	09:51
5	CCV1	1	Water			2.01	01/16/2018	09:53
6	J1014-02	1	Solid	20	20	8.89	01/16/2018	09:55
7	J1014-02DUP	1	Solid	20	20	8.90	01/16/2018	09:57
8	J1145-03	1	Solid	20	20	8.71	01/16/2018	09:59
9	CCV2	1	Water			12.01	01/16/2018	10:00

WORKLIST(Hardcopy Internal Chain)

LB 92632

Worklist Name : corrosivity011618

Worklist ID : 107461

Date : 1/16/2018 9:26:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/15/2018	Solid	J1014-02	Corrosivity	Cool 4 deg C	PSEG05	K52	HD-01-011518	01/15/2018	9045C
01/14/2018	Solid	J1145-03	Corrosivity	Cool 4 deg C	WSP001	K11	PB-1(8-21.5)	01/14/2018	9045C

Date/Time

01/16/18 9:30 AM

Received by:

SP

Relinquished by:

LB

Date/Time

01/16/18 10:15 AM

Received by:

SP

Relinquished by:

SP