

2192637

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/16/2018 13:33

Test: Reactive ~~Total~~ CN, ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.077	0.0	0.079	
ICB1	1.489	0.0	0.004	
CCV1	240.001	0.0	0.192	
CCB1	1.444	0.0	0.004	
PB105695BL	1.337	0.0	0.004	
J1014-02	1.091	0.0	0.003	
J1014-02DUP	1.383	0.0	0.004	
J1145-03	1.225	0.0	0.003	
CCV2	253.955	0.0	0.204	
CCB2	1.183	0.0	0.003	

N	10
Mean	59.919
SD	102.9748
CV%	171.86

AS.

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 16 13:27:21 2018

Tue Jan 16 13:31:14 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	0.9973	µg/l	1/16/2018 11:36:23	
5.0PPBCN	A	Reactive CN	P	4.5307	µg/l	1/16/2018 11:36:24	
10PPBCN	A	Reactive CN	P	9.95	µg/l	1/16/2018 11:36:25	
50PPBCN	A	Reactive CN	P	50.4944	µg/l	1/16/2018 11:36:26	
100PPBCN	A	Reactive CN	P	100.5114	µg/l	1/16/2018 11:36:27	
250PPBCN	A	Reactive CN	P	247.3243	µg/l	1/16/2018 11:36:28	
500PPBCN	A	Reactive CN	P	501.1918	µg/l	1/16/2018 11:36:29	
ICV1	S	Reactive CN	P	96.0773	µg/l	1/16/2018 13:27:22	
ICB1	S	Reactive CN	P	1.4889	µg/l	1/16/2018 13:27:23	
CCV1	S	Reactive CN	P	240.001	µg/l	1/16/2018 13:27:25	
CCB1	S	Reactive CN	P	1.4443	µg/l	1/16/2018 13:27:28	
PB105695BL	S	Reactive CN	P	1.3368	µg/l	1/16/2018 13:27:29	
J1014-02	S	Reactive CN	P	1.0905	µg/l	1/16/2018 13:27:30	
J1014-02DUP	S	Reactive CN	P	1.3834	µg/l	1/16/2018 13:27:31	
J1145-03	S	Reactive CN	P	1.2249	µg/l	1/16/2018 13:31:12	
CCV2	S	Reactive CN	P	253.955	µg/l	1/16/2018 13:31:13	
CCB2	S	Reactive CN	P	1.1833	µg/l	1/16/2018 13:31:14	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/16/2018 11:41


Test Total CN

Accepted 1/16/2018 11:41

Factor 1263
Bias 0.002

Coeff. of det. 0.999950

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.9973	0.0000	
2	5.0PPBCN	0.006	4.5307	5.0000	
3	10PPBCN	0.010	9.9500	10.0000	
4	50PPBCN	0.042	50.4944	50.0000	
5	100PPBCN	0.082	100.5114	100.0000	
6	250PPBCN	0.198	247.3243	250.0000	
7	500PPBCN	0.399	501.1918	500.0000	