



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: HIRAL

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB92643

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	01/16/2018	10:15
2	J1149-01		1	>100.0	>212.0	01/16/2018	10:46
3	J1149-01DUP		1	>100.0	>212.0	01/16/2018	11:18

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB92643

Worklist Name : FP011618

Worklist ID : 107459

Date : 1/16/2018 8:19:06 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2018	Water	J1149-01	Flash Point	Cool 4 deg C	THEP04	DrumPW		01/15/2018	1010A

Date/Time 01/16/18 8:25
 Received by: HP
 Relinquished by: CP

Date/Time 01/16/18 11:25
 Received by: CP
 Relinquished by: HP