



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB92645

Reagent/Standard	Lot/Loq #
TURBIDITY ICV 10 NTU	WP61019
TURBIDITY CCV 10 NTU	WP61020
Turbidity std, 40NTU	WP61014
turbidity std 5 NTU	WP61017
turbidity std 20 NTU	WP61015
Turbidity std 1 NTU	WP61018
turbidity std 0 NTU	WP61022
turbidity std 10 NTU	WP61016
10 NTU Standard 500 ml	W2242

Intercept: 0.0065

Slope: 0.992

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.089	01/17/2018	10:17
2	CAL2	1	1	1.011	01/17/2018	10:18
3	CAL3	5	1	4.970	01/17/2018	10:20
4	CAL4	10	1	9.960	01/17/2018	10:22
5	CAL5	20	1	19.600	01/17/2018	10:24
6	CAL6	40	1	39.800	01/17/2018	10:22



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB92645

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.970	10.044	01/17/2018	10:27
2	ICB		1	0.090	0.084	01/17/2018	10:28
3	CCV1	10	1	9.910	9.983	01/17/2018	10:29
4	CCB1		1	0.091	0.085	01/17/2018	10:31
5	LB92645BL		1	0.088	0.082	01/17/2018	10:32
6	J1160-01		1	1.020	1.022	01/17/2018	10:35
7	J1160-02		1	0.193	0.188	01/17/2018	10:37
8	J1160-02DUP		1	0.190	0.185	01/17/2018	10:39
9	J1160-03		1	0.226	0.221	01/17/2018	10:41
10	J1160-04		1	0.203	0.198	01/17/2018	10:44
11	CCV2	10	1	9.920	9.993	01/17/2018	10:46
12	CCB2		1	0.091	0.085	01/17/2018	10:49

0892645

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TUR011718

WorkList ID : 107498

Date : 1/17/2018 9:09:35 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2018	Water	J1160-01	A Turbidity	Cool 4 deg C	WARR01	L22	DW-1-1ST-FL-BOILER-RM-PC	01/16/2018	SM2130 B
01/18/2018	Water	J1160-02	Turbidity	Cool 4 deg C	WARR01	L22	DW-2-4TH-FL-RM-408	01/16/2018	SM2130 B
01/18/2018	Water	J1160-03	Turbidity	Cool 4 deg C	WARR01	L22	DW-3-8TH-FL-RM-815	01/16/2018	SM2130 B
01/18/2018	Water	J1160-04	✓ Turbidity	Cool 4 deg C	WARR01	L22	DW-4-14TH-FL-RM-1408	01/16/2018	SM2130 B

Date/Time 01-17-18 9:30
 Received by: AR
 Relinquished by: CR

Date/Time 01-17-18 10:55
 Received by: [Signature]
 Relinquished by: AR