



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92649

Temp.: 23.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/16/2018	16:20
2	CAL2	1	Water			7.00	01/16/2018	16:22
3	CAL3	1	Water			10.01	01/16/2018	16:24
4	ICV	1	Water			7.00	01/16/2018	16:26
5	CCV1	1	Water			2.01	01/16/2018	16:28
6	J1149-01	1	Water			7.38	01/16/2018	16:30
7	J1149-01DUP	1	Water			7.39	01/16/2018	16:32
8	J1160-01	1	Water			7.02	01/16/2018	16:34
9	J1160-02	1	Water			6.99	01/16/2018	16:36
10	J1160-03	1	Water			7.03	01/16/2018	16:38
11	J1160-04	1	Water			7.05	01/16/2018	16:40
12	CCV2	1	Water			12.01	01/16/2018	16:42

WORKLIST(Hardcopy Internal Chain)

LB 9 2649

Worklist Name : pH j 1149

Worklist ID : 107477

Date : 1/16/2018 4:01:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/15/2018	Water	J1149-01	pH	Cool 4 deg C	THEP04	L22	DRUM-PW	01/15/2018	SM 4500-PH B
01/16/2018	Water	J1160-01	pH	Cool 4 deg C	WARR01	L22	DW-1-1ST-FL-BOILER-RM-PC	01/16/2018	SM 4500-PH B
01/16/2018	Water	J1160-02	pH	Cool 4 deg C	WARR01	L22	DW-2-4TH-FL-RM-408	01/16/2018	SM 4500-PH B
01/16/2018	Water	J1160-03	pH	Cool 4 deg C	WARR01	L22	DW-3-8TH-FL-RM-815	01/16/2018	SM 4500-PH B
01/16/2018	Water	J1160-04	pH	Cool 4 deg C	WARR01	L22	DW-4-14TH-FL-RM-1408	01/16/2018	SM 4500-PH B

Date/Time

01/16/18 4:10 PM

Received by:

SP

Relinquished by:

SP

Date/Time

01/16/18 5:00 PM

Received by:

SP

Relinquished by:

SP