

7470A


LB 92650

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

36821



A= 0.0000e+000

B= 2.7418e-004

C= -9.8248e-003

Rho= 0.9998154

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	154	0.000	154				
0.2	0.200	0.244	0.044	927	0.0 %	927				
2.5	2.500	2.504	0.004	9170	0.0 %	9170				
5.0	5.000	4.862	-0.138	17770	0.0 %	17770				
7.5	7.500	7.471	-0.029	27285	0.0 %	27285				
10.0	10.000	10.086	0.086	36821	0.0 %	36821				

LB92650 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Jan 2018 10:36:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	154	-	0.0000	7470A	PPB	17 Jan 2018 10:51:29	S	Std
0.2 - 1	50.2	927	-	0.2000	7470A	PPB	17 Jan 2018 10:53:37	S	Std
2.5 - 1	52.5	9170	-	2.5000	7470A	PPB	17 Jan 2018 10:55:45	S	Std
5.0 - 1	55	17770	-	5.0000	7470A	PPB	17 Jan 2018 10:57:53	S	Std
7.5 - 1	57.5	27285	-	7.5000	7470A	PPB	17 Jan 2018 11:00:01	S	Std
10.0 - 1	510	36821	-	10.0000	7470A	PPB	17 Jan 2018 11:02:09	S	Std
ICV50 - 1	ICV50	14340	3.9218	-	7470A	PPB	17 Jan 2018 11:05:06	U	SMPL
ICB50 - 1	ICB50	40	0.0011	-	7470A	PPB	17 Jan 2018 11:07:12	U	SMPL
CCV73 - 1	CCV73	17720	4.8486	-	7470A	PPB	17 Jan 2018 11:09:20	U	SMPL
CCB73 - 1	CCB73	55	0.0053	-	7470A	PPB	17 Jan 2018 11:11:26	U	SMPL
CRA - 1	CRA	939	0.2476	-	7470A	PPB	17 Jan 2018 11:13:34	U	SMPL
HighStd - 1	HighStd	36878	10.1012	-	7470A	PPB	17 Jan 2018 11:15:40	U	SMPL
ChkStd - 1	ChkStd	25763	7.0538	-	7470A	PPB	17 Jan 2018 11:17:47	U	SMPL
PB105705BL - 1	PBW	-99	-0.0370	-	7470A	PPB	17 Jan 2018 11:19:56	U	SMPL
PB105705BS - 1	LCSW	13181	3.6041	-	7470A	PPB	17 Jan 2018 11:22:05	U	SMPL
PB105705TB - 1	PB105686TB	67	0.0085	-	7470A	PPB	17 Jan 2018 11:24:13	U	SMPL
J1014-02 - 1	HD-01-011518	137	0.0277	-	7470A	PPB	17 Jan 2018 11:26:21	U	SMPL
J1145-03 - 1	PB-1(8-21.5)	228	0.0527	-	7470A	PPB	17 Jan 2018 11:28:29	U	SMPL
J1145-03DUP - 1	PB-1(8-21.5)DUP	204	0.0461	-	7470A	PPB	17 Jan 2018 11:30:38	U	SMPL
J1145-03MS - 1	PB-1(8-21.5)MS	13168	3.6005	-	7470A	PPB	17 Jan 2018 11:32:45	U	SMPL
CCV74 - 1	CCV74	17643	4.8275	-	7470A	PPB	17 Jan 2018 11:34:52	U	SMPL
CCB74 - 1	CCB74	73	0.0102	-	7470A	PPB	17 Jan 2018 11:36:58	U	SMPL
J1145-03MSD - 1	PB-1(8-21.5)MSD	12826	3.5067	-	7470A	PPB	17 Jan 2018 11:39:05	U	SMPL
PB105706BL - 1	PBW	-56	-0.0252	-	7470A	PPB	17 Jan 2018 11:41:12	U	SMPL
PB105706BS - 1	LCSW	13612	3.7222	-	7470A	PPB	17 Jan 2018 11:46:03	U	SMPL
J1149-01 - 1	DRUM-PW	-43	-0.0216	-	7470A	PPB	17 Jan 2018 11:48:09	U	SMPL
J1149-01DUP - 1	DRUM-PWDUP	15	-0.0057	-	7470A	PPB	17 Jan 2018 11:50:17	U	SMPL
J1149-01MS - 1	DRUM-PWMS	14167	3.8744	-	7470A	PPB	17 Jan 2018 11:52:25	U	SMPL
J1149-01MSD - 1	DRUM-PWMSD	13905	3.8026	-	7470A	PPB	17 Jan 2018 11:54:33	U	SMPL
CCV75 - 1	CCV75	18098	4.9522	-	7470A	PPB	17 Jan 2018 11:56:41	U	SMPL
CCB75 - 1	CCB75	46	0.0028	-	7470A	PPB	17 Jan 2018 11:58:49	U	SMPL
J1145-03LX5 - 1		132	0.0264	-	7470A	PPB	17 Jan 2018 12:00:57	U	SMPL
J1145-03A - 1		14472	3.9580	-	7470A	PPB	17 Jan 2018 12:03:04	U	SMPL
J1149-01LX5 - 1		45	0.0025	-	7470A	PPB	17 Jan 2018 12:05:11	U	SMPL
J1149-01A - 1		14817	4.0526	-	7470A	PPB	17 Jan 2018 12:07:18	U	SMPL
CCV76 - 1	CCV76	17779	4.8647	-	7470A	PPB	17 Jan 2018 12:09:25	U	SMPL
CCB76 - 1	CCB76	38	0.0006	-	7470A	PPB	17 Jan 2018 12:11:32	U	SMPL