



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92653

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/17/2018	09:20
2	CAL2	1	Water			6.99	01/17/2018	09:22
3	CAL3	1	Water			10.01	01/17/2018	09:24
4	ICV	1	Water			7.00	01/17/2018	09:26
5	CCV1	1	Water			2.01	01/17/2018	09:28
6	J1021-02	1	Solid	20	20	7.27	01/17/2018	09:30
7	J1021-02DUP	1	Solid	20	20	7.28	01/17/2018	09:32
8	J1151-03	1	Solid	20	20	8.86	01/17/2018	09:34
9	J1153-01	1	Solid	20	20	9.72	01/17/2018	09:36
10	J1153-02	1	Solid	20	20	11.79	01/17/2018	09:38
11	J1154-03	1	Solid	20	20	9.07	01/17/2018	09:40
12	J1156-04	1	Solid	20	20	6.71	01/17/2018	09:42
13	J1156-13	1	Solid	20	20	8.80	01/17/2018	09:44
14	J1158-01	1	Solid	20	20	8.16	01/17/2018	09:46
15	J1158-02	1	Solid	20	20	6.39	01/17/2018	09:48
16	CCV2	1	Water			12.01	01/17/2018	09:50
17	J1158-03	1	Solid	20	20	7.65	01/17/2018	09:52
18	J1158-04	1	Solid	20	20	8.67	01/17/2018	09:54
19	J1158-05	1	Solid	20	20	8.63	01/17/2018	09:56
20	J1159-01	1	Solid	20	20	8.65	01/17/2018	09:58

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	J1163-06	1	Solid	20	20	4.46	01/17/2018	10:00
22	J1163-12	1	Solid	20	20	4.91	01/17/2018	10:02
23	J1167-02	1	Solid	20	20	5.74	01/17/2018	10:04
24	J1167-04	1	Solid	20	20	7.43	01/17/2018	10:06
25	J1167-06	1	Solid	20	20	7.33	01/17/2018	10:08
26	CCV3	1	Water			2.01	01/17/2018	10:10

WORKLIST(Hardcopy Internal Chain)

LB92653

WorkList Name : corrosivity011718

WorkList ID : 107490

Date : 1/17/2018 8:25:33 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/16/2018	Solid	J1021-02	Corrosivity	Cool 4 deg C	PSEG05	L52	SA-01-011618	01/16/2018	9045C
01/16/2018	Solid	J1151-03	Corrosivity	Cool 4 deg C	PSEG02	L52	OIL-GRAVEL	01/16/2018	9045C
01/11/2018	Solid	J1153-01	Corrosivity	Cool 4 deg C	ATCE02	L22	170921PS-01-LCC-POLY-SHE	01/11/2018	9045C
01/11/2018	Solid	J1153-02	Corrosivity	Cool 4 deg C	ATCE02	L22	170921PS-02-LCC-PEEL-AW	01/11/2018	9045C
01/16/2018	Solid	J1154-03	Corrosivity	Cool 4 deg C	PSEG01	L52	RT-1867-3407-4235	01/16/2018	9045C
01/16/2018	Solid	J1156-04	Corrosivity	Cool 4 deg C	PSEG01	L52	SP-COMP	01/16/2018	9045C
01/16/2018	Solid	J1156-13	Corrosivity	Cool 4 deg C	PSEG01	L52	OILY-STONES	01/16/2018	9045C
01/12/2018	Solid	J1158-01	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-1	01/12/2018	9045C
01/12/2018	Solid	J1158-02	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-2	01/12/2018	9045C
01/12/2018	Solid	J1158-03	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-3	01/12/2018	9045C
01/12/2018	Solid	J1158-04	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-4	01/12/2018	9045C
01/12/2018	Solid	J1158-05	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-5	01/12/2018	9045C
01/12/2018	Solid	J1159-01	Corrosivity	Cool 4 deg C	CLEA11		FD-1	01/12/2018	9045C
01/16/2018	Solid	J1163-06	Corrosivity	Cool 4 deg C	PSEG01	L52	TP-1-(0-4)	01/16/2018	9045C
01/16/2018	Solid	J1163-12	Corrosivity	Cool 4 deg C	PSEG01	L52	TP-2-(0-4)	01/16/2018	9045C
01/16/2018	Solid	J1167-02	Corrosivity	Cool 4 deg C	PSEG01		RO-72-11995	01/16/2018	9045C
01/16/2018	Solid	J1167-04	Corrosivity	Cool 4 deg C	PSEG01		RO-72-12000	01/16/2018	9045C
01/16/2018	Solid	J1167-06	Corrosivity	Cool 4 deg C	PSEG01	L52	DRUM-1-2	01/16/2018	9045C

Date/Time 01/17/18 9:10 AM

Received by: SP

Relinquished by: CP

Date/Time 01/17/18 10:30 AM

Received by: SP

Relinquished by: SP