



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92663

Temp.: 24.1 °C

SLOPE.: 99.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/17/2018	13:35
2	CAL2	1	Water			7.00	01/17/2018	13:37
3	CAL3	1	Water			10.01	01/17/2018	13:39
4	ICV	1	Water			7.00	01/17/2018	13:41
5	CCV1	1	Water			2.01	01/17/2018	13:43
6	J1176-01	1	Water			5.05	01/17/2018	13:45
7	J1176-01DUP	1	Water			5.06	01/17/2018	13:47
8	CCV2	1	Water			12.01	01/17/2018	13:50

WORKLIST(Hardcopy Internal Chain)

LB 92663

Worklist Name : pH J1176

Worklist ID : 107509

Date : 1/17/2018 1:21:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/16/2018	Water	J1176-01	pH	Cool 4 deg C	STER01	L31	POOL-WATER	01/16/2018	SM 4500-PH B

Date/Time 01/17/18 1:25 PM

Received by: SP

Relinquished by: SP

Date/Time 01/17/18 2:10 PM

Received by: SP

Relinquished by: SP