



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92676

Temp.: 17.6 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/17/2018	18:00
2	CAL2	1	Water			6.99	01/17/2018	18:01
3	CAL3	1	Water			10.02	01/17/2018	18:02
4	ICV	1	Water			7.00	01/17/2018	18:03
5	CCV1	1	Water			2.01	01/17/2018	18:05
6	J1173-02	1	Water			2.47	01/17/2018	18:06
7	J1173-02DUP	1	Water			2.48	01/17/2018	18:07
8	CCV2	1	Water			12.01	01/17/2018	18:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY j1173 WorkList ID : 107518 Date : 1/17/2018 5:45:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/17/2018	Water	J1173-02	Corrosivity	Cool 4 deg C	PSEG01	L52	UST-1-2	01/17/2018	9040C

Date/Time 01-17-18 5:50 PM
 Received by: JD
 Relinquished by: CB

Date/Time 01-17-18 6:40 PM
 Received by: JP
 Relinquished by: JP