



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB92688

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP60858
COD calibration std. 100 ppm	WP60855
COD calibration std. 10 ppm	WP60854
COD calibration std. 150 ppm	WP60857
COD calibration std. 0 ppm	WP60853
COD Digestion Vials Low Level 0-150Mg/L	W2291
COD ICV-LCS std, 50ppm	WP61077
COD CCV std, 50ppm	WP61078

Temp In (C): <u>150</u>	Date In: <u>01/18/2018</u>	Time In: <u>09:30</u>
Temp Out (C): <u>150</u>	Date Out: <u>01/18/2018</u>	Time Out: <u>11:30</u>

Intercept: -0.3904

Slope: 1.0063

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	01/10/2018	12:00
2	CAL2	10	1	Water	10.000	01/10/2018	12:00
3	CAL3	50	1	Water	49.000	01/10/2018	12:01
4	CAL4	100	1	Water	100.000	01/10/2018	12:01
5	CAL5	150	1	Water	151.000	01/10/2018	12:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB92688

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	50.075	01/10/2018	12:02
2	ICB				1	Water	0.000	0.388	01/10/2018	12:03
3	CCV1	50			1	Water	51.000	51.069	01/18/2018	12:30
4	CCB1				1	Water	0.000	0.388	01/18/2018	12:30
5	LB92688BL				1	Water	0.000	0.388	01/18/2018	12:31
6	LB92688BS	50			1	Water	50.000	50.075	01/18/2018	12:31
7	J1173-01				1	Water	40.000	40.138	01/18/2018	12:32
8	J1173-01DUP				1	Water	41.000	41.131	01/18/2018	12:32
9	J1173-01MS	50			1	Water	90.000	89.825	01/18/2018	12:33
10	J1173-01MSD	50			1	Water	92.000	91.812	01/18/2018	12:33
11	CCV2	50			1	Water	51.000	51.069	01/18/2018	12:34
12	CCB2				1	Water	0.000	0.388	01/18/2018	12:34