

7471B


LB 92693

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43458



A= 0.0000e+000

B= 2.2913e-004

C= -1.7737e-002

Rho= 0.9999425

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	28	0.000	28				
0.2	0.200	0.182	-0.018	872	0.0 %	872	0			
2.5	2.500	2.488	-0.012	10935	0.0 %	10935				
5.0	5.000	5.054	0.054	22136	0.0 %	22136				
7.5	7.500	7.547	0.047	33016	0.0 %	33016				
10.0	10.000	9.940	-0.060	43458	0.0 %	43458				

LB92693 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Jan 2018 14:53:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	28	-	0.0000	7471B	PPB	18 Jan 2018 14:57:19	S	Std
0.2 - 1	50.2	872	-	0.2000	7471B	PPB	18 Jan 2018 14:59:36	S	Std
2.5 - 1	52.5	10935	-	2.5000	7471B	PPB	18 Jan 2018 15:01:53	S	Std
5.0 - 1	55	22136	-	5.0000	7471B	PPB	18 Jan 2018 15:04:11	S	Std
7.5 - 1	57.5	33016	-	7.5000	7471B	PPB	18 Jan 2018 15:06:28	S	Std
10.0 - 1	50	43458	-	10.0000	7471B	PPB	18 Jan 2018 15:08:46	S	Std
ICV54 - 1	ICV54	16989	3.8750	-	7471B	PPB	18 Jan 2018 15:11:30	U	SMPL
ICB54 - 1	ICB54	-103	-0.0413	-	7471B	PPB	18 Jan 2018 15:13:46	U	SMPL
CCV90 - 1	CCV90	22034	5.0309	-	7471B	PPB	18 Jan 2018 15:16:02	U	SMPL
CCB90 - 1	CCB90	-136	-0.0489	-	7471B	PPB	18 Jan 2018 15:18:18	U	SMPL
CRA - 1	CRA	839	0.1745	-	7471B	PPB	18 Jan 2018 15:20:34	U	SMPL
HighStd - 1	HighStd	43784	10.0146	-	7471B	PPB	18 Jan 2018 15:22:50	U	SMPL
ChkStd - 1	ChkStd	30293	6.9233	-	7471B	PPB	18 Jan 2018 15:25:06	U	SMPL
PB105741BL - 1	PBS	-146	-0.0512	-	7471B	PPB	18 Jan 2018 15:27:24	U	SMPL
PB105741BS - 1	LCSS	16224	3.6997	-	7471B	PPB	18 Jan 2018 15:29:42	U	SMPL
J1010-06 - 1	BU-03-011718	2410	0.5345	-	7471B	PPB	18 Jan 2018 15:32:00	U	SMPL
J1010-06DUP - 1	BU-03-011718DUP	2095	0.4623	-	7471B	PPB	18 Jan 2018 15:34:16	U	SMPL
J1010-06MS - 1	BU-03-011718MS	18397	4.1976	-	7471B	PPB	18 Jan 2018 15:36:34	U	SMPL
J1010-06MSD - 1	BU-03-011718MSD	16866	3.8468	-	7471B	PPB	18 Jan 2018 15:38:52	U	SMPL
J1023-03 - 1	TR-04-011718	941	0.1979	-	7471B	PPB	18 Jan 2018 15:41:09	U	SMPL
CCV91 - 1	CCV91	22160	5.0598	-	7471B	PPB	18 Jan 2018 15:43:24	U	SMPL
CCB91 - 1	CCB91	-116	-0.0443	-	7471B	PPB	18 Jan 2018 15:45:40	U	SMPL
J1165-01 - 1	19A-4	1031	0.2185	-	7471B	PPB	18 Jan 2018 15:47:57	U	SMPL
J1165-02 - 1	19B-4	3181	0.7111	-	7471B	PPB	18 Jan 2018 15:50:13	U	SMPL
J1166-01 - 1	2A-15	1719	0.3761	-	7471B	PPB	18 Jan 2018 15:52:29	U	SMPL
J1166-02 - 1	18A-11	1396	0.3021	-	7471B	PPB	18 Jan 2018 15:54:45	U	SMPL
J1168-01 - 1	SB-44-2.0-2.5	6588	1.4918	-	7471B	PPB	18 Jan 2018 15:57:01	U	SMPL
J1168-03 - 1	SB-47-4.5-5.0	7486	1.6975	-	7471B	PPB	18 Jan 2018 15:59:18	U	SMPL
J1168-05 - 1	SB-57-10.5-11.0	106	0.0066	-	7471B	PPB	18 Jan 2018 16:01:35	U	SMPL
J1168-06 - 1	SB-57-0-2.0	10498	2.3877	-	7471B	PPB	18 Jan 2018 16:03:53	U	SMPL
J1168-08 - 1	SB-57-0-2.0-DUP	10451	2.3769	-	7471B	PPB	18 Jan 2018 16:06:10	U	SMPL
J1168-09 - 1	SB-57-10.5-11.0-DUP	134	0.0130	-	7471B	PPB	18 Jan 2018 16:08:26	U	SMPL
CCV92 - 1	CCV92	22088	5.0433	-	7471B	PPB	18 Jan 2018 16:10:42	U	SMPL
CCB92 - 1	CCB92	-131	-0.0478	-	7471B	PPB	18 Jan 2018 16:12:58	U	SMPL
J1171-01 - 1	VNJ-232	10980	2.4981	-	7471B	PPB	18 Jan 2018 16:15:15	U	SMPL
J1181-01 - 1	HAVERSTRAW-SE	-32	-0.0251	-	7471B	PPB	18 Jan 2018 16:17:31	U	SMPL
J1184-01 - 1	SCWC	501	0.0971	-	7471B	PPB	18 Jan 2018 16:19:47	U	SMPL
J1010-06LX5 - 1		433	0.0815	-	7471B	PPB	18 Jan 2018 16:22:03	U	SMPL
J1010-06A - 1		19877	4.5367	-	7471B	PPB	18 Jan 2018 16:24:20	U	SMPL
CCV93 - 1	CCV93	22071	5.0394	-	7471B	PPB	18 Jan 2018 16:26:37	U	SMPL
CCB93 - 1	CCB93	-132	-0.0480	-	7471B	PPB	18 Jan 2018 16:28:54	U	SMPL