



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92694

Temp.: 18.5 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/18/2018	14:45
2	CAL2	1	Water			6.99	01/18/2018	14:47
3	CAL3	1	Water			10.01	01/18/2018	14:49
4	ICV	1	Water			7.00	01/18/2018	14:51
5	CCV1	1	Water			2.01	01/18/2018	14:53
6	J1110-01	1	Water			6.85	01/18/2018	14:55
7	J1110-01DUP	1	Water			6.86	01/18/2018	14:57
8	J1110-02	1	Water			6.65	01/18/2018	14:59
9	J1110-03	1	Water			7.47	01/18/2018	15:01
10	J1198-01	1	Water			3.78	01/18/2018	15:03
11	J1198-02	1	Water			3.94	01/18/2018	15:05
12	J1198-03	1	Water			4.52	01/18/2018	15:07
13	CCV2	1	Water			12.01	01/18/2018	15:10

WORKLIST(Hardcopy Internal Chain)

LB 92694

Worklist Name : pH J1198

Worklist ID : 107560

Date : 1/18/2018 2:29:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/10/2018	Water	J1110-01	pH	Cool 4 deg C	THEP04	K22	ENV-106-1(MW)DRUMPW	01/10/2018	SM 4500-PH B
01/11/2018	Water	J1110-02	pH	Cool 4 deg C	THEP04	K22	ENV-116-6(MW)DRUMPW	01/11/2018	SM 4500-PH B
01/11/2018	Water	J1110-03	pH	Cool 4 deg C	THEP04	K22	ENV-125-11(MW)DRUMPW	01/11/2018	SM 4500-PH B
01/17/2018	Water	J1198-01	pH	Cool 4 deg C	STER01	L11	CELL-A-POOL-WATER(DE-IC	01/17/2018	SM 4500-PH B
01/17/2018	Water	J1198-02	pH	Cool 4 deg C	STER01	L11	CELL-B-C-POOL-WATER(DE-	01/17/2018	SM 4500-PH B
01/17/2018	Water	J1198-03	pH	Cool 4 deg C	STER01	L11	CELL-D-POOL-WATER(DE-IC	01/17/2018	SM 4500-PH B

Date/Time

01/18/18 2:40 PM

Received by:

SP

Relinquished by:

CP

Date/Time

01/18/18

Received by:

CP

Relinquished by:

SP