



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB92695

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	01/18/2018	09:00
2	J1110-01		1	>100.0	>212.0	01/18/2018	09:35
3	J1110-01DUP		1	>100.0	>212.0	01/18/2018	10:02
4	J1110-02		1	>100.0	>212.0	01/18/2018	10:35
5	J1110-03		1	>100.0	>212.0	01/18/2018	11:05
6	J1173-02		1	>100.0	>212.0	01/18/2018	11:40
7	J1183-02		1	>100.0	>212.0	01/18/2018	12:05
8	J1185-04		1	>100.0	>212.0	01/18/2018	12:30

Result = (Celsius * 1.8) + 32

LB 92695

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP011818

WorkList ID : 107528

Date : 1/18/2018 8:20:02 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2018	Water	J1110-01	Flash Point	Cool 4 deg C	THEP04	K22	ENV-106-1(MW)DRUMPW	01/10/2018	1010A
01/18/2018	Water	J1110-02	Flash Point	Cool 4 deg C	THEP04	K22	ENV-116-6(MW)DRUMPW	01/11/2018	1010A
01/18/2018	Water	J1110-03	Flash Point	Cool 4 deg C	THEP04	K22	ENV-125-11(MW)DRUMPW	01/11/2018	1010A
01/21/2018	Water	J1173-02	Flash Point	Cool 4 deg C	PSEG01	L52	UST-1-2	01/17/2018	1010A
01/24/2018	Water	J1183-02	Flash Point	Cool 4 deg C	PSEG01	L52	4473-4476	01/17/2018	1010A
01/24/2018	Water	J1185-04	Flash Point	Cool 4 deg C	PSEG01	L52	15182	01/17/2018	1010A

Date/Time 1/18/18 8:25 AM
 Received by: KS
 Relinquished by: CP

Date/Time 1/18/18 12:35 PM
 Received by: CP
 Relinquished by: KS