



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB92708
Analysis Date: 01/19/2018
BalanceID: WC SC-3
OvenID: EX Oven-1

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 01/19/2018
Extraction IN Time: 08:10
Extraction OUT Time: 08:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB92708BL	LB92708BL	WATER	2	1000	100	2.8741	2.8741	3.02	2.8741	2.8741	0	0
2	LB92708BS	LB92708BS	WATER	2	1000	100	2.9633	2.9633	3.01	2.9818	2.9818	0.0185	18.5
3	LB92708BSD	LB92708BSD	WATER	2	1000	100	2.8103	2.8103	3.03	2.829	2.829	0.0187	18.7
4	J1173-01	UST-1-2	WATER	2	1000	100	2.9176	2.9176	3.04	2.918	2.918	0.0004	0.4



QC Batch# LB92708 Test: TPH Analysis Date: 01/19/2018

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2303
pH Paper 0-14	W2287
Sodium Sulfate	EP1806
1:1 HCL	WP59249
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD	10.00 ML	WP58005
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0019 In OVEN TEMP : 70 Dessicator Time In : 09:45
1.0000 gram Balance: 1.0004 In Time: 09:10
Bal Check Time: 08:30

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 11:15
1.0000 gram Balance: 1.0005 Out Time: 10:45
Bal Check Time: 11:50

2392708

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J1200

WorkList ID : 107572

Date : 1/19/2018 7:59:55 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2018	Water	J1173-01-K	TPH	Conc H2SO4 to	PSEG01	L52	UST-1-2	01/17/2018	1664A

Date/Time 01-19-18 8:00AM

Received by: Jh

Relinquished by: [Signature]

Date/Time 01-19-18 4:20PM

Received by: [Signature]

Relinquished by: Jh