



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92711

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
IODINE SOLUTION .025N 1L	W2261
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/18/2018	17:30
2	CAL2	1	Water			7.00	01/18/2018	17:31
3	CAL3	1	Water			10.04	01/18/2018	17:32
4	ICV	1	Water			6.99	01/18/2018	17:34
5	CCV1	1	Water			2.01	01/18/2018	17:35
6	J1018-07	1	Solid	20.02	20	8.28	01/18/2018	17:36
7	J1018-07DUP	1	Solid	20.03	20	8.29	01/18/2018	17:37
8	J1018-09	1	Solid	20.04	20	8.14	01/18/2018	17:38
9	J1191-01	1	Solid	20.02	20	6.90	01/18/2018	17:40
10	J1202-03	1	Solid	20.04	20	7.54	01/18/2018	17:41
11	J1203-02	1	Solid	20.05	20	8.78	01/18/2018	17:42
12	CCV2	1	Water			12.02	01/18/2018	17:44

1392711

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1018

WorkList ID : 107567

Date : 1/18/2018 5:18:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2018	Solid	J1018-07	Corrosivity	Cool 4 deg C	PSEG05	L52	OR-02-011818-A	01/18/2018	9045C
01/18/2018	Solid	J1018-09	Corrosivity	Cool 4 deg C	PSEG05	L52	OR-02-011818-B	01/18/2018	9045C
01/18/2018	Solid	J1191-01	Corrosivity	Cool 4 deg C	PSEG02	L52	OILY-DEBRIS	01/18/2018	9045C
01/18/2018	Solid	J1202-03	Corrosivity	Cool 4 deg C	PSEG01	L52	ROLLOFF-1-5	01/18/2018	9045C
01/18/2018	Solid	J1203-02	Corrosivity	Cool 4 deg C	PSEG02	L52	RO-285	01/18/2018	9045C

Date/Time 01-18-18 5:20 PM
Received by: J1
Relinquished by: CD

Date/Time 01-18-18 6:35 PM
Received by: CD
Relinquished by: JP