


7470A

LB 92712
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36567



A= 0.0000e+000

B= 2.7648e-004

C= -1.0175e-002

Rho= 0.9995021

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	31	0.000	31				
0.2	0.200	0.199	-0.001	758	0.0 %	758				
2.5	2.500	2.504	0.004	9094	0.0 %	9094				
5.0	5.000	5.131	0.131	18594	0.0 %	18594				
7.5	7.500	7.267	-0.233	26321	0.0 %	26321				
10.0	10.000	10.100	0.100	36567	0.0 %	36567				

LB92712 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jan 2018 10:31:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7470A	PPB	19 Jan 2018 10:39:34	S	Std
0.2 - 1	50.2	758	-	0.2000	7470A	PPB	19 Jan 2018 10:41:42	S	Std
2.5 - 1	52.5	9094	-	2.5000	7470A	PPB	19 Jan 2018 10:43:51	S	Std
5.0 - 1	55	18594	-	5.0000	7470A	PPB	19 Jan 2018 10:45:58	S	Std
7.5 - 1	57.5	26321	-	7.5000	7470A	PPB	19 Jan 2018 10:48:06	S	Std
10.0 - 1	510	36567	-	10.0000	7470A	PPB	19 Jan 2018 10:50:14	S	Std
ICV55 - 1	ICV55	14261	3.9328	-	7470A	PPB	19 Jan 2018 10:53:28	U	SMPL
ICB55 - 1	ICB55	-97	-0.0370	-	7470A	PPB	19 Jan 2018 10:55:34	U	SMPL
CCV94 - 1	CCV94	18686	5.1562	-	7470A	PPB	19 Jan 2018 10:57:41	U	SMPL
CCB94 - 1	CCB94	-103	-0.0387	-	7470A	PPB	19 Jan 2018 10:59:47	U	SMPL
CRA - 1	CRA	708	0.1856	-	7470A	PPB	19 Jan 2018 11:01:54	U	SMPL
HighStd - 1	HighStd	36589	10.1061	-	7470A	PPB	19 Jan 2018 11:04:01	U	SMPL
ChkStd - 1	ChkStd	25630	7.0761	-	7470A	PPB	19 Jan 2018 11:06:08	U	SMPL
PB105764BL - 1	PBW	-136	-0.0478	-	7470A	PPB	19 Jan 2018 11:08:17	U	SMPL
PB105764BS - 1	LCSW	13653	3.7647	-	7470A	PPB	19 Jan 2018 11:10:26	U	SMPL
PB105764TB - 1	PB105739TB	-92	-0.0356	-	7470A	PPB	19 Jan 2018 11:12:33	U	SMPL
J1010-07 - 1	BU-03-011718	8	-0.0080	-	7470A	PPB	19 Jan 2018 11:14:41	U	SMPL
J1010-07DUP - 1	BU-03-011718DUP	40	0.0009	-	7470A	PPB	19 Jan 2018 11:16:49	U	SMPL
J1010-07MS - 1	BU-03-011718MS	12966	3.5747	-	7470A	PPB	19 Jan 2018 11:18:58	U	SMPL
J1010-07MSD - 1	BU-03-011718MSD	13128	3.6195	-	7470A	PPB	19 Jan 2018 11:21:04	U	SMPL
CCV95 - 1	CCV95	18844	5.1999	-	7470A	PPB	19 Jan 2018 11:23:11	U	SMPL
CCB95 - 1	CCB95	-91	-0.0353	-	7470A	PPB	19 Jan 2018 11:25:17	U	SMPL
J1023-04 - 1	TR-04-011718	1	-0.0099	-	7470A	PPB	19 Jan 2018 11:27:24	U	SMPL
J1159-01 - 1	FD-1	41	0.0012	-	7470A	PPB	19 Jan 2018 11:29:31	U	SMPL
J1168-02 - 1	SB-44-COMP	21	-0.0044	-	7470A	PPB	19 Jan 2018 11:31:38	U	SMPL
J1168-04 - 1	SB-47-COMP	58	0.0059	-	7470A	PPB	19 Jan 2018 11:33:45	U	SMPL
J1168-07 - 1	SB-57-COMP	91	0.0150	-	7470A	PPB	19 Jan 2018 11:35:53	U	SMPL
J1168-10 - 1	SB-57-COMP-DUP	64	0.0075	-	7470A	PPB	19 Jan 2018 11:38:00	U	SMPL
J1171-02 - 1	VNJ-232	30	-0.0019	-	7470A	PPB	19 Jan 2018 11:40:08	U	SMPL
J1173-02 - 1	UST-1-2	20	-0.0046	-	7470A	PPB	19 Jan 2018 11:42:16	U	SMPL
J1183-01 - 1	4469-73	76	0.0108	-	7470A	PPB	19 Jan 2018 11:44:25	U	SMPL
J1183-02 - 1	4473-4476	48	0.0031	-	7470A	PPB	19 Jan 2018 11:46:32	U	SMPL
CCV96 - 1	CCV96	18756	5.1756	-	7470A	PPB	19 Jan 2018 11:48:39	U	SMPL
CCB96 - 1	CCB96	-87	-0.0342	-	7470A	PPB	19 Jan 2018 11:50:46	U	SMPL
J1184-02 - 1	SCWC	24	-0.0035	-	7470A	PPB	19 Jan 2018 11:52:54	U	SMPL
J1185-01 - 1	15180	17	-0.0055	-	7470A	PPB	19 Jan 2018 11:55:01	U	SMPL
J1185-02 - 1	15192	-20	-0.0157	-	7470A	PPB	19 Jan 2018 11:57:08	U	SMPL
J1185-03 - 1	15183-90	59	0.0061	-	7470A	PPB	19 Jan 2018 11:59:15	U	SMPL
J1185-04 - 1	15182	6	-0.0085	-	7470A	PPB	19 Jan 2018 12:01:23	U	SMPL
PB105766BL - 1	PBW	14	-0.0063	-	7470A	PPB	19 Jan 2018 12:03:31	U	SMPL
PB105766BS - 1	LCSW	13605	3.7514	-	7470A	PPB	19 Jan 2018 12:05:38	U	SMPL
J1173-01 - 1	UST-1-2	-223	-0.0718	-	7470A	PPB	19 Jan 2018 12:07:47	U	SMPL
J1173-01DUP - 1	UST-1-2DUP	-157	-0.0536	-	7470A	PPB	19 Jan 2018 12:09:55	U	SMPL
J1173-01MS - 1	UST-1-2MS	13057	3.5999	-	7470A	PPB	19 Jan 2018 12:12:02	U	SMPL
CCV97 - 1	CCV97	18488	5.1015	-	7470A	PPB	19 Jan 2018 12:14:09	U	SMPL
CCB97 - 1	CCB97	-64	-0.0279	-	7470A	PPB	19 Jan 2018 12:16:16	U	SMPL
J1173-01MSD - 1	UST-1-2MSD	13198	3.6389	-	7470A	PPB	19 Jan 2018 12:18:23	U	SMPL
J1189-08 - 1	TW-1	5856	1.6089	-	7470A	PPB	19 Jan 2018 12:20:30	U	SMPL
J1010-07LX5 - 1		-44	-0.0223	-	7470A	PPB	19 Jan 2018 12:22:38	U	SMPL
J1010-07A - 1		14071	3.8802	-	7470A	PPB	19 Jan 2018 12:24:45	U	SMPL
J1173-01LX5 - 1		-86	-0.0340	-	7470A	PPB	19 Jan 2018 12:26:53	U	SMPL
J1173-01A - 1		14569	4.0179	-	7470A	PPB	19 Jan 2018 12:29:01	U	SMPL
CCV98 - 1	CCV98	18730	5.1684	-	7470A	PPB	19 Jan 2018 12:31:08	U	SMPL
CCB98 - 1	CCB98	-92	-0.0356	-	7470A	PPB	19 Jan 2018 12:33:17	U	SMPL