



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92714

Temp.: 10.0 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/19/2018	10:05
2	CAL2	1	Water			7.00	01/19/2018	10:07
3	CAL3	1	Water			10.01	01/19/2018	10:09
4	ICV	1	Water			7.00	01/19/2018	10:11
5	CCV1	1	Water			2.01	01/19/2018	10:13
6	J1204-01	1	Water			6.29	01/19/2018	10:15
7	J1204-01DUP	1	Water			6.30	01/19/2018	10:17
8	J1204-02	1	Water			6.50	01/19/2018	10:19
9	J1204-03	1	Water			6.57	01/19/2018	10:21
10	J1204-04	1	Water			6.29	01/19/2018	10:23
11	J1204-05	1	Water			6.65	01/19/2018	10:25
12	J1204-06	1	Water			6.70	01/19/2018	10:27
13	J1204-07	1	Water			6.65	01/19/2018	10:29
14	J1204-08	1	Water			6.71	01/19/2018	10:31
15	CCV2	1	Water			12.01	01/19/2018	10:33

WORKLIST(Hardcopy Internal Chain)

LB 92714

Worklist Name : pH J1204

Worklist ID : 107592

Date : 1/19/2018 9:57:49 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/17/2018	Water	J1204-01	pH	Cool 4 deg C	DIST01	L23	1-BSMT-CITY-WATER	01/17/2018	SM 4500-PH B
01/17/2018	Water	J1204-02	pH	Cool 4 deg C	DIST01	L23	2-FL7-ZN2-HOT-DRAIN	01/17/2018	SM 4500-PH B
01/17/2018	Water	J1204-03	pH	Cool 4 deg C	DIST01	L23	3-14BE-MBBTH-TUB	01/17/2018	SM 4500-PH B
01/17/2018	Water	J1204-04	pH	Cool 4 deg C	DIST01	L23	4-14BE-MBBTH-SHWR	01/17/2018	SM 4500-PH B
01/17/2018	Water	J1204-05	pH	Cool 4 deg C	DIST01	L23	5-14BE-KITCHEN-SINK	01/17/2018	SM 4500-PH B
01/17/2018	Water	J1204-06	pH	Cool 4 deg C	DIST01	L23	6-ROOF-PUMP-RM-FAUCET	01/17/2018	SM 4500-PH B
01/17/2018	Water	J1204-07	pH	Cool 4 deg C	DIST01	L23	7-14BE-KITCHEN-SINK	01/17/2018	SM 4500-PH B
01/17/2018	Water	J1204-08	pH	Cool 4 deg C	DIST01	L23	8-14BE-MBBTH-TUB	01/17/2018	SM 4500-PH B

Date/Time 01/19/18 10:00 AM
 Received by: SP
 Relinquished by: SP

Date/Time 01/19/18 10:45 AM
 Received by: SP
 Relinquished by: SP