


7471B

LB 92720
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44394



A= 0.0000e+000

B= 2.2874e-004

C= -2.7038e-002

Rho= 0.9996129

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	68	0.000	68				
0.2	0.200	0.208	0.008	1026	0.0 %	1026	0			
2.5	2.500	2.602	0.102	11495	0.0 %	11495				
5.0	5.000	4.960	-0.040	21804	0.0 %	21804				
7.5	7.500	7.314	-0.186	32092	0.0 %	32092				
10.0	10.000	10.128	0.128	44394	0.0 %	44394				

LB92720 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 19 Jan 2018 14:13:24					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	68	-	0.0000	7471B	PPB	19 Jan 2018 14:22:58	S	Std
0.2 - 1	SD.2	1026	-	0.2000	7471B	PPB	19 Jan 2018 14:25:16	S	Std
2.5 - 1	SD.5	11495	-	2.5000	7471B	PPB	19 Jan 2018 14:27:33	S	Std
5.0 - 1	SD	21804	-	5.0000	7471B	PPB	19 Jan 2018 14:29:51	S	Std
7.5 - 1	SD.5	32092	-	7.5000	7471B	PPB	19 Jan 2018 14:32:07	S	Std
10.0 - 1	SD	44394	-	10.0000	7471B	PPB	19 Jan 2018 14:37:05	S	Std
ICV56 - 1	ICV56	17382	3.9489	-	7471B	PPB	19 Jan 2018 14:39:44	U	SMPL
ICB56 - 1	ICB56	-124	-0.0554	-	7471B	PPB	19 Jan 2018 14:42:01	U	SMPL
CCV01 - 1	CCV01	21735	4.9446	-	7471B	PPB	19 Jan 2018 14:44:16	U	SMPL
CCB01 - 1	CCB01	-125	-0.0556	-	7471B	PPB	19 Jan 2018 14:46:32	U	SMPL
CRA - 1	CRA	1029	0.2083	-	7471B	PPB	19 Jan 2018 14:48:49	U	SMPL
HighStd - 1	HighStd	44626	10.1806	-	7471B	PPB	19 Jan 2018 14:51:05	U	SMPL
ChkStd - 1	ChkStd	30760	7.0089	-	7471B	PPB	19 Jan 2018 14:53:21	U	SMPL
PB105763BL - 1	PBS	-156	-0.0627	-	7471B	PPB	19 Jan 2018 14:55:40	U	SMPL
PB105763BS - 1	LCSS	14945	3.3914	-	7471B	PPB	19 Jan 2018 14:57:58	U	SMPL
J1018-06 - 1	OR-02-011818-A	5575	1.2482	-	7471B	PPB	19 Jan 2018 15:00:14	U	SMPL
J1018-08 - 1	OR-02-011818-B	2222	0.4812	-	7471B	PPB	19 Jan 2018 15:02:32	U	SMPL
J1187-01 - 1	EPE-106-4A(5-10)	18345	4.1692	-	7471B	PPB	19 Jan 2018 15:04:49	U	SMPL
J1187-02 - 1	EPE-106-4A(34-36)	103	-0.0035	-	7471B	PPB	19 Jan 2018 15:07:07	U	SMPL
J1187-03 - 1	EPE-106-4A(37-39)	154	0.0082	-	7471B	PPB	19 Jan 2018 15:09:23	U	SMPL
CCV02 - 1	CCV02	21862	4.9736	-	7471B	PPB	19 Jan 2018 15:11:38	U	SMPL
CCB02 - 1	CCB02	-133	-0.0575	-	7471B	PPB	19 Jan 2018 15:13:54	U	SMPL
J1187-04 - 1	EPE-106-4A(40-42)	235	0.0267	-	7471B	PPB	19 Jan 2018 15:16:10	U	SMPL
J1187-05 - 1	EPE-106-4A(46-48)	192	0.0169	-	7471B	PPB	19 Jan 2018 15:18:25	U	SMPL
J1187-06 - 1	EPE-106-4A(55-57)	534	0.0951	-	7471B	PPB	19 Jan 2018 15:20:41	U	SMPL
J1187-07 - 1	EPE-106-4A(65-67)	547	0.0981	-	7471B	PPB	19 Jan 2018 15:22:57	U	SMPL
J1187-08 - 1	EPE-106-4A(70-72)	1167	0.2399	-	7471B	PPB	19 Jan 2018 15:25:13	U	SMPL
J1187-09 - 1	EPE-125-9(0-5)	8199	1.8484	-	7471B	PPB	19 Jan 2018 15:27:30	U	SMPL
J1187-10 - 1	EPE-125-9(5-10)	9361	2.1142	-	7471B	PPB	19 Jan 2018 15:29:48	U	SMPL
J1189-01 - 1	33C-4	25429	5.7895	-	7471B	PPB	19 Jan 2018 15:32:05	U	SMPL
J1189-04 - 1	22C-2	18550	4.2160	-	7471B	PPB	19 Jan 2018 15:44:16	U	SMPL
J1189-04DUP - 1	22C-2DUP	18241	4.1454	-	7471B	PPB	19 Jan 2018 15:46:32	U	SMPL
CCV03 - 1	CCV03	22044	5.0153	-	7471B	PPB	19 Jan 2018 15:48:48	U	SMPL
CCB03 - 1	CCB03	-324	-0.1011	-	7471B	PPB	19 Jan 2018 15:51:04	U	SMPL
J1189-04MS - 1	22C-2MS	36026	8.2135	-	7471B	PPB	19 Jan 2018 16:01:54	U	SMPL
J1189-04MSD - 1	22C-2MSD	36783	8.3866	-	7471B	PPB	19 Jan 2018 16:04:10	U	SMPL
J1189-05 - 1	2B-2	13835	3.1376	-	7471B	PPB	19 Jan 2018 16:06:28	U	SMPL
J1189-06 - 1	4-2	4766	1.0631	-	7471B	PPB	19 Jan 2018 16:08:46	U	SMPL
J1189-07 - 1	30A-8	570	0.1033	-	7471B	PPB	19 Jan 2018 16:11:02	U	SMPL
J1202-01 - 1	ROLLOFF-1-5	9783	2.2107	-	7471B	PPB	19 Jan 2018 16:13:19	U	SMPL
J1203-01 - 1	RO-285	21259	4.8357	-	7471B	PPB	19 Jan 2018 16:15:36	U	SMPL
J1189-04LX5 - 1		3498	0.7731	-	7471B	PPB	19 Jan 2018 16:17:54	U	SMPL
J1189-04A - 1		36429	8.3057	-	7471B	PPB	19 Jan 2018 16:20:13	U	SMPL
CCV04 - 1	CCV04	21669	4.9295	-	7471B	PPB	19 Jan 2018 16:22:29	U	SMPL
CCB04 - 1	CCB04	-132	-0.0572	-	7471B	PPB	19 Jan 2018 16:24:46	U	SMPL