


ISM02.4

LB92734
INSTRUMENT ID: CV1

Linear

 μ Abs.:

46475



A= 0.0000e+000

B= 2.1690e-004

C= -3.8456e-002

Rho= 0.9997475

Accept=Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o D
0.00	0.000	-0.026	-0.026	59	0.000	59					-
0.2	0.200	0.182	-0.018	1016	0.0 %	1016					-9
2.5	2.500	2.535	0.035	11863	0.0 %	11863					1
5.0	5.000	5.119	0.119	23778	0.0 %	23778					2
7.5	7.500	7.348	-0.152	34054	0.0 %	34054					-2
10.0	10.000	10.042	0.042	46475	0.0 %	46475					0

NB
1/19/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	59	-		0 ISM02.4	PPB	1/19/2018 17:00	Std
	0.2 S0.2	1016	-		0.2 ISM02.4	PPB	1/19/2018 17:02	Std
	2.5 S2.5	11863	-		2.5 ISM02.4	PPB	1/19/2018 17:07	Std
	5 S5	23778	-		5 ISM02.4	PPB	1/19/2018 17:09	Std
	7.5 S7.5	34054	-		7.5 ISM02.4	PPB	1/19/2018 17:11	Std
	10 S10	46475	-		10 ISM02.4	PPB	1/19/2018 17:14	Std
ICV	ICV	17843	3.8317 -		ISM02.4	PPB	1/19/2018 17:16	SMPL
ICB	ICB	-70	-0.0536 -		ISM02.4	PPB	1/19/2018 17:19	SMPL
CCV005	CCV005	23538	5.067 -		ISM02.4	PPB	1/19/2018 17:21	SMPL
CCB005	CCB005	-62	-0.0519 -		ISM02.4	PPB	1/19/2018 17:23	SMPL
PB105765BL	PBS004	102	-0.0163 -		ISM02.4	PPB	1/19/2018 17:25	SMPL
J1175-01	MF6A00	1736	0.3381 -		ISM02.4	PPB	1/19/2018 17:28	SMPL
J1175-02	MF6A00D	1716	0.3337 -		ISM02.4	PPB	1/19/2018 17:30	SMPL
J1175-03	MF6A00S	13708	2.9348 -		ISM02.4	PPB	1/19/2018 17:32	SMPL
J1175-04	MF6A01	1327	0.2494 -		ISM02.4	PPB	1/19/2018 17:34	SMPL
J1175-05	MF6A04	1070	0.1936 -		ISM02.4	PPB	1/19/2018 17:37	SMPL
J1175-07	MF6A57	340	0.0353 -		ISM02.4	PPB	1/19/2018 17:39	SMPL
J1175-08	MF6A58	308	0.0283 -		ISM02.4	PPB	1/19/2018 17:41	SMPL
J1175-09	MF6A59	374	0.0427 -		ISM02.4	PPB	1/19/2018 17:44	SMPL
J1175-10	MF6A60	173	-0.0009 -		ISM02.4	PPB	1/19/2018 17:46	SMPL
J1175-11	MF6A61	325	0.032 -		ISM02.4	PPB	1/19/2018 17:48	SMPL
J1175-12	MF6A62	235	0.0125 -		ISM02.4	PPB	1/19/2018 17:50	SMPL
J1175-13	MF6A46	1860	0.365 -		ISM02.4	PPB	1/19/2018 17:53	SMPL
J1175-14	MF6A47	17481	3.7532 -		ISM02.4	PPB	1/19/2018 17:55	SMPL
J1175-15	MF6A48	20905	4.4959 -		ISM02.4	PPB	1/19/2018 17:57	SMPL
J1175-16	MF6A49	1087	0.1973 -		ISM02.4	PPB	1/19/2018 18:00	SMPL
J1175-17	MF6A50	1060	0.1915 -		ISM02.4	PPB	1/19/2018 18:02	SMPL
J1175-18	MF6A51	951	0.1678 -		ISM02.4	PPB	1/19/2018 18:04	SMPL
J1175-19	MF6A52	168	-0.002 -		ISM02.4	PPB	1/19/2018 18:06	SMPL
CCV006	CCV006	23707	5.1037 -		ISM02.4	PPB	1/19/2018 18:09	SMPL
CCB006	CCB006	-86	-0.0571 -		ISM02.4	PPB	1/19/2018 18:11	SMPL
J1175-20	MF6A53	932	0.1637 -		ISM02.4	PPB	1/19/2018 18:13	SMPL
J1175-21	MF6A54	368	0.0414 -		ISM02.4	PPB	1/19/2018 18:15	SMPL
J1175-22	MF6A55	277	0.0216 -		ISM02.4	PPB	1/19/2018 18:18	SMPL
J1175-06	MF6A56	400	0.0483 -		ISM02.4	PPB	1/19/2018 18:20	SMPL
J1175-23	MF6A56D	408	0.05 -		ISM02.4	PPB	1/19/2018 18:22	SMPL
J1175-24	MF6A56S	12935	2.7672 -		ISM02.4	PPB	1/19/2018 18:25	SMPL
PB105800BL	PBS004	3	-0.0378 -		ISM02.4	PPB	1/19/2018 18:27	SMPL
J1217-01	MB0AE4	954	0.1685 -		ISM02.4	PPB	1/19/2018 18:29	SMPL
J1217-02	MB0AE5	592	0.09 -		ISM02.4	PPB	1/19/2018 18:31	SMPL
J1217-03	MB0AE6	694	0.1121 -		ISM02.4	PPB	1/19/2018 18:34	SMPL
J1217-04	MB0AE7	6686	1.4118 -		ISM02.4	PPB	1/19/2018 18:36	SMPL
J1217-05	MB0AE8	463	0.062 -		ISM02.4	PPB	1/19/2018 18:38	SMPL
J1217-06	MB0AE8D	522	0.0748 -		ISM02.4	PPB	1/19/2018 18:41	SMPL
J1217-07	MB0AE8S	12494	2.6715 -		ISM02.4	PPB	1/19/2018 18:43	SMPL
CCV007	CCV007	23559	5.0716 -		ISM02.4	PPB	1/19/2018 18:45	SMPL

LB92734

INSTRUMENT ID : CV1

CCB007

CCB007

-32 -0.0454 -

ISM02.4 PPB

1/19/2018 18:47 SMPL