



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92740

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
IODINE SOLUTION .025N 1L	W2261
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	01/19/2018	18:00
2	CAL2	1	Water			7.01	01/19/2018	18:01
3	CAL3	1	Water			10.02	01/19/2018	18:02
4	ICV	1	Water			7.00	01/19/2018	18:03
5	CCV1	1	Water			2.01	01/19/2018	18:05
6	J1186-01	1	Solid	20.02	20	7.27	01/19/2018	18:06
7	J1186-01DUP	1	Solid	20.03	20	7.28	01/19/2018	18:09
8	J1186-02	1	Solid	20.04	20	8.53	01/19/2018	18:10
9	J1186-03	1	Solid	20.01	20	6.86	01/19/2018	18:11
10	J1186-04	1	Solid	20.02	20	6.65	01/19/2018	18:12
11	J1186-05	1	Solid	20.05	20	8.26	01/19/2018	18:14
12	J1186-06	1	Solid	20.03	20	6.90	01/19/2018	18:15
13	J1188-01	1	Solid	20.02	20	7.58	01/19/2018	18:16
14	J1188-02	1	Solid	20.02	20	7.15	01/19/2018	18:19
15	J1210-03	1	Solid	20.02	20	8.55	01/19/2018	18:21
16	CCV2	1	Water			12.02	01/19/2018	18:22
17	J1210-06	1	Solid	20.04	20	7.63	01/19/2018	18:24
18	J1210-09	1	Solid	20.02	20	8.34	01/19/2018	18:25
19	J1210-12	1	Solid	20.03	20	7.91	01/19/2018	18:26
20	J1212-04	1	Solid	20.01	20	8.88	01/19/2018	18:27

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	J1214-03	1	Solid	20.03	20	6.93	01/19/2018	18:29
22	J1214-04	1	Solid	20.01	20	7.08	01/19/2018	18:30
23	J1218-02	1	Solid	20.02	20	7.88	01/19/2018	18:31
24	J1229-03	1	Solid	20.04	20	5.90	01/19/2018	18:32
25	CCV3	1	Water			12.01	01/19/2018	18:33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1186

WorkList ID : 107608

Date : 1/19/2018 2:05:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/16/2018	Solid	J1186-01	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-1	01/16/2018	9045C
01/16/2018	Solid	J1186-02	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-2	01/16/2018	9045C
01/16/2018	Solid	J1186-03	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-3	01/16/2018	9045C
01/16/2018	Solid	J1186-04	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-4	01/16/2018	9045C
01/16/2018	Solid	J1186-05	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-5	01/16/2018	9045C
01/16/2018	Solid	J1186-06	Corrosivity	Cool 4 deg C	CLEA11	L42	OWS-6	01/16/2018	9045C
01/16/2018	Solid	J1188-01	Corrosivity	Cool 4 deg C	CLEA11		OWS-1	01/16/2018	9045C
01/16/2018	Solid	J1188-02	Corrosivity	Cool 4 deg C	CLEA11		OWS-2	01/16/2018	9045C
01/19/2018	Solid	J1210-03	Corrosivity	Cool 4 deg C	PSEG01		RT-3992	01/19/2018	9045C
01/19/2018	Solid	J1210-06	Corrosivity	Cool 4 deg C	PSEG01	L52	RT-4248	01/19/2018	9045C
01/19/2018	Solid	J1210-09	Corrosivity	Cool 4 deg C	PSEG01	L52	R-4220	01/19/2018	9045C
01/19/2018	Solid	J1210-12	Corrosivity	Cool 4 deg C	PSEG01	L52	RT-2898	01/19/2018	9045C
01/19/2018	Solid	J1212-04	Corrosivity	Cool 4 deg C	PSEG02	L52	OILY-STONES	01/19/2018	9045C
01/18/2018	Solid	J1214-03	Corrosivity	Cool 4 deg C	VIA01	M13	WC-PH3-MGP-01	01/18/2018	9045C
01/18/2018	Solid	J1214-04	Corrosivity	Cool 4 deg C	VIA01	M13	WC-PH3-MGP-02	01/18/2018	9045C
01/19/2018	Solid	J1218-02	Corrosivity	Cool 4 deg C	PSEG02	M52	OILY-GRAVEL	01/19/2018	9045C
01/18/2018	Solid	J1229-03	Corrosivity	Cool 4 deg C	TECT01	M22	TEL-COMP-1	01/18/2018	9045C

Date/Time 01-19-18 2:10 PM
 Received by: TV
 Relinquished by: CP

Date/Time 01-19-18 7:00 PM
 Received by: SD
 Relinquished by: JP