



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92742

Temp.: 22.1 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
IODINE SOLUTION .025N 1L	W2261
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	01/19/2018	18:00
2	CAL2	1	Water			7.01	01/19/2018	18:01
3	CAL3	1	Water			10.02	01/19/2018	18:02
4	ICV	1	Water			7.00	01/19/2018	18:03
5	CCV1	1	Water			2.02	01/19/2018	18:38
6	J1213-03	1	Water			7.53	01/19/2018	18:39
7	J1213-03DUP	1	Water			7.54	01/19/2018	18:40
8	J1230-03	1	Water			6.81	01/19/2018	18:41
9	CCV2	1	Water			12.01	01/19/2018	18:43

2392742

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 11213

WorkList ID : 107623

Date : 1/19/2018 6:03:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/19/2018	Water	J1213-03-17	pH	Cool 4 deg C	PSEG01	L52	MANHOLE	01/19/2018	SM 4500-PH B
01/19/2018	Water	J1230-03-17	pH	Cool 4 deg C	PSEG02	M52	EXCAVATING	01/19/2018	SM 4500-PH B

Date/Time 01-19-18 6:05 PM
Received by: JH
Relinquished by: tR

Date/Time 01-19-18 6:00 PM
Received by: CH
Relinquished by: JH