


7470A

LB 92753
INSTRUMENT ID: CV2

Linear

 μ Abs.:

36419



A= 0.0000e+000

B= 2.7304e-004

C= 1.6086e-002

Rho= 0.9999486

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	-134	0.000	-134				
0.2	0.200	0.226	0.026	768	0.0 %	768				
2.5	2.500	2.486	-0.014	9045	0.0 %	9045				
5.0	5.000	4.978	-0.022	18173	0.0 %	18173				
7.5	7.500	7.571	0.071	27669	0.0 %	27669				
10.0	10.000	9.960	-0.040	36419	0.0 %	36419				

LB92753 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Jan 2018 10:27:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-134	-	0.0000	7470A	PPB	22 Jan 2018 11:21:11	S	Std
0.2 - 1	50.2	768	-	0.2000	7470A	PPB	22 Jan 2018 11:23:19	S	Std
2.5 - 1	52.5	9045	-	2.5000	7470A	PPB	22 Jan 2018 11:25:27	S	Std
5.0 - 1	55	18173	-	5.0000	7470A	PPB	22 Jan 2018 11:27:35	S	Std
7.5 - 1	57.5	27669	-	7.5000	7470A	PPB	22 Jan 2018 11:29:43	S	Std
10.0 - 1	510	36419	-	10.0000	7470A	PPB	22 Jan 2018 11:31:50	S	Std
ICV57 - 1	ICV57	14146	3.8785	-	7470A	PPB	22 Jan 2018 11:34:32	U	SMPL
ICB57 - 1	ICB57	-245	-0.0508	-	7470A	PPB	22 Jan 2018 11:36:38	U	SMPL
CCV08 - 1	CCV08	18336	5.0226	-	7470A	PPB	22 Jan 2018 11:38:44	U	SMPL
CCB08 - 1	CCB08	-243	-0.0503	-	7470A	PPB	22 Jan 2018 11:40:51	U	SMPL
CRA - 1	CRA	780	0.2291	-	7470A	PPB	22 Jan 2018 11:42:58	U	SMPL
HighStd - 1	HighStd	36611	10.0124	-	7470A	PPB	22 Jan 2018 11:45:05	U	SMPL
ChkStd - 1	ChkStd	24585	6.7288	-	7470A	PPB	22 Jan 2018 11:47:11	U	SMPL
PB105801BL - 1	PBW	-151	-0.0251	-	7470A	PPB	22 Jan 2018 11:49:20	U	SMPL
PB105801BS - 1	LCSW	13741	3.7680	-	7470A	PPB	22 Jan 2018 11:51:28	U	SMPL
PB105801TB - 1	PB105768TB	-74	-0.0041	-	7470A	PPB	22 Jan 2018 11:53:36	U	SMPL
J1018-07 - 1	OR-02-011818-A	-66	-0.0019	-	7470A	PPB	22 Jan 2018 11:55:45	U	SMPL
J1018-07DUP - 1	OR-02-011818-ADUP	-2	0.0155	-	7470A	PPB	22 Jan 2018 11:57:53	U	SMPL
J1018-07MS - 1	OR-02-011818-AMS	13780	3.7786	-	7470A	PPB	22 Jan 2018 12:00:01	U	SMPL
J1018-07MSD - 1	OR-02-011818-AMSD	13540	3.7131	-	7470A	PPB	22 Jan 2018 12:02:08	U	SMPL
CCV09 - 1	CCV09	18221	4.9912	-	7470A	PPB	22 Jan 2018 12:04:15	U	SMPL
CCB09 - 1	CCB09	-286	-0.0620	-	7470A	PPB	22 Jan 2018 12:06:22	U	SMPL
J1018-09 - 1	OR-02-011818-B	-18	0.0112	-	7470A	PPB	22 Jan 2018 12:08:29	U	SMPL
J1191-01 - 1	OILY-DEBRIS	-82	-0.0063	-	7470A	PPB	22 Jan 2018 12:10:35	U	SMPL
J1202-03 - 1	ROLLOFF-1-5	-14	0.0123	-	7470A	PPB	22 Jan 2018 12:12:41	U	SMPL
J1203-02 - 1	RO-285	-26	0.0090	-	7470A	PPB	22 Jan 2018 12:14:48	U	SMPL
PB105802BL - 1	PBW	-57	0.0005	-	7470A	PPB	22 Jan 2018 12:16:56	U	SMPL
PB105802BS - 1	LCSW	14298	3.9200	-	7470A	PPB	22 Jan 2018 12:19:03	U	SMPL
PB105802TB - 1	PB105767TB	-125	-0.0180	-	7470A	PPB	22 Jan 2018 12:21:12	U	SMPL
J1208-01 - 1	WC-9871-01	-81	-0.0060	-	7470A	PPB	22 Jan 2018 12:23:20	U	SMPL
J1208-01DUP - 1	WC-9871-01DUP	-24	0.0095	-	7470A	PPB	22 Jan 2018 12:25:28	U	SMPL
J1208-01MS - 1	WC-9871-01MS	13595	3.7281	-	7470A	PPB	22 Jan 2018 12:27:35	U	SMPL
CCV10 - 1	CCV10	18232	4.9942	-	7470A	PPB	22 Jan 2018 12:29:42	U	SMPL
CCB10 - 1	CCB10	-273	-0.0585	-	7470A	PPB	22 Jan 2018 12:31:49	U	SMPL
J1208-01MSD - 1	WC-9871-01MSD	13564	3.7196	-	7470A	PPB	22 Jan 2018 12:33:57	U	SMPL
J1208-02 - 1	WC-7317-01	-144	-0.0232	-	7470A	PPB	22 Jan 2018 12:36:04	U	SMPL
J1208-03 - 1	WC-7317-02	-59	-0.0000	-	7470A	PPB	22 Jan 2018 12:38:11	U	SMPL
J1208-04 - 1	WC-12R12-01	-97	-0.0104	-	7470A	PPB	22 Jan 2018 12:40:18	U	SMPL
J1208-05 - 1	WC-14-01	-46	0.0035	-	7470A	PPB	22 Jan 2018 12:42:25	U	SMPL
J1208-06 - 1	WC-14R30-01	-59	-0.0000	-	7470A	PPB	22 Jan 2018 12:44:33	U	SMPL
J1208-07 - 1	WC-14R30-02	-34	0.0068	-	7470A	PPB	22 Jan 2018 12:46:41	U	SMPL
J1208-08 - 1	WC-12R12-02	-35	0.0065	-	7470A	PPB	22 Jan 2018 12:48:49	U	SMPL
J1208-09 - 1	WC-42-02	-42	0.0046	-	7470A	PPB	22 Jan 2018 12:50:57	U	SMPL
J1208-10 - 1	WC-43-02	-68	-0.0025	-	7470A	PPB	22 Jan 2018 12:53:04	U	SMPL
CCV11 - 1	CCV11	18550	5.0810	-	7470A	PPB	22 Jan 2018 12:55:11	U	SMPL
CCB11 - 1	CCB11	-269	-0.0574	-	7470A	PPB	22 Jan 2018 12:57:18	U	SMPL
J1208-11 - 1	WC-096613-01	-64	-0.0014	-	7470A	PPB	22 Jan 2018 12:59:26	U	SMPL
J1208-12 - 1	WC-096613-02	-49	0.0027	-	7470A	PPB	22 Jan 2018 13:01:33	U	SMPL
J1208-13 - 1	WC-50-01	-20	0.0106	-	7470A	PPB	22 Jan 2018 13:03:40	U	SMPL
J1208-14 - 1	WC-114767779-01	-1	0.0158	-	7470A	PPB	22 Jan 2018 13:05:48	U	SMPL
J1208-15 - 1	WC-11R-01	1	0.0164	-	7470A	PPB	22 Jan 2018 13:07:55	U	SMPL
J1208-16 - 1	WC-53-01	-4	0.0150	-	7470A	PPB	22 Jan 2018 13:10:03	U	SMPL
J1208-17 - 1	WC-52-01	14	0.0199	-	7470A	PPB	22 Jan 2018 13:12:11	U	SMPL
J1208-18 - 1	WC-42-01	8	0.0183	-	7470A	PPB	22 Jan 2018 13:14:19	U	SMPL
J1208-19 - 1	WC-43-01	-23	0.0098	-	7470A	PPB	22 Jan 2018 13:16:28	U	SMPL
J1208-20 - 1	WC-50-02	30	0.0243	-	7470A	PPB	22 Jan 2018 13:18:35	U	SMPL
CCV12 - 1	CCV12	17851	4.8902	-	7470A	PPB	22 Jan 2018 13:20:42	U	SMPL
CCB12 - 1	CCB12	-283	-0.0612	-	7470A	PPB	22 Jan 2018 13:22:49	U	SMPL
PB105803BL - 1	PBW	-28	0.0084	-	7470A	PPB	22 Jan 2018 13:24:56	U	SMPL
PB105803BS - 1	LCSW	14097	3.8652	-	7470A	PPB	22 Jan 2018 13:27:03	U	SMPL
J1204-01 - 1	1-BSMT-CITY-WATER	-286	-0.0620	-	7470A	PPB	22 Jan 2018 13:29:10	U	SMPL
J1204-02 - 1	2-FL7-ZN2-HOT-DRAIN	-201	-0.0388	-	7470A	PPB	22 Jan 2018 13:31:17	U	SMPL
J1204-03 - 1	3-14BE-MBBTH-TUB	-190	-0.0358	-	7470A	PPB	22 Jan 2018 13:33:24	U	SMPL
J1204-04 - 1	4-14BE-MBBTH-SHWR	-186	-0.0347	-	7470A	PPB	22 Jan 2018 13:35:32	U	SMPL
J1204-05 - 1	5-14BE-KITCHEN-SINK	-197	-0.0377	-	7470A	PPB	22 Jan 2018 13:37:40	U	SMPL
J1204-05DUP - 1	5-14BE-KITCHEN-SINKDUP	-188	-0.0352	-	7470A	PPB	22 Jan 2018 13:39:48	U	SMPL
J1204-05MS - 1	5-14BE-KITCHEN-SINKMS	14440	3.9588	-	7470A	PPB	22 Jan 2018 13:41:57	U	SMPL
J1204-05MSD - 1	5-14BE-KITCHEN-SINKMSD	14133	3.8750	-	7470A	PPB	22 Jan 2018 13:44:04	U	SMPL
CCV13 - 1	CCV13	18510	5.0701	-	7470A	PPB	22 Jan 2018 13:46:11	U	SMPL
CCB13 - 1	CCB13	-281	-0.0606	-	7470A	PPB	22 Jan 2018 13:48:19	U	SMPL
I1204-06 - 1	6-ROOF-PUMP-RM-FAUCET	-187	-0.0350	-	7470A	PPB	22 Jan 2018 13:50:27	U	SMPL
I1204-07 - 1	7-14BE-KITCHEN-SINK	-171	-0.0306	-	7470A	PPB	22 Jan 2018 13:52:35	U	SMPL
I1204-08 - 1	8-14BE-MBBTH-TUB	-197	-0.0377	-	7470A	PPB	22 Jan 2018 13:54:43	U	SMPL
I1018-07LX5 - 1		-123	-0.0175	-	7470A	PPB	22 Jan 2018 13:56:50	U	SMPL
I1018-07A - 1		14456	3.9632	-	7470A	PPB	22 Jan 2018 13:58:57	U	SMPL
I1208-01LX5 - 1		-221	-0.0443	-	7470A	PPB	22 Jan 2018 14:01:05	U	SMPL
I1208-01A - 1		13939	3.8220	-	7470A	PPB	22 Jan 2018 14:03:13	U	SMPL
I1204-05LX5 - 1		-240	-0.0494	-	7470A	PPB	22 Jan 2018 14:05:21	U	SMPL
I1204-05A - 1		14453	3.9624	-	7470A	PPB	22 Jan 2018 14:07:30	U	SMPL
CCV14 - 1	CCV14	17511	4.7973	-	7470A	PPB	22 Jan 2018 14:09:37	U	SMPL
CCB14 - 1	CCB14	-340	-0.0767	-	7470A	PPB	22 Jan 2018 14:11:45	U	SMPL