


7471B

LB 92764
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44805



A= 0.0000e+000

B= 2.2110e-004

C= 1.0062e-002

Rho= 0.9997928

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	-9	0.000	-9				
0.2	0.200	0.209	0.009	898	0.0 %	898	0			
2.5	2.500	2.465	-0.035	11105	0.0 %	11105				
5.0	5.000	4.950	-0.050	22342	0.0 %	22342				
7.5	7.500	7.652	0.152	34562	0.0 %	34562				
10.0	10.000	9.916	-0.084	44805	0.0 %	44805				

LB92764 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Jan 2018 15:00:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-9	-	0.00007471B	-	PPB	22 Jan 2018 15:03:32	S	Std
0.2 - 1	50.2	898	-	0.20007471B	-	PPB	22 Jan 2018 15:05:49	S	Std
2.5 - 1	52.5	11105	-	2.50007471B	-	PPB	22 Jan 2018 15:08:07	S	Std
5.0 - 1	55	22342	-	5.00007471B	-	PPB	22 Jan 2018 15:10:24	S	Std
7.5 - 1	57.5	34562	-	7.50007471B	-	PPB	22 Jan 2018 15:12:41	S	Std
10.0 - 1	510	44805	-	10.00007471B	-	PPB	22 Jan 2018 15:14:58	S	Std
ICV59 - 1	ICV59	17060	3.7820	-7471B	-	PPB	22 Jan 2018 15:17:48	U	SMPL
ICB59 - 1	ICB59	-151	-0.0233	-7471B	-	PPB	22 Jan 2018 15:20:03	U	SMPL
CCV15 - 1	CCV15	21683	4.8042	-7471B	-	PPB	22 Jan 2018 15:22:19	U	SMPL
CCB15 - 1	CCB15	-142	-0.0213	-7471B	-	PPB	22 Jan 2018 15:24:34	U	SMPL
CRA - 1	CRA	884	0.2055	-7471B	-	PPB	22 Jan 2018 15:26:50	U	SMPL
HighStd - 1	HighStd	44621	9.8757	-7471B	-	PPB	22 Jan 2018 15:29:06	U	SMPL
ChkStd - 1	ChkStd	30947	6.8524	-7471B	-	PPB	22 Jan 2018 15:31:21	U	SMPL
PB105828BL - 1	PBS	-127	-0.0180	-7471B	-	PPB	22 Jan 2018 15:33:40	U	SMPL
PB105828BS - 1	LCSS	15918	3.5295	-7471B	-	PPB	22 Jan 2018 15:35:57	U	SMPL
J1210-01 - 1	RT-3992	25479	5.6434	-7471B	-	PPB	22 Jan 2018 15:38:14	U	SMPL
J1210-01DUP - 1	RT-3992DUP	22669	5.0222	-7471B	-	PPB	22 Jan 2018 15:40:31	U	SMPL
J1210-01MS - 1	RT-3992MS	43081	9.5352	-7471B	-	PPB	22 Jan 2018 15:45:31	U	SMPL
J1210-01MSD - 1	RT-3992MSD	40672	9.0026	-7471B	-	PPB	22 Jan 2018 15:47:48	U	SMPL
J1210-04 - 1	RT-4248	2444	0.5504	-7471B	-	PPB	22 Jan 2018 15:50:05	U	SMPL
CCV16 - 1	CCV16	21485	4.7604	-7471B	-	PPB	22 Jan 2018 15:52:23	U	SMPL
CCB16 - 1	CCB16	-151	-0.0233	-7471B	-	PPB	22 Jan 2018 15:54:39	U	SMPL
J1210-07 - 1	R-4220	5521	1.2307	-7471B	-	PPB	22 Jan 2018 15:56:55	U	SMPL
J1210-10 - 1	RT-2898	7942	1.7660	-7471B	-	PPB	22 Jan 2018 15:59:11	U	SMPL
J1212-03 - 1	OILY-STONES	185	0.0510	-7471B	-	PPB	22 Jan 2018 16:01:27	U	SMPL
J1212-05 - 1	OIL	5	0.0112	-7471B	-	PPB	22 Jan 2018 16:03:43	U	SMPL
J1213-01 - 1	CP-1-3	513	0.1235	-7471B	-	PPB	22 Jan 2018 16:06:00	U	SMPL
J1214-01 - 1	WC-PH3-MGP-01	11408	2.5324	-7471B	-	PPB	22 Jan 2018 16:08:16	U	SMPL
J1214-02 - 1	WC-PH3-MGP-02	4060	0.9077	-7471B	-	PPB	22 Jan 2018 16:10:33	U	SMPL
J1218-01 - 1	OILY-GRAVEL	117	0.0359	-7471B	-	PPB	22 Jan 2018 16:12:51	U	SMPL
J1228-22 - 1	COMP-1	3989	0.8920	-7471B	-	PPB	22 Jan 2018 16:15:08	U	SMPL
J1228-23 - 1	COMP-2	1686	0.3828	-7471B	-	PPB	22 Jan 2018 16:17:24	U	SMPL
CCV17 - 1	CCV17	21181	4.6932	-7471B	-	PPB	22 Jan 2018 16:19:41	U	SMPL
CCB17 - 1	CCB17	-122	-0.0169	-7471B	-	PPB	22 Jan 2018 16:21:57	U	SMPL
J1228-24 - 1	COMP-3	1062	0.2449	-7471B	-	PPB	22 Jan 2018 16:24:13	U	SMPL
J1228-25 - 1	COMP-4	933	0.2163	-7471B	-	PPB	22 Jan 2018 16:26:30	U	SMPL
J1228-26 - 1	COMP-5	340	0.0852	-7471B	-	PPB	22 Jan 2018 16:28:46	U	SMPL
J1229-02 - 1	TEL-COMP-1	918	0.2130	-7471B	-	PPB	22 Jan 2018 16:31:03	U	SMPL
J1235-05 - 1	32-2	828	0.1931	-7471B	-	PPB	22 Jan 2018 16:33:20	U	SMPL
J1235-07 - 1	30-2	5861	1.3059	-7471B	-	PPB	22 Jan 2018 16:35:37	U	SMPL
J1240-01 - 1	CA-1	2418	0.5447	-7471B	-	PPB	22 Jan 2018 16:37:54	U	SMPL
PB105829BL - 1	PBS	0	0.0101	-7471B	-	PPB	22 Jan 2018 16:42:56	U	SMPL
PB105829BS - 1	LCSS	15538	3.4455	-7471B	-	PPB	22 Jan 2018 16:45:13	U	SMPL
J1236-01 - 1	CA-3	107839	23.8532	-7471B	-	PPB	22 Jan 2018 16:47:29	U	SMPL
CCV18 - 1	CCV18	20447	4.5309	-7471B	-	PPB	22 Jan 2018 16:49:45	U	SMPL
CCB18 - 1	CCB18	-225	-0.0397	-7471B	-	PPB	22 Jan 2018 16:52:04	U	SMPL
J1236-02 - 1	33E-6	16246	3.6020	-7471B	-	PPB	22 Jan 2018 16:54:21	U	SMPL
J1236-03 - 1	6-2	2605	0.5860	-7471B	-	PPB	22 Jan 2018 16:56:38	U	SMPL
J1236-04 - 1	20A-5	308	0.0782	-7471B	-	PPB	22 Jan 2018 16:58:54	U	SMPL
J1236-07 - 1	31A-5	5655	1.2604	-7471B	-	PPB	22 Jan 2018 17:01:10	U	SMPL
J1238-01 - 1	SB-06-0-2.0	9872	2.1928	-7471B	-	PPB	22 Jan 2018 17:03:27	U	SMPL
J1238-02 - 1	SB-06-0-2.0-DUP	5997	1.3360	-7471B	-	PPB	22 Jan 2018 17:05:43	U	SMPL
J1238-03 - 1	SB-07-0-2.0	7638	1.6988	-7471B	-	PPB	22 Jan 2018 17:08:01	U	SMPL
J1238-04 - 1	SB-08-0-2.0	5835	1.3002	-7471B	-	PPB	22 Jan 2018 17:10:18	U	SMPL
J1238-05 - 1	SB-09-0-2.0	31473	6.9687	-7471B	-	PPB	22 Jan 2018 17:12:36	U	SMPL
J1238-06 - 1	SB-10-0-2.0	12655	2.8081	-7471B	-	PPB	22 Jan 2018 17:14:52	U	SMPL
CCV19 - 1	CCV19	21805	4.8311	-7471B	-	PPB	22 Jan 2018 17:17:09	U	SMPL
CCB19 - 1	CCB19	-164	-0.0262	-7471B	-	PPB	22 Jan 2018 17:19:26	U	SMPL
J1238-07 - 1	SB-11-0-2.0	7367	1.6389	-7471B	-	PPB	22 Jan 2018 17:21:42	U	SMPL
J1238-08 - 1	SB-01-6.5-7.0	740	0.1737	-7471B	-	PPB	22 Jan 2018 17:23:59	U	SMPL
J1238-10 - 1	SB-02-6.5-7.0	2949	0.6621	-7471B	-	PPB	22 Jan 2018 17:26:15	U	SMPL
J1238-12 - 1	SB-03-7.5-8.0	12408	2.7535	-7471B	-	PPB	22 Jan 2018 17:28:31	U	SMPL
J1238-14 - 1	SB-04-9.5-10.0	3500	0.7839	-7471B	-	PPB	22 Jan 2018 17:30:48	U	SMPL
J1238-16 - 1	SB-05-9.0-10.0	4360	0.9741	-7471B	-	PPB	22 Jan 2018 17:33:05	U	SMPL
J1238-18 - 1	SB-06-9.5-10.0	1136	0.2612	-7471B	-	PPB	22 Jan 2018 17:35:22	U	SMPL
J1238-20 - 1	SB-06-9.5-10.0-DUP	1335	0.3052	-7471B	-	PPB	22 Jan 2018 17:37:40	U	SMPL
J1238-22 - 1	SB-07-9.5-10.0	895	0.2079	-7471B	-	PPB	22 Jan 2018 17:39:58	U	SMPL
J1238-22DUP - 1	SB-07-9.5-10.0DUP	1015	0.2345	-7471B	-	PPB	22 Jan 2018 17:42:15	U	SMPL
CCV20 - 1	CCV20	21912	4.8548	-7471B	-	PPB	22 Jan 2018 17:44:31	U	SMPL
CCB20 - 1	CCB20	-161	-0.0255	-7471B	-	PPB	22 Jan 2018 17:46:48	U	SMPL
J1238-22MS - 1	SB-07-9.5-10.0MS	16724	3.7077	-7471B	-	PPB	22 Jan 2018 17:49:05	U	SMPL
J1238-22MSD - 1	SB-07-9.5-10.0MSD	16148	3.5804	-7471B	-	PPB	22 Jan 2018 17:51:21	U	SMPL
PB105830BL - 1	PBS	-118	-0.0160	-7471B	-	PPB	22 Jan 2018 17:53:38	U	SMPL
PB105830BS - 1	LCSS	15086	3.3456	-7471B	-	PPB	22 Jan 2018 17:55:55	U	SMPL
1241-01 - 1	EPE-108-4(0-5)	444	0.1082	-7471B	-	PPB	22 Jan 2018 17:58:11	U	SMPL
1241-02 - 1	EPE-108-4(5-10)	131	0.0390	-7471B	-	PPB	22 Jan 2018 18:00:29	U	SMPL
1241-03 - 1	EPE-106-4A(85-87)	463	0.1124	-7471B	-	PPB	22 Jan 2018 18:02:46	U	SMPL
1241-03DUP - 1	EPE-106-4A(85-87)DUP	525	0.1261	-7471B	-	PPB	22 Jan 2018 18:05:04	U	SMPL
1241-03MS - 1	EPE-106-4A(85-87)MS	16423	3.6412	-7471B	-	PPB	22 Jan 2018 18:07:21	U	SMPL
1241-03MSD - 1	EPE-106-4A(85-87)MSD	16198	3.5914	-7471B	-	PPB	22 Jan 2018 18:09:38	U	SMPL
CV21 - 1	CCV21	21729	4.8143	-7471B	-	PPB	22 Jan 2018 18:11:55	U	SMPL

LB92764 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Jan 2018 15:00:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB21 - 1	CCB21	-259	-0.0472	-7471B	PPB		22 Jan 2018 18:14:12	U	SMPL
J1241-04 - 1	EPE-106-4A(90-92)	1100	0.2533	-7471B	PPB		22 Jan 2018 18:16:29	U	SMPL
J1241-05 - 1	EPE-106-4A(110-112)	188	0.0516	-7471B	PPB		22 Jan 2018 18:18:46	U	SMPL
J1241-06 - 1	EPE-121-2(0-5)	10642	2.3630	-7471B	PPB		22 Jan 2018 18:21:03	U	SMPL
J1210-01LX5 - 1		4552	1.0165	-7471B	PPB		22 Jan 2018 18:23:19	U	SMPL
J1210-01A - 1		40621	8.9913	-7471B	PPB		22 Jan 2018 18:25:36	U	SMPL
J1238-22LX5 - 1		-158	-0.0249	-7471B	PPB		22 Jan 2018 18:27:53	U	SMPL
J1238-22A - 1		17705	3.9246	-7471B	PPB		22 Jan 2018 18:30:12	U	SMPL
J1241-03LX5 - 1		-125	-0.0176	-7471B	PPB		22 Jan 2018 18:32:30	U	SMPL
J1241-03A - 1		17213	3.8158	-7471B	PPB		22 Jan 2018 18:34:48	U	SMPL
J1236-01DLX5 - 1		21422	4.7464	-7471B	PPB		22 Jan 2018 18:37:04	U	SMPL
CCV22 - 1	CCV22	21738	4.8163	-7471B	PPB		22 Jan 2018 18:39:21	U	SMPL
CCB22 - 1	CCB22	-278	-0.0514	-7471B	PPB		22 Jan 2018 18:41:39	U	SMPL