

7470A


LB92784

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37528



A= 0.0000e+000

B= 2.6873e-004

C= 6.9476e-003

Rho= 0.9998463

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	-44	0.000	-44				
0.2	0.200	0.233	0.033	840	0.0 %	840				
2.5	2.500	2.543	0.043	9436	0.0 %	9436				
5.0	5.000	4.936	-0.064	18341	0.0 %	18341				
7.5	7.500	7.402	-0.098	27518	0.0 %	27518				
10.0	10.000	10.092	0.092	37528	0.0 %	37528				

LB92784 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jan 2018 10:51:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-44	-	0.0000	7470A	PPB	23 Jan 2018 11:00:09	S	Std
0.2 - 1	50.2	840	-	0.2000	7470A	PPB	23 Jan 2018 11:02:17	S	Std
2.5 - 1	52.5	9436	-	2.5000	7470A	PPB	23 Jan 2018 11:04:25	S	Std
5.0 - 1	55	18341	-	5.0000	7470A	PPB	23 Jan 2018 11:06:34	S	Std
7.5 - 1	57.5	27518	-	7.5000	7470A	PPB	23 Jan 2018 11:08:42	S	Std
10.0 - 1	510	37528	-	10.0000	7470A	PPB	23 Jan 2018 11:10:49	S	Std
ICV60 - 1	ICV60	14151	3.8098	-	7470A	PPB	23 Jan 2018 11:13:39	U	SMPL
ICB60 - 1	ICB60	-152	-0.0339	-	7470A	PPB	23 Jan 2018 11:15:45	U	SMPL
CCV23 - 1	CCV23	18176	4.8914	-	7470A	PPB	23 Jan 2018 11:17:52	U	SMPL
CCB23 - 1	CCB23	-161	-0.0363	-	7470A	PPB	23 Jan 2018 11:19:59	U	SMPL
CRA - 1	CRA	838	0.2321	-	7470A	PPB	23 Jan 2018 11:22:06	U	SMPL
HighStd - 1	HighStd	37883	10.1873	-	7470A	PPB	23 Jan 2018 11:24:12	U	SMPL
ChkStd - 1	ChkStd	26171	7.0399	-	7470A	PPB	23 Jan 2018 11:26:19	U	SMPL
PB105867BL - 1	PBW	-118	-0.0248	-	7470A	PPB	23 Jan 2018 11:28:27	U	SMPL
PB105867BS - 1	LCSW	14211	3.8259	-	7470A	PPB	23 Jan 2018 11:30:36	U	SMPL
PB105867TB - 1	PB105806TB	-70	-0.0119	-	7470A	PPB	23 Jan 2018 11:32:43	U	SMPL
J1186-01 - 1	OVS-1	39	0.0174	-	7470A	PPB	23 Jan 2018 11:34:51	U	SMPL
J1186-01DUP - 1	OVS-1DUP	77	0.0276	-	7470A	PPB	23 Jan 2018 11:36:59	U	SMPL
J1186-01MS - 1	OVS-1MS	13973	3.7619	-	7470A	PPB	23 Jan 2018 11:39:07	U	SMPL
J1186-01MSD - 1	OVS-1MSD	14733	3.9662	-	7470A	PPB	23 Jan 2018 11:41:14	U	SMPL
CCV24 - 1	CCV24	18366	4.9425	-	7470A	PPB	23 Jan 2018 11:43:21	U	SMPL
CCB24 - 1	CCB24	-171	-0.0390	-	7470A	PPB	23 Jan 2018 11:45:28	U	SMPL
J1186-02 - 1	OVS-2	201	0.0610	-	7470A	PPB	23 Jan 2018 11:47:34	U	SMPL
J1186-03 - 1	OVS-3	203	0.0615	-	7470A	PPB	23 Jan 2018 11:49:41	U	SMPL
J1186-04 - 1	OVS-4	79	0.0282	-	7470A	PPB	23 Jan 2018 11:51:48	U	SMPL
J1186-05 - 1	OVS-5	64	0.0241	-	7470A	PPB	23 Jan 2018 11:53:55	U	SMPL
J1186-06 - 1	OVS-6	50	0.0204	-	7470A	PPB	23 Jan 2018 11:56:02	U	SMPL
J1188-01 - 1	OVS-1	61	0.0233	-	7470A	PPB	23 Jan 2018 11:58:10	U	SMPL
J1188-02 - 1	OVS-2	20	0.0123	-	7470A	PPB	23 Jan 2018 12:00:18	U	SMPL
J1210-03 - 1	RT-3992	27	0.0142	-	7470A	PPB	23 Jan 2018 12:02:26	U	SMPL
J1210-06 - 1	RT-4248	16	0.0112	-	7470A	PPB	23 Jan 2018 12:04:34	U	SMPL
J1210-09 - 1	R-4220	9	0.0094	-	7470A	PPB	23 Jan 2018 12:06:41	U	SMPL
CCV25 - 1	CCV25	18115	4.8750	-	7470A	PPB	23 Jan 2018 12:08:48	U	SMPL
CCB25 - 1	CCB25	-169	-0.0385	-	7470A	PPB	23 Jan 2018 12:10:55	U	SMPL
J1210-12 - 1	RT-2898	0	0.0069	-	7470A	PPB	23 Jan 2018 12:13:02	U	SMPL
J1212-04 - 1	OILY-STONES	48	0.0198	-	7470A	PPB	23 Jan 2018 12:15:09	U	SMPL
J1212-10 - 1	OILY-DEBRIS	71	0.0260	-	7470A	PPB	23 Jan 2018 12:17:16	U	SMPL
J1214-03 - 1	WC-PH3-MGP-01	57	0.0223	-	7470A	PPB	23 Jan 2018 12:19:23	U	SMPL
J1214-04 - 1	WC-PH3-MGP-02	79	0.0282	-	7470A	PPB	23 Jan 2018 12:21:31	U	SMPL
J1218-02 - 1	OILY-GRAVEL	39	0.0174	-	7470A	PPB	23 Jan 2018 12:23:39	U	SMPL
J1218-09 - 1	OILY-DEBRIS	53	0.0212	-	7470A	PPB	23 Jan 2018 12:25:47	U	SMPL
J1230-02 - 1	OILY-DEBRIS	70	0.0258	-	7470A	PPB	23 Jan 2018 12:27:55	U	SMPL
J1230-07 - 1	CO-1-6	65	0.0244	-	7470A	PPB	23 Jan 2018 12:30:04	U	SMPL
PB105868BL - 1	PBW	98	0.0333	-	7470A	PPB	23 Jan 2018 12:32:11	U	SMPL
CCV26 - 1	CCV26	17820	4.7957	-	7470A	PPB	23 Jan 2018 12:34:18	U	SMPL
CCB26 - 1	CCB26	79	0.0282	-	7470A	PPB	23 Jan 2018 12:36:25	U	SMPL
PB105868BS - 1	LCSW	12956	3.4886	-	7470A	PPB	23 Jan 2018 12:38:33	U	SMPL
PB105868TB - 1	PB105840TB	-18	0.0021	-	7470A	PPB	23 Jan 2018 12:40:40	U	SMPL
J1213-04 - 1	MANHOLE	20	0.0123	-	7470A	PPB	23 Jan 2018 12:42:47	U	SMPL
J1229-03 - 1	TEL-COMP-1	27	0.0142	-	7470A	PPB	23 Jan 2018 12:44:54	U	SMPL
J1230-04 - 1	EXCAVATION	48	0.0198	-	7470A	PPB	23 Jan 2018 12:47:02	U	SMPL
J1238-09 - 1	SB-01-COMP	65	0.0244	-	7470A	PPB	23 Jan 2018 12:49:09	U	SMPL
J1238-11 - 1	SB-02-COMP	40	0.0177	-	7470A	PPB	23 Jan 2018 12:51:18	U	SMPL
J1238-13 - 1	SB-03-COMP	58	0.0225	-	7470A	PPB	23 Jan 2018 12:53:26	U	SMPL
J1238-15 - 1	SB-04-COMP	67	0.0250	-	7470A	PPB	23 Jan 2018 12:55:34	U	SMPL
J1238-17 - 1	SB-05-COMP	46	0.0193	-	7470A	PPB	23 Jan 2018 12:57:42	U	SMPL
CCV27 - 1	CCV27	17996	4.8430	-	7470A	PPB	23 Jan 2018 12:59:49	U	SMPL
CCB27 - 1	CCB27	65	0.0244	-	7470A	PPB	23 Jan 2018 13:01:56	U	SMPL
J1238-19 - 1	SB-06-COMP	-19	0.0018	-	7470A	PPB	23 Jan 2018 13:04:03	U	SMPL
J1238-21 - 1	SB-06-COMP-DUP	2	0.0075	-	7470A	PPB	23 Jan 2018 13:06:10	U	SMPL
J1238-23 - 1	SB-07-COMP	8	0.0091	-	7470A	PPB	23 Jan 2018 13:08:17	U	SMPL
J1240-03 - 1	CA-1	27	0.0142	-	7470A	PPB	23 Jan 2018 13:10:24	U	SMPL
J1240-03DUP - 1	CA-1DUP	-22	0.0010	-	7470A	PPB	23 Jan 2018 13:12:32	U	SMPL
J1240-03MS - 1	CA-1MS	13222	3.6601	-	7470A	PPB	23 Jan 2018 13:14:40	U	SMPL
J1240-03MSD - 1	CA-1MSD	13388	3.6047	-	7470A	PPB	23 Jan 2018 13:16:47	U	SMPL
PB105869BL - 1	PBW	-63	-0.0100	-	7470A	PPB	23 Jan 2018 13:18:56	U	SMPL
PB105869BS - 1	LCSW	13832	3.7240	-	7470A	PPB	23 Jan 2018 13:21:04	U	SMPL
J1213-03 - 1	MANHOLE	-248	-0.0597	-	7470A	PPB	23 Jan 2018 13:23:11	U	SMPL
CCV28 - 1	CCV28	17900	4.8172	-	7470A	PPB	23 Jan 2018 13:25:18	U	SMPL
CCB28 - 1	CCB28	42	0.0182	-	7470A	PPB	23 Jan 2018 13:27:25	U	SMPL
J1230-03 - 1	EXCAVATION	734	0.2042	-	7470A	PPB	23 Jan 2018 13:29:33	U	SMPL
J1230-03MS - 1	EXCAVATIONMS	15539	4.1828	-	7470A	PPB	23 Jan 2018 13:33:47	U	SMPL
J1230-03MSD - 1	EXCAVATIONMSD	15046	4.0503	-	7470A	PPB	23 Jan 2018 13:35:55	U	SMPL
J1235-01 - 1	TW-3	1339	0.3668	-	7470A	PPB	23 Jan 2018 13:38:02	U	SMPL
J1236-06 - 1	TW-2	7189	1.9389	-	7470A	PPB	23 Jan 2018 13:40:10	U	SMPL
J1240-02 - 1	PT-1	133	0.0427	-	7470A	PPB	23 Jan 2018 13:42:18	U	SMPL
J1186-01LX5 - 1		108	0.0360	-	7470A	PPB	23 Jan 2018 13:44:26	U	SMPL
J1186-01A - 1		15441	4.1564	-	7470A	PPB	23 Jan 2018 13:46:34	U	SMPL
J1230-03DUP - 1	EXCAVATION	670	0.1870	-	7470A	PPB	23 Jan 2018 13:51:18	U	SMPL
CCV29 - 1	CCV29	17707	4.7654	-	7470A	PPB	23 Jan 2018 13:53:26	U	SMPL
CCB29 - 1	CCB29	58	0.0225	-	7470A	PPB	23 Jan 2018 13:55:34	U	SMPL

LB92784 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jan 2018 10:51:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
J1240-03LX5 - 1		114	0.0376		-7470A	PPB	23 Jan 2018 13:57:42	U	SMPL
J1240-03A - 1		14368	3.8681		-7470A	PPB	23 Jan 2018 13:59:49	U	SMPL
J1230-03LX5 - 1		-33	-0.0019		-7470A	PPB	23 Jan 2018 14:01:57	U	SMPL
J1230-03A - 1		14805	3.9855		-7470A	PPB	23 Jan 2018 14:04:04	U	SMPL
CCV30 - 1	CCV30	17884	4.8129		-7470A	PPB	23 Jan 2018 14:07:58	U	SMPL
CCB30 - 1	CCB30	11	0.0099		-7470A	PPB	23 Jan 2018 14:10:05	U	SMPL