



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92789

Temp.: 19.8 °C

SLOPE.: 99.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/22/2018	16:15
2	CAL2	1	Water			6.99	01/22/2018	16:17
3	CAL3	1	Water			10.01	01/22/2018	16:19
4	ICV	1	Water			7.00	01/22/2018	16:21
5	CCV1	1	Water			2.01	01/22/2018	16:23
6	J1252-01	1	Water			5.02	01/22/2018	16:25
7	J1252-01DUP	1	Water			5.03	01/22/2018	16:27
8	CCV2	1	Water			12.01	01/22/2018	16:30

WORKLIST(Hardcopy Internal Chain)

LB92789

WorkList Name : pH J1252

WorkList ID : 107732

Date : 1/22/2018 04:00:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/13/2017	Water	J1252-01	pH	Cool 4 deg C	STER13	M41	STERIGENICS-TUSTIN-MAT1	11/13/2017	SM 4500-PH B

Date/Time

01/22/18 4:05 PM

Received by:

SP

Relinquished by:

SP

Date/Time

01/22/18 4:40 PM

Received by:

SP

Relinquished by:

SP