


ISM02.4

LB 92791
INSTRUMENT ID: CV1

Linear

 μ Abs.:

46697



A= 0.0000e+000

B= 2.1612e-004

C= 3.1841e-003

Rho= 0.9997806

Accept=Accepted

slope
1/101250

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.001	-0.001	-21	0.000	-21					4
0.2	0.200	0.209	0.009	952	0.0 %	952					1
2.5	2.500	2.532	0.032	11703	0.0 %	11703					0
5.0	5.000	5.023	0.023	23226	0.0 %	23226					-2
7.5	7.500	7.342	-0.158	33955	0.0 %	33955					1
10.0	10.000	10.095	0.095	46697	0.0 %	46697					

NB
1/23/18

LB92791
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-21	-		0 ISM02.4	PPB	1/23/2018 14:37	Std
	0.2 S0.2	952	-		0.2 ISM02.4	PPB	1/23/2018 14:39	Std
	2.5 S2.5	11703	-		2.5 ISM02.4	PPB	1/23/2018 14:41	Std
	5 S5	23226	-		5 ISM02.4	PPB	1/23/2018 14:43	Std
	7.5 S7.5	33955	-		7.5 ISM02.4	PPB	1/23/2018 14:46	Std
	10 S10	46697	-		10 ISM02.4	PPB	1/23/2018 14:48	Std
ICV	ICV	17503	3.786 -		ISM02.4	PPB	1/23/2018 14:51	SMPL
ICB	ICB	-11	0.0008 -		ISM02.4	PPB	1/23/2018 14:53	SMPL
CCV031	CCV031	21803	4.7153 -		ISM02.4	PPB	1/23/2018 14:55	SMPL
CCB031	CCB031	-29	-0.0031 -		ISM02.4	PPB	1/23/2018 14:58	SMPL
PB105874BL	PBS004	140	0.0334 -		ISM02.4	PPB	1/23/2018 15:00	SMPL
J1224-01	MF6A02	1616	0.3524 -		ISM02.4	PPB	1/23/2018 15:02	SMPL
J1224-02	MF6A03	1725	0.376 -		ISM02.4	PPB	1/23/2018 15:04	SMPL
J1224-03	MF6A05	1923	0.4188 -		ISM02.4	PPB	1/23/2018 15:07	SMPL
J1224-04	MF6A06	1379	0.3012 -		ISM02.4	PPB	1/23/2018 15:09	SMPL
J1224-05	MF6A07	1362	0.2975 -		ISM02.4	PPB	1/23/2018 15:11	SMPL
J1224-06	MF6A08	1346	0.2941 -		ISM02.4	PPB	1/23/2018 15:14	SMPL
J1224-07	MF6A09	1011	0.2217 -		ISM02.4	PPB	1/23/2018 15:16	SMPL
J1224-08	MF6A10	2931	0.6366 -		ISM02.4	PPB	1/23/2018 15:18	SMPL
J1224-09	MF6A11	1113	0.2437 -		ISM02.4	PPB	1/23/2018 15:20	SMPL
J1224-10	MF6A12	1825	0.3976 -		ISM02.4	PPB	1/23/2018 15:23	SMPL
J1224-11	MF6A16	1236	0.2703 -		ISM02.4	PPB	1/23/2018 15:25	SMPL
J1224-12	MF6A16D	1160	0.2539 -		ISM02.4	PPB	1/23/2018 15:27	SMPL
J1224-13	MF6A16S	13271	2.8713 -		ISM02.4	PPB	1/23/2018 15:30	SMPL
J1224-14	MF6A17	1157	0.2532 -		ISM02.4	PPB	1/23/2018 15:32	SMPL
J1224-15	MF6A24	1750	0.3814 -		ISM02.4	PPB	1/23/2018 15:34	SMPL
J1224-16	MF6A13	947	0.2079 -		ISM02.4	PPB	1/23/2018 15:36	SMPL
J1224-17	MF6A14	1143	0.2502 -		ISM02.4	PPB	1/23/2018 15:39	SMPL
J1224-18	MF6A15	1746	0.3805 -		ISM02.4	PPB	1/23/2018 15:41	SMPL
J1224-19	MF6A18	888	0.1951 -		ISM02.4	PPB	1/23/2018 15:43	SMPL
CCV032	CCV032	22970	4.9675 -		ISM02.4	PPB	1/23/2018 15:46	SMPL
CCB032	CCB032	-26	-0.0024 -		ISM02.4	PPB	1/23/2018 15:48	SMPL
J1224-20	MF6A19	1058	0.2318 -		ISM02.4	PPB	1/23/2018 15:50	SMPL
J1224-21	MF6A20	1566	0.3416 -		ISM02.4	PPB	1/23/2018 15:52	SMPL
J1224-22	MF6A21	1318	0.288 -		ISM02.4	PPB	1/23/2018 15:55	SMPL
PB105875BL	PBS004	149	0.0354 -		ISM02.4	PPB	1/23/2018 15:57	SMPL
J1225-01	MF6A63	609	0.1348 -		ISM02.4	PPB	1/23/2018 15:59	SMPL
J1225-02	MF6A64	664	0.1467 -		ISM02.4	PPB	1/23/2018 16:01	SMPL
J1225-03	MF6A65	489	0.1089 -		ISM02.4	PPB	1/23/2018 16:04	SMPL
J1225-04	MF6A66	671	0.1482 -		ISM02.4	PPB	1/23/2018 16:06	SMPL
J1225-05	MF6A67	354	0.0797 -		ISM02.4	PPB	1/23/2018 16:08	SMPL
J1225-06	MF6A68	747	0.1646 -		ISM02.4	PPB	1/23/2018 16:11	SMPL
J1225-07	MF6A69	402	0.0901 -		ISM02.4	PPB	1/23/2018 16:13	SMPL
J1225-08	MF6A70	414	0.0927 -		ISM02.4	PPB	1/23/2018 16:15	SMPL
J1225-09	MF6A71	587	0.13 -		ISM02.4	PPB	1/23/2018 16:17	SMPL
J1225-10	MF6A72	314	0.071 -		ISM02.4	PPB	1/23/2018 16:20	SMPL

LB92791
INSTRUMENT ID : CV1

J1225-11	MF6A73	683	0.1508 -	ISM02.4	PPB	1/23/2018 16:22	SMPL
J1225-12	MF6A74	489	0.1089 -	ISM02.4	PPB	1/23/2018 16:24	SMPL
J1225-13	MF6A75	586	0.1298 -	ISM02.4	PPB	1/23/2018 16:26	SMPL
J1225-14	MF6A76	894	0.1964 -	ISM02.4	PPB	1/23/2018 16:29	SMPL
J1225-15	MF6A76D	863	0.1897 -	ISM02.4	PPB	1/23/2018 16:34	SMPL
J1225-16	MF6A76S	12811	2.7719 -	ISM02.4	PPB	1/23/2018 16:36	SMPL
CCV033	CCV033	22762	4.9226 -	ISM02.4	PPB	1/23/2018 16:39	SMPL
CCB033	CCB033	-24	-0.002 -	ISM02.4	PPB	1/23/2018 16:41	SMPL
J1225-17	MF6A77	161	0.038 -	ISM02.4	PPB	1/23/2018 16:43	SMPL
J1225-18	MF6A79	1687	0.3678 -	ISM02.4	PPB	1/23/2018 16:46	SMPL
J1225-19	MF6A80	408	0.0914 -	ISM02.4	PPB	1/23/2018 16:48	SMPL
PB105876BL	PBS004	68	0.0179 -	ISM02.4	PPB	1/23/2018 16:50	SMPL
J1226-01	MF6A95	10022	2.1692 -	ISM02.4	PPB	1/23/2018 16:52	SMPL
J1226-02	MF6A96	548	0.1216 -	ISM02.4	PPB	1/23/2018 16:55	SMPL
J1226-03	MF6A96D	688	0.1519 -	ISM02.4	PPB	1/23/2018 16:57	SMPL
J1226-04	MF6A96S	12627	2.7322 -	ISM02.4	PPB	1/23/2018 16:59	SMPL
J1226-05	MF6A97	708	0.1562 -	ISM02.4	PPB	1/23/2018 17:01	SMPL
J1226-06	MF6A98	466	0.1039 -	ISM02.4	PPB	1/23/2018 17:04	SMPL
J1226-07	MF6A99	265	0.0605 -	ISM02.4	PPB	1/23/2018 17:06	SMPL
J1226-08	MF6B00	772	0.17 -	ISM02.4	PPB	1/23/2018 17:08	SMPL
J1226-09	MF6B01	728	0.1605 -	ISM02.4	PPB	1/23/2018 17:11	SMPL
J1226-10	MF6B02	571	0.1266 -	ISM02.4	PPB	1/23/2018 17:13	SMPL
J1226-13	MF6B03	403	0.0903 -	ISM02.4	PPB	1/23/2018 17:15	SMPL
J1226-14	MF6B04	375	0.0842 -	ISM02.4	PPB	1/23/2018 17:17	SMPL
J1226-15	MF6B06	972	0.2133 -	ISM02.4	PPB	1/23/2018 17:20	SMPL
J1226-16	MF6B07	466	0.1039 -	ISM02.4	PPB	1/23/2018 17:22	SMPL
J1226-17	MF6B08	590	0.1307 -	ISM02.4	PPB	1/23/2018 17:24	SMPL
J1226-18	MF6B09	456	0.1017 -	ISM02.4	PPB	1/23/2018 17:27	SMPL
CCV034	CCV034	23277	5.0339 -	ISM02.4	PPB	1/23/2018 17:29	SMPL
CCB034	CCB034	-61	-0.01 -	ISM02.4	PPB	1/23/2018 17:31	SMPL
J1226-19	MF6B10	623	0.1378 -	ISM02.4	PPB	1/23/2018 17:33	SMPL
J1226-20	MF6B11	187	0.0436 -	ISM02.4	PPB	1/23/2018 17:36	SMPL
PB105877BL	PBS004	51	0.0142 -	ISM02.4	PPB	1/23/2018 17:38	SMPL
J1227-01	MF6A22	1501	0.3276 -	ISM02.4	PPB	1/23/2018 17:40	SMPL
J1227-02	MF6A23	3136	0.6809 -	ISM02.4	PPB	1/23/2018 17:43	SMPL
J1227-03	MF6A25	1640	0.3576 -	ISM02.4	PPB	1/23/2018 17:45	SMPL
J1227-04	MF6A81	687	0.1517 -	ISM02.4	PPB	1/23/2018 17:47	SMPL
J1227-05	MF6A82	562	0.1246 -	ISM02.4	PPB	1/23/2018 17:49	SMPL
J1227-06	MF6A83	794	0.1748 -	ISM02.4	PPB	1/23/2018 17:52	SMPL
J1227-07	MF6A84	854	0.1878 -	ISM02.4	PPB	1/23/2018 17:54	SMPL
J1227-08	MF6A85	469	0.1045 -	ISM02.4	PPB	1/23/2018 17:56	SMPL
J1227-09	MF6B22	918	0.2016 -	ISM02.4	PPB	1/23/2018 17:59	SMPL
J1227-10	MF6B23	485	0.108 -	ISM02.4	PPB	1/23/2018 18:01	SMPL
J1227-11	MF6B24	235	0.054 -	ISM02.4	PPB	1/23/2018 18:03	SMPL
J1227-12	MF6B25	613	0.1357 -	ISM02.4	PPB	1/23/2018 18:05	SMPL
J1227-13	MF6B26	498	0.1108 -	ISM02.4	PPB	1/23/2018 18:08	SMPL
J1227-14	MF6B12	14679	3.1757 -	ISM02.4	PPB	1/23/2018 18:10	SMPL

LB92791
INSTRUMENT ID : CV1

J1227-15	MF6B13	436	0.0974 -	ISM02.4	PPB	1/23/2018 18:12 SMPL
J1227-16	MF6B14	362	0.0814 -	ISM02.4	PPB	1/23/2018 18:15 SMPL
J1227-17	MF6B15	274	0.0624 -	ISM02.4	PPB	1/23/2018 18:17 SMPL
CCV035	CCV035	23125	5.001 -	ISM02.4	PPB	1/23/2018 18:19 SMPL
CCB035	CCB035	-46	-0.0068 -	ISM02.4	PPB	1/23/2018 18:21 SMPL
J1227-18	MF6B16	322	0.0728 -	ISM02.4	PPB	1/23/2018 18:24 SMPL
J1227-19	MF6B16D	295	0.0669 -	ISM02.4	PPB	1/23/2018 18:26 SMPL
J1227-20	MF6B16S	11984	2.5932 -	ISM02.4	PPB	1/23/2018 18:28 SMPL
J1227-21	MF6B17	6048	1.3103 -	ISM02.4	PPB	1/23/2018 18:31 SMPL
J1227-22	MF6B18	1038	0.2275 -	ISM02.4	PPB	1/23/2018 18:33 SMPL
CCV036	CCV036	23137	5.0036 -	ISM02.4	PPB	1/23/2018 18:37 SMPL
CCB036	CCB036	-49	-0.0074 -	ISM02.4	PPB	1/23/2018 18:40 SMPL