

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: Aparna

Date: 01/24/2018

DATE & TIME IN: 01/23/2018 16:00

OVEN TEMP: 103 °C

Run Number: LB92795

DATE & TIME OUT: 01/24/2018 09:15

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB92795BL	LB92795BL	76.9615	76.9615	100	76.9615	76.9615	76.9615	0	0
2	J1192-20	1002	85.4562	85.4562	100	85.4981	85.4981	85.4981	0.0419	419
3	J1262-01	BELL-0118	82.4763	82.4763	100	82.5803	82.5803	82.5803	0.104	1040
4	J1264-01	2-AVE-09	84.0521	84.0521	100	84.21	84.21	84.21	0.1579	1579
5	J1264-01DUP	2-AVE-09DUP	79.9826	79.9826	100	80.1459	80.1459	80.1459	0.1633	1633

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS012318

WorkList ID : 107760

Date : 1/23/2018 3:06:48 PM

LB92795

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2018	Water	J1262-01	TS	Cool 4 deg C	DIST01	L62	BELL-0118	01/22/2018	SM2540 B
01/29/2018	Water	J1264-01	TS	Cool 4 deg C	DIST01	L62	2-AVE-09	01/22/2018	SM2540 B

Date/Time 01/23/18, 15:20
 Received by: AS
 Relinquished by: CP

Date/Time 01/23/18, 15:55
 Received by: CP
 Relinquished by: AS