



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92816

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
IODINE SOLUTION .025N 1L	W2261
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.01 Control Limits[+/- 0.1].

True Value of CCV4 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/23/2018	17:10
2	CAL2	1	Water			6.99	01/23/2018	17:11
3	CAL3	1	Water			10.03	01/23/2018	17:12
4	ICV	1	Water			7.01	01/23/2018	17:15
5	CCV1	1	Water			2.01	01/23/2018	17:16
6	J1259-01	1	Solid	20.02	20	8.38	01/23/2018	17:18
7	J1259-01DUP	1	Solid	20.03	20	8.39	01/23/2018	17:19
8	J1259-02	1	Solid	20.01	20	9.03	01/23/2018	17:20
9	J1259-03	1	Solid	20.02	20	10.97	01/23/2018	17:21
10	J1259-04	1	Solid	20.05	20	9.82	01/23/2018	17:24
11	J1259-05	1	Solid	20.04	20	11.12	01/23/2018	17:25
12	J1259-06	1	Solid	20.01	20	9.74	01/23/2018	17:26
13	J1259-07	1	Solid	20.02	20	8.70	01/23/2018	17:30
14	J1259-08	1	Solid	20.04	20	9.01	01/23/2018	17:31
15	J1259-09	1	Solid	20.03	20	10.46	01/23/2018	17:32
16	CCV2	1	Water			12.02	01/23/2018	17:33
17	J1259-10	1	Solid	20.05	20	8.88	01/23/2018	17:35
18	J1259-11	1	Solid	20.02	20	8.69	01/23/2018	17:36
19	J1259-12	1	Solid	20.01	20	10.06	01/23/2018	17:39
20	J1259-13	1	Solid	20.02	20	8.87	01/23/2018	17:40

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	J1259-14	1	Solid	20.03	20	8.89	01/23/2018	17:41
22	J1259-15	1	Solid	20.05	20	8.45	01/23/2018	17:42
23	J1259-16	1	Solid	20.02	20	8.75	01/23/2018	17:44
24	J1259-17	1	Solid	20.01	20	11.04	01/23/2018	17:46
25	J1259-18	1	Solid	20.04	20	8.89	01/23/2018	17:47
26	J1259-19	1	Solid	20.03	20	8.18	01/23/2018	17:48
27	CCV3	1	Water			2.02	01/23/2018	17:49
28	J1279-02	1	Solid	20.02	20	7.86	01/23/2018	17:50
29	J1281-08	1	Solid	20.03	20	7.63	01/23/2018	17:51
30	J1281-08DUP	1	Solid	20.05	20	7.64	01/23/2018	18:00
31	CCV4	1	Water			12.01	01/23/2018	18:01

23 92816

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1281

WorkList ID : 107768

Date : 1/23/2018 4:51:17 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2018	Solid	J1259-01	Corrosivity	Cool 4 deg C	EART12	M12	WC-01(0-5)	01/18/2018	9045C
01/18/2018	Solid	J1259-02	Corrosivity	Cool 4 deg C	EART12	M12	WC-02(0-5)	01/18/2018	9045C
01/18/2018	Solid	J1259-03	Corrosivity	Cool 4 deg C	EART12	M12	WC-03(0-5)	01/18/2018	9045C
01/18/2018	Solid	J1259-04	Corrosivity	Cool 4 deg C	EART12	M12	WC-04(0-5)	01/18/2018	9045C
01/18/2018	Solid	J1259-05	Corrosivity	Cool 4 deg C	EART12	M12	WC-05(0-5)	01/18/2018	9045C
01/18/2018	Solid	J1259-06	Corrosivity	Cool 4 deg C	EART12	M12	WC-06(5-16)	01/18/2018	9045C
01/18/2018	Solid	J1259-07	Corrosivity	Cool 4 deg C	EART12	M12	WC-07(5-15)	01/18/2018	9045C
01/18/2018	Solid	J1259-08	Corrosivity	Cool 4 deg C	EART12	M12	WC-08(5-15)	01/18/2018	9045C
01/18/2018	Solid	J1259-09	Corrosivity	Cool 4 deg C	EART12	M12	WC-9(5-15)	01/18/2018	9045C
01/19/2018	Solid	J1259-10	Corrosivity	Cool 4 deg C	EART12	M12	WC-10(0-5)	01/19/2018	9045C
01/19/2018	Solid	J1259-11	Corrosivity	Cool 4 deg C	EART12	M12	WC-11(0-5)	01/19/2018	9045C
01/19/2018	Solid	J1259-12	Corrosivity	Cool 4 deg C	EART12	M12	WC-12(0-5)	01/19/2018	9045C
01/19/2018	Solid	J1259-13	Corrosivity	Cool 4 deg C	EART12	M12	WC-13(0-5)	01/19/2018	9045C
01/19/2018	Solid	J1259-14	Corrosivity	Cool 4 deg C	EART12	M12	WC-14(0-5)	01/19/2018	9045C
01/19/2018	Solid	J1259-15	Corrosivity	Cool 4 deg C	EART12	M12	WC-15(0-5)	01/19/2018	9045C
01/19/2018	Solid	J1259-16	Corrosivity	Cool 4 deg C	EART12	M12	WC-16(0-5)	01/19/2018	9045C
01/19/2018	Solid	J1259-17	Corrosivity	Cool 4 deg C	EART12	M12	WC-17(0-5)	01/19/2018	9045C
01/19/2018	Solid	J1259-18	Corrosivity	Cool 4 deg C	EART12	M12	WC-19(5-15)	01/19/2018	9045C
01/19/2018	Solid	J1259-19	Corrosivity	Cool 4 deg C	EART12	M12	WC-20(5-15)	01/19/2018	9045C
01/20/2018	Solid	J1279-02	Corrosivity	Cool 4 deg C	WSPO01	M23	PB-1(25-32)	01/20/2018	9045C
01/23/2018	Solid	J1281-08	Corrosivity	Cool 4 deg C	PSEG01	M33	SP-A	01/23/2018	9045C

Date/Time 01-23-18 5:00 PM
 Received by: JB
 Relinquished by: [Signature]

Date/Time 01-23-18 4:00 PM
 Received by: [Signature]
 Relinquished by: JB