



Analytical Summary Report

Analysis Method: 120.1

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB92822

Reagent/Standard	Lot/Log #
conductivity standard	WP59787

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB92822BL		1	23.7	0.95	0.97	01/22/2018	10:15
2	LB92822BS	1413	1	22.9	1330.00	1385.58	01/22/2018	10:22
3	J1192-14		1	23.0	1051.00	1092.74	01/22/2018	10:32

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$