



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB92824

Reagent/Standard	Lot/Loq #
TURBIDITY ICV 10 NTU	WP61255
TURBIDITY CCV 10 NTU	WP61251
Turbidity std, 40NTU	WP61247
turbidity std 5 NTU	WP61250
turbidity std 20 NTU	WP61248
Turbidity std 1 NTU	WP61254
turbidity std 0 NTU	WP61246
turbidity std 10 NTU	WP61249
TURBIDITY LOD STD, 0.5NTU	WP61252
TURBIDITY LOQ STD, 2.0NTU	WP61253

Intercept: -0.0603

Slope: 0.9515

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.145	01/24/2018	09:30
2	CAL2	1	1	0.945	01/24/2018	09:36
3	CAL3	5	1	4.660	01/24/2018	09:44
4	CAL4	10	1	9.200	01/24/2018	09:50
5	CAL5	20	1	18.900	01/24/2018	09:56
6	CAL6	40	1	38.100	01/24/2018	10:00



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB92824

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.670	10.226	01/24/2018	10:06
2	ICB		1	0.172	0.244	01/24/2018	10:13
3	CCV1	10	1	9.890	10.457	01/24/2018	10:20
4	CCB1		1	0.180	0.253	01/24/2018	10:27
5	LB92824BL		1	0.176	0.248	01/24/2018	10:35
6	J1161-10		1	0.629	0.724	01/24/2018	10:43
7	J1161-11		1	1.930	2.092	01/24/2018	10:50
8	J1161-12		1	0.635	0.731	01/24/2018	10:57
9	J1190-02		1	5.700	6.054	01/24/2018	11:05
10	J1192-10		1	2.920	3.132	01/24/2018	11:15
11	CCV2	10	1	9.760	10.321	01/24/2018	11:23
12	CCB2		1	0.178	0.250	01/24/2018	11:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name : J1190 Wet PT

WorkList ID : 107547

Date : 1/18/2018 11:10:31 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2018	Water	J1190-01	Turbidity	Cool 4 deg C	CHEM02	QA Office	PT-TURB-WS	01/16/2018	180.1
01/18/2018	Water	J1190-02	Turbidity	Cool 4 deg C	CHEM02	QA Office	PT-TURB-WS	01/16/2018	SM2130 B

Date/Time 1/19/18 9:30 AM

Received by: AP

Relinquished by: MT

Date/Time 2/19/18 4:00 PM

Received by: MT

Relinquished by: AP