



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB92825

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP61251
Turbidity std, 40NTU	WP61247
turbidity std 5 NTU	WP61250
turbidity std 20 NTU	WP61248
Turbidity std 1 NTU	WP61254
turbidity std 0 NTU	WP61246
turbidity std 10 NTU	WP61249
TURBIDITY LOD STD, 0.5NTU	WP61252
TURBIDITY LOQ STD, 2.0NTU	WP61253

Intercept: -0.0603

Slope: 0.9515

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.145	01/24/2018	09:30
2	CAL2	1	1	0.945	01/24/2018	09:36
3	CAL3	5	1	4.660	01/24/2018	09:44
4	CAL4	10	1	9.200	01/24/2018	09:50
5	CAL5	20	1	18.900	01/24/2018	09:56
6	CAL6	40	1	38.100	01/24/2018	10:00



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB92825

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.670	10.226	01/24/2018	10:06
2	ICB		1	0.172	0.244	01/24/2018	10:13
3	CCV1	10	1	9.760	10.321	01/24/2018	11:23
4	CCB1		1	0.178	0.250	01/24/2018	11:30
5	LB92825BL		1	0.176	0.248	01/24/2018	11:34
6	J1161-07		1	0.630	0.725	01/24/2018	11:40
7	J1161-08		1	1.910	2.071	01/24/2018	11:45
8	J1161-09		1	0.605	0.699	01/24/2018	11:50
9	J1190-01		1	5.720	6.075	01/24/2018	11:56
10	J1192-09		1	2.950	3.164	01/24/2018	12:04
11	CCV2	10	1	9.550	10.100	01/24/2018	12:10
12	CCB2		1	0.171	0.243	01/24/2018	12:16