


7470A

LB92828
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35404



A= 0.0000e+000

B= 2.8021e-004

C= -1.4974e-002

Rho= 0.9996951

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.049	-0.049	-122	0.000	-122				
0.2	0.200	0.190	-0.010	732	0.0 %	732				
2.5	2.500	2.595	0.095	9315	0.0 %	9315				
5.0	5.000	4.911	-0.089	17581	0.0 %	17581				
7.5	7.500	7.647	0.147	27342	0.0 %	27342				
10.0	10.000	9.906	-0.094	35404	0.0 %	35404				

LB92828 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Jan 2018 14:30:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-122	-	0.0000	7470A	PPB	24 Jan 2018 14:31:58	S	Std
0.2 - 1	50.2	732	-	0.2000	7470A	PPB	24 Jan 2018 14:36:54	S	Std
2.5 - 1	52.5	9315	-	2.5000	7470A	PPB	24 Jan 2018 14:39:02	S	Std
5.0 - 1	55	17581	-	5.0000	7470A	PPB	24 Jan 2018 14:41:10	S	Std
7.5 - 1	57.5	27342	-	7.5000	7470A	PPB	24 Jan 2018 14:43:18	S	Std
10.0 - 1	57.0	35404	-	10.0000	7470A	PPB	24 Jan 2018 14:47:52	S	Std
ICV62 - 1	ICV62	13946	3.8929	-7470A	PPB	PPB	24 Jan 2018 14:50:20	U	SMPL
ICB62 - 1	ICB62	-253	-0.0859	-7470A	PPB	PPB	24 Jan 2018 14:52:26	U	SMPL
CCV44 - 1	CCV44	17888	4.9975	-7470A	PPB	PPB	24 Jan 2018 14:54:32	U	SMPL
CCB44 - 1	CCB44	-239	-0.0819	-7470A	PPB	PPB	24 Jan 2018 14:56:39	U	SMPL
CRA - 1	CRA	685	0.1770	-7470A	PPB	PPB	24 Jan 2018 14:58:46	U	SMPL
HighStd - 1	HighStd	35614	9.9646	-7470A	PPB	PPB	24 Jan 2018 15:00:52	U	SMPL
ChkStd - 1	ChkStd	25840	7.2257	-7470A	PPB	PPB	24 Jan 2018 15:02:59	U	SMPL
PB105921BL - 1	PBW	-154	-0.0581	-7470A	PPB	PPB	24 Jan 2018 15:05:08	U	SMPL
PB105921BS - 1	LCSW	14238	3.9747	-7470A	PPB	PPB	24 Jan 2018 15:07:16	U	SMPL
PB105921TB - 1	PB105871TB	-105	-0.0444	-7470A	PPB	PPB	24 Jan 2018 15:09:24	U	SMPL
J1015-13 - 1	CL-01-012218-A	-26	-0.0223	-7470A	PPB	PPB	24 Jan 2018 15:11:32	U	SMPL
J1015-13DUP - 1	CL-01-012218-ADUP	20	-0.0094	-7470A	PPB	PPB	24 Jan 2018 15:13:40	U	SMPL
J1015-13MS - 1	CL-01-012218-AMS	12940	3.6110	-7470A	PPB	PPB	24 Jan 2018 15:15:49	U	SMPL
J1015-13MSD - 1	CL-01-012218-AMSD	13416	3.7444	-7470A	PPB	PPB	24 Jan 2018 15:17:55	U	SMPL
CCV45 - 1	CCV45	17962	5.0182	-7470A	PPB	PPB	24 Jan 2018 15:20:02	U	SMPL
CCB45 - 1	CCB45	-241	-0.0825	-7470A	PPB	PPB	24 Jan 2018 15:22:09	U	SMPL
J1015-15 - 1	CL-01-012218-B	15	-0.0108	-7470A	PPB	PPB	24 Jan 2018 15:24:16	U	SMPL
J1015-17 - 1	CL-01-012218-C	150	0.0271	-7470A	PPB	PPB	24 Jan 2018 15:26:23	U	SMPL
J1015-19 - 1	CL-01-012218-D	42	-0.0032	-7470A	PPB	PPB	24 Jan 2018 15:28:30	U	SMPL
J1016-06 - 1	OK-01-012218	15	-0.0108	-7470A	PPB	PPB	24 Jan 2018 15:30:37	U	SMPL
J1022-09 - 1	SU-04-012218	-5	-0.0164	-7470A	PPB	PPB	24 Jan 2018 15:32:45	U	SMPL
J1254-01 - 1	OWS-1	-6	-0.0167	-7470A	PPB	PPB	24 Jan 2018 15:34:52	U	SMPL
J1254-02 - 1	OWS-2	3	-0.0141	-7470A	PPB	PPB	24 Jan 2018 15:37:00	U	SMPL
J1254-03 - 1	OWS-3	0	-0.0150	-7470A	PPB	PPB	24 Jan 2018 15:39:09	U	SMPL
J1254-04 - 1	OWS-4	-17	-0.0197	-7470A	PPB	PPB	24 Jan 2018 15:41:17	U	SMPL
J1254-05 - 1	OWS-5	-4	-0.0161	-7470A	PPB	PPB	24 Jan 2018 15:43:24	U	SMPL
CCV46 - 1	CCV46	17945	5.0135	-7470A	PPB	PPB	24 Jan 2018 15:45:31	U	SMPL
CCB46 - 1	CCB46	-242	-0.0828	-7470A	PPB	PPB	24 Jan 2018 15:47:37	U	SMPL
J1254-06 - 1	OWS-6	31	-0.0063	-7470A	PPB	PPB	24 Jan 2018 15:49:45	U	SMPL
J1255-01 - 1	OWS-1	-2	-0.0155	-7470A	PPB	PPB	24 Jan 2018 15:51:51	U	SMPL
J1255-02 - 1	OWS-2	31	-0.0063	-7470A	PPB	PPB	24 Jan 2018 15:53:58	U	SMPL
J1255-03 - 1	OWS-3	6	-0.0133	-7470A	PPB	PPB	24 Jan 2018 15:56:05	U	SMPL
J1255-04 - 1	OWS-4	-3	-0.0158	-7470A	PPB	PPB	24 Jan 2018 15:58:12	U	SMPL
J1258-02 - 1	CONCRETE-PIERS	0	-0.0150	-7470A	PPB	PPB	24 Jan 2018 16:00:19	U	SMPL
J1266-04 - 1	MH-198	-28	-0.0228	-7470A	PPB	PPB	24 Jan 2018 16:02:27	U	SMPL
J1266-08 - 1	MH-199	-27	-0.0225	-7470A	PPB	PPB	24 Jan 2018 16:04:35	U	SMPL
J1266-12 - 1	MH-200	-16	-0.0195	-7470A	PPB	PPB	24 Jan 2018 16:06:43	U	SMPL
CCV47 - 1	CCV47	19071	5.3290	-7470A	PPB	PPB	24 Jan 2018 16:08:50	U	SMPL
CCB47 - 1	CCB47	-262	-0.0884	-7470A	PPB	PPB	24 Jan 2018 16:10:58	U	SMPL
J1015-13LX5 - 1		59	0.0016	-7470A	PPB	PPB	24 Jan 2018 16:13:05	U	SMPL
J1015-13A - 1		13855	3.8674	-7470A	PPB	PPB	24 Jan 2018 16:15:12	U	SMPL
CCV48 - 1	CCV48	18758	5.2413	-7470A	PPB	PPB	24 Jan 2018 16:17:19	U	SMPL
CCB48 - 1	CCB48	-278	-0.0929	-7470A	PPB	PPB	24 Jan 2018 16:19:27	U	SMPL