


7471B

LB 92829
 INSTRUMENT ID: CV1

Linear

μ Abs.:

44656



A= 0.0000e+000

B= 2.2329e-004

C= -3.8398e-003

Rho= 0.9999449

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	48	0.000	48				
0.2	0.200	0.223	0.023	1015	0.0 %	1015	0			
2.5	2.500	2.435	-0.065	10924	0.0 %	10924				
5.0	5.000	5.013	0.013	22468	0.0 %	22468				
7.5	7.500	7.555	0.055	33852	0.0 %	33852				
10.0	10.000	9.967	-0.033	44656	0.0 %	44656				

LB92829 INSTRUMRNT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Jan 2018 16:34:38

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	48	-	0.0000	7471B	PPB	24 Jan 2018 16:39:34	S	Std
0.2 - 1	502	1015	-	0.2000	7471B	PPB	24 Jan 2018 16:41:51	S	Std
2.5 - 1	525	10924	-	2.5000	7471B	PPB	24 Jan 2018 16:44:09	S	Std
5.0 - 1	55	22468	-	5.0000	7471B	PPB	24 Jan 2018 16:46:27	S	Std
7.5 - 1	575	33852	-	7.5000	7471B	PPB	24 Jan 2018 16:48:44	S	Std
10.0 - 1	510	44656	-	10.0000	7471B	PPB	24 Jan 2018 16:56:23	S	Std
ICV63 - 1	ICV63	16830	3.7541		-7471B	PPB	24 Jan 2018 17:01:41	U	SMPL
ICB63 - 1	ICB63	-25	-0.0094		-7471B	PPB	24 Jan 2018 17:03:56	U	SMPL
CCV49 - 1	CCV49	21485	4.7935		-7471B	PPB	24 Jan 2018 17:06:12	U	SMPL
CCB49 - 1	CCB49	-57	-0.0166		-7471B	PPB	24 Jan 2018 17:08:28	U	SMPL
CRA - 1	CRA	965	0.2116		-7471B	PPB	24 Jan 2018 17:10:44	U	SMPL
HighStd - 1	HighStd	45296	10.1101		-7471B	PPB	24 Jan 2018 17:13:00	U	SMPL
ChkStd - 1	ChkStd	31378	7.0024		-7471B	PPB	24 Jan 2018 17:15:16	U	SMPL
PB105878BL - 1	PBS	-234	-0.0561		-7471B	PPB	24 Jan 2018 17:17:34	U	SMPL
PB105878BS - 1	LCSS	15679	3.4971		-7471B	PPB	24 Jan 2018 17:19:52	U	SMPL
J1249-01 - 1	AB-1(0-2)	462	0.0993		-7471B	PPB	24 Jan 2018 17:22:09	U	SMPL
J1249-01DUP - 1	AB-1(0-2)DUP	332	0.0703		-7471B	PPB	24 Jan 2018 17:24:26	U	SMPL
J1249-01MS - 1	AB-1(0-2)MS	16447	3.6685		-7471B	PPB	24 Jan 2018 17:26:43	U	SMPL
J1249-01MSD - 1	AB-1(0-2)MSD	16674	3.7192		-7471B	PPB	24 Jan 2018 17:29:01	U	SMPL
J1249-03 - 1	AB-2(0-2)	390	0.0832		-7471B	PPB	24 Jan 2018 17:31:17	U	SMPL
CCV50 - 1	CCV50	21543	4.8064		-7471B	PPB	24 Jan 2018 17:33:32	U	SMPL
CCB50 - 1	CCB50	-64	-0.0181		-7471B	PPB	24 Jan 2018 17:35:48	U	SMPL
J1249-05 - 1	AB-3(0-2)	487	0.1049		-7471B	PPB	24 Jan 2018 17:38:05	U	SMPL
J1249-07 - 1	AB-4(0-2)	415	0.0888		-7471B	PPB	24 Jan 2018 17:40:21	U	SMPL
J1249-09 - 1	AB-5(0-2)	334	0.0707		-7471B	PPB	24 Jan 2018 17:42:37	U	SMPL
J1249-11 - 1	AB-6(0-2)	253	0.0527		-7471B	PPB	24 Jan 2018 17:44:53	U	SMPL
J1249-13 - 1	AB-7(0-2)	552	0.1194		-7471B	PPB	24 Jan 2018 17:47:10	U	SMPL
J1249-15 - 1	AB-8(0-2)	484	0.1042		-7471B	PPB	24 Jan 2018 17:49:27	U	SMPL
J1249-17 - 1	AB-9(0-2)	701	0.1527		-7471B	PPB	24 Jan 2018 17:51:44	U	SMPL
J1249-19 - 1	AB-11(0-2)	517	0.1116		-7471B	PPB	24 Jan 2018 17:54:02	U	SMPL
J1250-01 - 1	AB-10(0-2)	318	0.0672		-7471B	PPB	24 Jan 2018 17:56:19	U	SMPL
J1250-03 - 1	AB-12(0-2)	756	0.1650		-7471B	PPB	24 Jan 2018 17:58:35	U	SMPL
CCV51 - 1	CCV51	22615	5.0458		-7471B	PPB	24 Jan 2018 18:00:52	U	SMPL
CCB51 - 1	CCB51	-55	-0.0161		-7471B	PPB	24 Jan 2018 18:03:08	U	SMPL
J1250-05 - 1	AB-13(0-2)	562	0.1216		-7471B	PPB	24 Jan 2018 18:05:25	U	SMPL
J1250-07 - 1	AB-14(0-2)	405	0.0866		-7471B	PPB	24 Jan 2018 18:07:41	U	SMPL
J1250-09 - 1	AB-15(0-2)	569	0.1232		-7471B	PPB	24 Jan 2018 18:09:57	U	SMPL
J1250-11 - 1	AB-16(0-2)	691	0.1505		-7471B	PPB	24 Jan 2018 18:12:13	U	SMPL
J1250-13 - 1	AB-17(0-2)	301	0.0634		-7471B	PPB	24 Jan 2018 18:14:30	U	SMPL
J1250-15 - 1	AB-18(0-2)	283	0.0594		-7471B	PPB	24 Jan 2018 18:16:47	U	SMPL
J1250-17 - 1	AB-19(0-2)	243	0.0504		-7471B	PPB	24 Jan 2018 18:19:05	U	SMPL
J1250-19 - 1	AB-20(0-2)	309	0.0652		-7471B	PPB	24 Jan 2018 18:21:22	U	SMPL
PB105879BL - 1	PBS	-5	-0.0050		-7471B	PPB	24 Jan 2018 18:23:40	U	SMPL
PB105879BS - 1	LCSS	16382	3.6540		-7471B	PPB	24 Jan 2018 18:25:56	U	SMPL
CCV52 - 1	CCV52	22414	5.0009		-7471B	PPB	24 Jan 2018 18:28:12	U	SMPL
CCB52 - 1	CCB52	-101	-0.0264		-7471B	PPB	24 Jan 2018 18:30:29	U	SMPL
J1015-12 - 1	CL-01-012218-A	1702	0.3762		-7471B	PPB	24 Jan 2018 18:32:45	U	SMPL
J1015-14 - 1	CL-01-012218-B	1465	0.3233		-7471B	PPB	24 Jan 2018 18:35:01	U	SMPL
J1015-16 - 1	CL-01-012218-C	4127	0.9177		-7471B	PPB	24 Jan 2018 18:37:17	U	SMPL
J1015-18 - 1	CL-01-012218-D	3748	0.8330		-7471B	PPB	24 Jan 2018 18:39:33	U	SMPL
J1016-05 - 1	OK-01-012218	786	0.1717		-7471B	PPB	24 Jan 2018 18:41:50	U	SMPL
J1022-08 - 1	SU-04-012218	1154	0.2538		-7471B	PPB	24 Jan 2018 18:44:07	U	SMPL
J1256-01 - 1	DB-OF-E-W-(SW0-4)	992	2.2272		-7471B	PPB	24 Jan 2018 18:46:24	U	SMPL
J1256-02 - 1	DB-OF-E-W-(SW4-8)	1561	0.3447		-7471B	PPB	24 Jan 2018 18:48:42	U	SMPL
J1256-03 - 1	DB-OF-E-W-(NW0-4)	11792	2.6291		-7471B	PPB	24 Jan 2018 18:51:00	U	SMPL
J1256-04 - 1	DB-OF-E-W-(NW4-8)	5803	1.2919		-7471B	PPB	24 Jan 2018 18:53:16	U	SMPL
CCV53 - 1	CCV53	22510	5.0223		-7471B	PPB	24 Jan 2018 18:55:32	U	SMPL
CCB53 - 1	CCB53	-102	-0.0266		-7471B	PPB	24 Jan 2018 18:57:49	U	SMPL
J1256-05 - 1	DB-OF-N-S-(WW0-4)	6368	1.4180		-7471B	PPB	24 Jan 2018 19:00:05	U	SMPL
J1256-06 - 1	DB-OF-N-S-(WW4-8)	6523	1.4527		-7471B	PPB	24 Jan 2018 19:02:21	U	SMPL
J1256-07 - 1	DB-OF-N-S-(EW0-4)	3522	0.7826		-7471B	PPB	24 Jan 2018 19:04:37	U	SMPL
J1256-08 - 1	DB-OF-N-S-(EW4-8)	2869	0.6368		-7471B	PPB	24 Jan 2018 19:06:54	U	SMPL
J1256-09 - 1	8-SAN-1-(WW1-4)	8680	1.9343		-7471B	PPB	24 Jan 2018 19:09:10	U	SMPL
J1256-10 - 1	8-SAN-1-(EW1-4)	3029	0.6725		-7471B	PPB	24 Jan 2018 19:11:27	U	SMPL
J1256-10DUP - 1	8-SAN-1-(EW1-4)DUP	2631	0.5836		-7471B	PPB	24 Jan 2018 19:13:44	U	SMPL
J1256-10MS - 1	8-SAN-1-(EW1-4)MS	19004	4.2395		-7471B	PPB	24 Jan 2018 19:16:01	U	SMPL
J1256-10MSD - 1	8-SAN-1-(EW1-4)MSD	19692	4.3931		-7471B	PPB	24 Jan 2018 19:18:19	U	SMPL
J1258-01 - 1	CONCRETE-PIERS	243	0.0504		-7471B	PPB	24 Jan 2018 19:20:36	U	SMPL
CCV54 - 1	CCV54	22332	4.9826		-7471B	PPB	24 Jan 2018 19:22:53	U	SMPL
CCB54 - 1	CCB54	-74	-0.0204		-7471B	PPB	24 Jan 2018 19:25:09	U	SMPL
J1249-01LX5 - 1		99	0.0183		-7471B	PPB	24 Jan 2018 19:27:25	U	SMPL
J1249-01A - 1		17615	3.9293		-7471B	PPB	24 Jan 2018 19:29:43	U	SMPL
J1256-10LX5 - 1		525	0.1134		-7471B	PPB	24 Jan 2018 19:31:59	U	SMPL
J1256-10A - 1		20063	4.4759		-7471B	PPB	24 Jan 2018 19:34:16	U	SMPL
CCV55 - 1	CCV55	22259	4.9663		-7471B	PPB	24 Jan 2018 19:36:33	U	SMPL
CCB55 - 1	CCB55	-112	-0.0288		-7471B	PPB	24 Jan 2018 19:38:49	U	SMPL