


7470A

LB 92849
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36273



A= 0.0000e+000

B= 2.7422e-004

C= 1.1056e-002

Rho= 0.9998653

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	46	0.000	46				
0.2	0.200	0.236	0.036	819	0.0 %	819				
2.5	2.500	2.400	-0.100	8713	0.0 %	8713				
5.0	5.000	4.991	-0.009	18160	0.0 %	18160				
7.5	7.500	7.592	0.092	27644	0.0 %	27644				
10.0	10.000	9.958	-0.042	36273	0.0 %	36273				

LB92849 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Jan 2018 08:44:26

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7470A	PPB	25 Jan 2018 09:04:53	S	Std
0.2 - 1	312	819	-	0.2000	7470A	PPB	25 Jan 2018 09:07:01	S	Std
2.5 - 1	52.5	8713	-	2.5000	7470A	PPB	25 Jan 2018 09:09:09	S	Std
5.0 - 1	55	18160	-	5.0000	7470A	PPB	25 Jan 2018 09:11:17	S	Std
7.5 - 1	57.5	27644	-	7.5000	7470A	PPB	25 Jan 2018 09:13:25	S	Std
10.0 - 1	511	36273	-	10.0000	7470A	PPB	25 Jan 2018 09:15:33	S	Std
ICV64 - 1	ICV64	13563	3.7303	-	7470A	PPB	25 Jan 2018 09:18:06	U	SMPL
ICB64 - 1	ICB64	-59	-0.0051	-	7470A	PPB	25 Jan 2018 09:20:12	U	SMPL
CCV56 - 1	CCV56	18678	5.1329	-	7470A	PPB	25 Jan 2018 09:22:19	U	SMPL
CCB56 - 1	CCB56	-127	-0.0238	-	7470A	PPB	25 Jan 2018 09:24:26	U	SMPL
CRA - 1	CRA	728	0.2107	-	7470A	PPB	25 Jan 2018 09:26:33	U	SMPL
HighStd - 1	HighStd	37079	10.1789	-	7470A	PPB	25 Jan 2018 09:28:39	U	SMPL
ChkStd - 1	ChkStd	25337	6.9590	-	7470A	PPB	25 Jan 2018 09:30:46	U	SMPL
PB105976BL - 1	PBW	-175	-0.0369	-	7470A	PPB	25 Jan 2018 09:32:55	U	SMPL
PB105976BS - 1	LCSW	12672	3.4860	-	7470A	PPB	25 Jan 2018 09:35:04	U	SMPL
PB105976TB - 1	PB105924TB	-76	-0.0098	-	7470A	PPB	25 Jan 2018 09:37:11	U	SMPL
J1259-01 - 1	WC-01(0-5)	-44	-0.0010	-	7470A	PPB	25 Jan 2018 09:39:20	U	SMPL
J1259-01DUP - 1	WC-01(0-5)DUP	-56	-0.0043	-	7470A	PPB	25 Jan 2018 09:41:28	U	SMPL
J1259-01MS - 1	WC-01(0-5)MS	13390	3.6829	-	7470A	PPB	25 Jan 2018 09:43:36	U	SMPL
J1259-01MSD - 1	WC-01(0-5)MSD	12539	3.4495	-	7470A	PPB	25 Jan 2018 09:45:42	U	SMPL
CCV57 - 1	CCV57	18585	5.1074	-	7470A	PPB	25 Jan 2018 09:47:49	U	SMPL
CCB57 - 1	CCB57	-82	-0.0114	-	7470A	PPB	25 Jan 2018 09:49:56	U	SMPL
J1259-02 - 1	WC-02(0-5)	-18	0.0061	-	7470A	PPB	25 Jan 2018 09:52:04	U	SMPL
J1259-03 - 1	WC-03(0-5)	38	0.0215	-	7470A	PPB	25 Jan 2018 09:54:10	U	SMPL
J1259-04 - 1	WC-04(0-5)	-27	0.0037	-	7470A	PPB	25 Jan 2018 09:56:17	U	SMPL
J1259-05 - 1	WC-05(0-5)	-2	0.0105	-	7470A	PPB	25 Jan 2018 09:58:24	U	SMPL
J1259-06 - 1	WC-06(5-16)	-52	-0.0032	-	7470A	PPB	25 Jan 2018 10:00:32	U	SMPL
J1259-07 - 1	WC-07(5-15)	18	0.0160	-	7470A	PPB	25 Jan 2018 10:02:39	U	SMPL
J1259-08 - 1	WC-08(5-15)	4	0.0122	-	7470A	PPB	25 Jan 2018 10:04:47	U	SMPL
J1259-09 - 1	WC-9(5-15)	14	0.0149	-	7470A	PPB	25 Jan 2018 10:06:56	U	SMPL
J1259-10 - 1	WC-10(0-5)	-16	0.0067	-	7470A	PPB	25 Jan 2018 10:09:04	U	SMPL
J1259-11 - 1	WC-11(0-5)	21	0.0168	-	7470A	PPB	25 Jan 2018 10:11:11	U	SMPL
CCV58 - 1	CCV58	18301	5.0296	-	7470A	PPB	25 Jan 2018 10:13:18	U	SMPL
CCB58 - 1	CCB58	-79	-0.0106	-	7470A	PPB	25 Jan 2018 10:15:25	U	SMPL
J1259-12 - 1	WC-12(0-5)	24	0.0176	-	7470A	PPB	25 Jan 2018 10:17:33	U	SMPL
J1259-13 - 1	WC-13(0-5)	-24	0.0045	-	7470A	PPB	25 Jan 2018 10:19:40	U	SMPL
J1259-14 - 1	WC-14(0-5)	-3	0.0102	-	7470A	PPB	25 Jan 2018 10:21:46	U	SMPL
J1259-15 - 1	WC-15(0-5)	-43	-0.0007	-	7470A	PPB	25 Jan 2018 10:23:54	U	SMPL
J1259-16 - 1	WC-16(0-5)	19	0.0163	-	7470A	PPB	25 Jan 2018 10:26:01	U	SMPL
J1259-17 - 1	WC-17(0-5)	-4	0.0100	-	7470A	PPB	25 Jan 2018 10:28:08	U	SMPL
J1259-18 - 1	WC-19(5-15)	4	0.0122	-	7470A	PPB	25 Jan 2018 10:30:16	U	SMPL
J1259-19 - 1	WC-20(5-15)	-22	0.0050	-	7470A	PPB	25 Jan 2018 10:32:25	U	SMPL
J1279-02 - 1	PB-1(25-32)	-44	-0.0010	-	7470A	PPB	25 Jan 2018 10:34:32	U	SMPL
J1281-08 - 1	SP-A	-37	0.0009	-	7470A	PPB	25 Jan 2018 10:36:39	U	SMPL
CCV59 - 1	CCV59	18473	5.0767	-	7470A	PPB	25 Jan 2018 10:38:46	U	SMPL
CCB59 - 1	CCB59	-67	-0.0073	-	7470A	PPB	25 Jan 2018 10:40:53	U	SMPL
PB105987BL - 1	PBW	-9	0.0086	-	7470A	PPB	25 Jan 2018 10:43:01	U	SMPL
PB105987BS - 1	LCSW	12918	3.5534	-	7470A	PPB	25 Jan 2018 10:45:08	U	SMPL
J1270-01 - 1	GROUNDWATER	-230	-0.0520	-	7470A	PPB	25 Jan 2018 10:47:15	U	SMPL
J1270-01DUP - 1	GROUNDWATERDUP	-129	-0.0243	-	7470A	PPB	25 Jan 2018 10:49:22	U	SMPL
J1270-01MS - 1	GROUNDWATERMS	15663	4.3062	-	7470A	PPB	25 Jan 2018 10:51:30	U	SMPL
J1270-01MSD - 1	GROUNDWATERMSD	15301	4.2069	-	7470A	PPB	25 Jan 2018 10:53:38	U	SMPL
J1272-10 - 1	FB	-271	-0.0633	-	7470A	PPB	25 Jan 2018 10:55:46	U	SMPL
J1300-06 - 1	MANHOLE	1454	0.4098	-	7470A	PPB	25 Jan 2018 10:57:54	U	SMPL
CCV60 - 1	CCV60	18406	5.0584	-	7470A	PPB	25 Jan 2018 11:00:03	U	SMPL
CCB60 - 1	CCB60	-75	-0.0095	-	7470A	PPB	25 Jan 2018 11:02:10	U	SMPL
J1259-01LX5 - 1		28	0.0187	-	7470A	PPB	25 Jan 2018 11:04:18	U	SMPL
J1259-01A - 1		13803	3.7961	-	7470A	PPB	25 Jan 2018 11:06:24	U	SMPL
J1270-01LX5 - 1		-109	-0.0188	-	7470A	PPB	25 Jan 2018 11:08:31	U	SMPL
J1270-01A - 1		15123	4.1581	-	7470A	PPB	25 Jan 2018 11:10:38	U	SMPL
CCV61 - 1	CCV61	18300	5.0293	-	7470A	PPB	25 Jan 2018 11:12:46	U	SMPL
CCB61 - 1	CCB61	-77	-0.0101	-	7470A	PPB	25 Jan 2018 11:14:54	U	SMPL