



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB92852
Analysis Date: 01/25/2018
BalanceID: WC SC-4
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 01/25/2018
Extraction IN Time: 09:00
Extraction OUT Time: 09:30

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB92852BL	LB92852BL	WATER	2	1000	100	2.8711	2.8711	3.02	2.8711	2.8711	0	0
2	LB92852BS	LB92852BS	WATER	2	1000	100	2.7687	2.7687	3.01	2.7876	2.7876	0.0189	18.9
3	LB92852BSD	LB92852BSD	WATER	2	1000	100	2.8108	2.8108	3.04	2.8296	2.8296	0.0188	18.8
4	J1300-06	MANHOLE	WATER	2	1000	100	2.989	2.989	3.03	3.0002	3.0002	0.0112	11.2



QC Batch# LB92852 Test: TPH Analysis Date: 01/25/2018

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2317
pH Paper 0-14	W2287
Sodium Sulfate	EP1806
1:1 HCL	WP59249
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD	10.00 ML	WP58005
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 70 Dessicator Time In : 10:20
1.0000 gram Balance: 1.0003 In Time: 09:45
Bal Check Time: 09:15

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 12:00
1.0000 gram Balance: 1.0005 Out Time: 11:20
Bal Check Time: 12:30

LB 92852

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J1300 WorkList ID : 107864 Date : 1/25/2018 8:45:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/28/2018	Water	J1300-06-12	TPH	Conc H2SO4 to	PSEG01	M52	MANHOLE	01/24/2018	1664A

Date/Time 01-25-17 8:50 AM
Received by: JP
Relinquished by: CD

Date/Time 01-25-17 10:30 AM
Received by: CD
Relinquished by: JP