

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB92855

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP60984
calibration std. phosphate 1 ppm	WP61269
calibration std. phosphate 0.5 ppm	WP61268
calibration std. phosphate 0.3 ppm	WP61267
calibration std. 0 ppm	WP61264
phosphate ICV std.	WP61271
phosphate CCV std.	WP61270
calibration std. phosphate 0.1 ppm	WP61266
calibration std. phosphate 0.05 ppm	WP61265
Combined reagent	WP61273
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036
PHOSPHATE LOD STD, 0.01PPM	WP61274
PHOSPATE LOQ STD, 0.1PPM	WP61275

Intercept: -0.0011Slope: 0.6502Regression: 0.999819

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/25/2018	11:00
2	CAL2	0.05	1	50	50	0.033	01/25/2018	11:01
3	CAL3	0.10	1	50	50	0.065	01/25/2018	11:01
4	CAL4	0.30	1	50	50	0.185	01/25/2018	11:02
5	CAL5	0.50	1	50	50	0.329	01/25/2018	11:02
6	CAL6	1.00	1	50	50	0.649	01/25/2018	11:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB92855

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.494	01/25/2018	11:03
2	ICB		1	50	50	0.000	0.002	01/25/2018	11:04
3	CCV1	0.50	1	50	50	0.327	0.505	01/25/2018	11:05
4	CCB1		1	50	50	0.000	0.002	01/25/2018	11:05
5	PB105999BL		1	50	50	0.000	0.002	01/25/2018	11:06
6	PB105999BS	0.50	1	50	50	0.315	0.486	01/25/2018	11:06
7	J1161-07		1	50	50	0.011	0.019	01/25/2018	11:07
8	J1161-08		1	50	50	0.039	0.062	01/25/2018	11:07
9	J1161-09		1	50	50	0.009	0.016	01/25/2018	11:08
10	J1199-01		1	50	50	1.512	2.327	01/25/2018	11:08
11	J1199-01DUP		1	50	50	1.510	2.324	01/25/2018	11:08
12	J1199-01MS	0.50	1	50	50	1.531	2.356	01/25/2018	11:09
13	J1199-01MSD	0.50	1	50	50	1.535	2.363	01/25/2018	11:10
14	J1199-01		10	50	50	0.512	0.789	01/25/2018	11:25
15	CCV2	0.50	1	50	50	0.323	0.498	01/25/2018	11:25
16	CCB2		1	50	50	0.000	0.002	01/25/2018	11:26
17	J1199-01DUP		10	50	50	0.502	0.774	01/25/2018	11:26
18	J1199-01MS	0.50	10	50	50	0.548	0.845	01/25/2018	11:27
19	J1199-01MSD	0.50	10	50	50	0.550	0.848	01/25/2018	11:27
20	CCV3	0.50	1	50	50	0.325	0.502	01/25/2018	11:28
21	CCB3		1	50	50	0.000	0.002	01/25/2018	11:28