


ISM02.4

LB 92860
INSTRUMENT 11: CV1

Linear

 μ Abs.:

47154



A= 0.0000e+000

B= 2.1412e-004

C= 1.0807e-002

Rho= 0.9997373

Accept=Accepted

slope
 $y = 1.0807e-002 x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Q/D
0.00	0.000	0.040	0.040	135	0.000	135					7
0.2	0.200	0.214	0.014	951	0.0 %	951					2
2.5	2.500	2.555	0.055	11884	0.0 %	11884					-3
5.0	5.000	4.852	-0.148	22612	0.0 %	22612					-1
7.5	7.500	7.431	-0.069	34654	0.0 %	34654					1
10.0	10.000	10.107	0.107	47154	0.0 %	47154					

NS
1/25/18

LB92860
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	135	-		0 ISM02.4	PPB	1/25/2018 12:22	Std
	0.2 S0.2	951	-		0.2 ISM02.4	PPB	1/25/2018 12:24	Std
	2.5 S2.5	11884	-		2.5 ISM02.4	PPB	1/25/2018 12:26	Std
	5 S5	22612	-		5 ISM02.4	PPB	1/25/2018 12:29	Std
	7.5 S7.5	34654	-		7.5 ISM02.4	PPB	1/25/2018 12:31	Std
	10 S10	47154	-		10 ISM02.4	PPB	1/25/2018 12:33	Std
ICV	ICV	17776	3.8169 -		ISM02.4	PPB	1/25/2018 12:36	SMPL
ICB	ICB	-59	-0.0018 -		ISM02.4	PPB	1/25/2018 12:38	SMPL
CCV062	CCV062	22782	4.8888 -		ISM02.4	PPB	1/25/2018 12:41	SMPL
CCB062	CCB062	2	0.0112 -		ISM02.4	PPB	1/25/2018 12:43	SMPL
PB105984BL	PBS004	116	0.0356 -		ISM02.4	PPB	1/25/2018 12:45	SMPL
J1242-01	MF6B19	964	0.2172 -		ISM02.4	PPB	1/25/2018 12:47	SMPL
J1242-02	MF6B20	380	0.0922 -		ISM02.4	PPB	1/25/2018 12:50	SMPL
J1242-03	MF6B21	718	0.1645 -		ISM02.4	PPB	1/25/2018 12:52	SMPL
J1242-04	MF6B27	266	0.0678 -		ISM02.4	PPB	1/25/2018 12:54	SMPL
J1242-05	MF6B28	613	0.1421 -		ISM02.4	PPB	1/25/2018 12:56	SMPL
J1242-06	MF6B29	775	0.1767 -		ISM02.4	PPB	1/25/2018 12:59	SMPL
J1242-07	MF6B30	537	0.1258 -		ISM02.4	PPB	1/25/2018 13:01	SMPL
J1242-08	MF6B31	492	0.1162 -		ISM02.4	PPB	1/25/2018 13:03	SMPL
J1242-09	MF6B32	438	0.1046 -		ISM02.4	PPB	1/25/2018 13:06	SMPL
J1242-10	MF6B33	506	0.1192 -		ISM02.4	PPB	1/25/2018 13:08	SMPL
J1242-11	MF6B34	228	0.0596 -		ISM02.4	PPB	1/25/2018 13:10	SMPL
J1242-12	MF6B35	1414	0.3136 -		ISM02.4	PPB	1/25/2018 13:12	SMPL
J1242-13	MF6B36	279	0.0705 -		ISM02.4	PPB	1/25/2018 13:15	SMPL
J1242-14	MF6B36D	287	0.0723 -		ISM02.4	PPB	1/25/2018 13:17	SMPL
J1242-15	MF6B36S	12198	2.6226 -		ISM02.4	PPB	1/25/2018 13:19	SMPL
J1242-16	MF6B47	887	0.2007 -		ISM02.4	PPB	1/25/2018 13:21	SMPL
J1242-17	MF6B48	309	0.077 -		ISM02.4	PPB	1/25/2018 13:24	SMPL
J1242-18	MF6B49	367	0.0894 -		ISM02.4	PPB	1/25/2018 13:26	SMPL
J1242-19	MF6B50	260	0.0665 -		ISM02.4	PPB	1/25/2018 13:28	SMPL
CCV063	CCV063	22855	4.9044 -		ISM02.4	PPB	1/25/2018 13:31	SMPL
CCB063	CCB063	-3	0.0102 -		ISM02.4	PPB	1/25/2018 13:33	SMPL
J1242-20	MF6B51	374	0.0909 -		ISM02.4	PPB	1/25/2018 13:35	SMPL
PB105985BL	PBS004	106	0.0335 -		ISM02.4	PPB	1/25/2018 13:37	SMPL
J1277-01	MF6B57	585	0.1361 -		ISM02.4	PPB	1/25/2018 13:40	SMPL
J1277-02	MF6B58	471	0.1117 -		ISM02.4	PPB	1/25/2018 13:42	SMPL
J1277-03	MF6B59	297	0.0744 -		ISM02.4	PPB	1/25/2018 13:44	SMPL
J1277-04	MF6B60	328	0.081 -		ISM02.4	PPB	1/25/2018 13:47	SMPL
J1277-05	MF6B61	981	0.2209 -		ISM02.4	PPB	1/25/2018 13:49	SMPL
J1277-06	MF6B62	275	0.0697 -		ISM02.4	PPB	1/25/2018 13:51	SMPL
J1277-07	MF6B63	492	0.1162 -		ISM02.4	PPB	1/25/2018 13:53	SMPL
J1277-08	MF6B64	284	0.0716 -		ISM02.4	PPB	1/25/2018 13:56	SMPL
J1277-09	MF6B65	412	0.099 -		ISM02.4	PPB	1/25/2018 13:58	SMPL
J1277-10	MF6B66	549	0.1284 -		ISM02.4	PPB	1/25/2018 14:00	SMPL
J1277-11	MF6B67	366	0.0892 -		ISM02.4	PPB	1/25/2018 14:03	SMPL
J1277-12	MF6B68	350	0.0857 -		ISM02.4	PPB	1/25/2018 14:05	SMPL

INSTRUMENT ID : CV1

J1277-13	MF6B69	273	0.0693 -	ISM02.4	PPB	1/25/2018 14:07	SMPL
J1277-14	MF6B70	223	0.0586 -	ISM02.4	PPB	1/25/2018 14:09	SMPL
J1277-15	MF6B71	348	0.0853 -	ISM02.4	PPB	1/25/2018 14:12	SMPL
J1277-16	MF6B72	472	0.1119 -	ISM02.4	PPB	1/25/2018 14:14	SMPL
J1277-17	MF6B73	1196	0.2669 -	ISM02.4	PPB	1/25/2018 14:16	SMPL
J1277-18	MF6B74	302	0.0755 -	ISM02.4	PPB	1/25/2018 14:18	SMPL
CCV064	CCV064	22810	4.8948 -	ISM02.4	PPB	1/25/2018 14:21	SMPL
CCB064	CCB064	19	0.0149 -	ISM02.4	PPB	1/25/2018 14:23	SMPL
J1277-19	MF6B75	332	0.0819 -	ISM02.4	PPB	1/25/2018 14:25	SMPL
J1277-20	MF6B75D	376	0.0913 -	ISM02.4	PPB	1/25/2018 14:28	SMPL
J1277-21	MF6B75S	12301	2.6447 -	ISM02.4	PPB	1/25/2018 14:30	SMPL
J1277-22	MF6B76	137	0.0401 -	ISM02.4	PPB	1/25/2018 14:32	SMPL
PB105986BL	PBS004	61	0.0239 -	ISM02.4	PPB	1/25/2018 14:34	SMPL
J1278-01	MF6B77	264	0.0673 -	ISM02.4	PPB	1/25/2018 14:37	SMPL
J1278-02	MF6B78	750	0.1714 -	ISM02.4	PPB	1/25/2018 14:39	SMPL
J1278-03	MF6B79	644	0.1487 -	ISM02.4	PPB	1/25/2018 14:41	SMPL
J1278-04	MF6B80	181	0.0496 -	ISM02.4	PPB	1/25/2018 14:43	SMPL
J1278-05	MF6B81	334	0.0823 -	ISM02.4	PPB	1/25/2018 14:46	SMPL
J1278-06	MF6B82	217	0.0573 -	ISM02.4	PPB	1/25/2018 14:48	SMPL
J1278-07	MF6B83	294	0.0738 -	ISM02.4	PPB	1/25/2018 14:50	SMPL
J1278-08	MF6B84	402	0.0969 -	ISM02.4	PPB	1/25/2018 14:53	SMPL
J1278-09	MF6B85	229	0.0598 -	ISM02.4	PPB	1/25/2018 14:55	SMPL
J1278-10	MF6B86	267	0.068 -	ISM02.4	PPB	1/25/2018 14:57	SMPL
J1278-11	MF6B87	444	0.1059 -	ISM02.4	PPB	1/25/2018 14:59	SMPL
J1278-12	MF6B88	374	0.0909 -	ISM02.4	PPB	1/25/2018 15:02	SMPL
J1278-13	MF6B89	595	0.1382 -	ISM02.4	PPB	1/25/2018 15:04	SMPL
J1278-14	MF6B90	563	0.1314 -	ISM02.4	PPB	1/25/2018 15:06	SMPL
J1278-15	MF6B91	957	0.2157 -	ISM02.4	PPB	1/25/2018 15:09	SMPL
CCV065	CCV065	22922	4.9188 -	ISM02.4	PPB	1/25/2018 15:11	SMPL
CCB065	CCB065	-9	0.0089 -	ISM02.4	PPB	1/25/2018 15:13	SMPL
J1278-16	MF6B92	242	0.0626 -	ISM02.4	PPB	1/25/2018 15:15	SMPL
J1278-17	MF6B93	278	0.0703 -	ISM02.4	PPB	1/25/2018 15:18	SMPL
J1278-18	MF6C34	250	0.0643 -	ISM02.4	PPB	1/25/2018 15:20	SMPL
J1278-19	MF6C35	270	0.0686 -	ISM02.4	PPB	1/25/2018 15:22	SMPL
J1278-20	MF6C36	431	0.1031 -	ISM02.4	PPB	1/25/2018 15:25	SMPL
J1278-21	MF6C36D	409	0.0984 -	ISM02.4	PPB	1/25/2018 15:27	SMPL
J1278-22	MF6C36S	12249	2.6335 -	ISM02.4	PPB	1/25/2018 15:29	SMPL
CCV066	CCV066	22776	4.8875 -	ISM02.4	PPB	1/25/2018 15:31	SMPL
CCB066	CCB066	-42	0.0018 -	ISM02.4	PPB	1/25/2018 15:34	SMPL