

7471B


LB 92869

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41873



A= 0.0000e+000

B= 2.3556e-004

C= -2.5309e-002

Rho= 0.9993909

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	96	0.000	96				
0.2	0.200	0.219	0.019	1037	0.0 %	1037	0			
2.5	2.500	2.444	-0.056	10482	0.0 %	10482				
5.0	5.000	4.943	-0.057	21092	0.0 %	21092				
7.5	7.500	7.758	0.258	33042	0.0 %	33042				
10.0	10.000	9.838	-0.162	41873	0.0 %	41873				

LB92863 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Jan 2018 17:14:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	96	-	0.00007471B	PPB		25 Jan 2018 17:18:47	S	Std
0.2 - 1	50.2	1037	-	0.20007471B	PPB		25 Jan 2018 17:21:04	S	Std
2.5 - 1	52.5	10482	-	2.50007471B	PPB		25 Jan 2018 17:23:22	S	Std
5.0 - 1	55	21092	-	5.00007471B	PPB		25 Jan 2018 17:25:39	S	Std
7.5 - 1	57.5	33042	-	7.50007471B	PPB		25 Jan 2018 17:27:57	S	Std
10.0 - 1	57.5	41873	-	10.00007471B	PPB		25 Jan 2018 17:32:52	S	Std
ICV65 - 1	ICV65	17869	4.1840	-7471B	PPB		25 Jan 2018 17:38:51	U	SMPL
ICB65 - 1	ICB65	-88	-0.0460	-7471B	PPB		25 Jan 2018 17:41:07	U	SMPL
CCV67 - 1	CCV67	21832	5.1175	-7471B	PPB		25 Jan 2018 17:43:22	U	SMPL
CCB67 - 1	CCB67	-107	-0.0505	-7471B	PPB		25 Jan 2018 17:45:38	U	SMPL
CRA - 1	CRA	984	0.2065	-7471B	PPB		25 Jan 2018 17:47:55	U	SMPL
HighStd - 1	HighStd	42096	9.8910	-7471B	PPB		25 Jan 2018 17:50:10	U	SMPL
ChkStd - 1	ChkStd	30493	7.1577	-7471B	PPB		25 Jan 2018 17:52:26	U	SMPL
PB105988BL - 1	PBS	-160	-0.0630	-7471B	PPB		25 Jan 2018 17:54:45	U	SMPL
PB105988BS - 1	LCSS	15597	3.6488	-7471B	PPB		25 Jan 2018 17:57:02	U	SMPL
J1020-01 - 1	PL-01-012418-A	1351	0.2929	-7471B	PPB		25 Jan 2018 17:59:19	U	SMPL
J1020-01DUP - 1	PL-01-012418-ADUP	1181	0.2529	-7471B	PPB		25 Jan 2018 18:01:36	U	SMPL
J1020-01MS - 1	PL-01-012418-AMS	17506	4.0985	-7471B	PPB		25 Jan 2018 18:03:53	U	SMPL
J1020-01MSD - 1	PL-01-012418-AMSD	17795	4.1665	-7471B	PPB		25 Jan 2018 18:06:11	U	SMPL
J1020-03 - 1	PL-01-012418-B	1297	0.2802	-7471B	PPB		25 Jan 2018 18:08:27	U	SMPL
CCV68 - 1	CCV68	22023	5.1625	-7471B	PPB		25 Jan 2018 18:10:43	U	SMPL
CCB68 - 1	CCB68	-98	-0.0484	-7471B	PPB		25 Jan 2018 18:12:59	U	SMPL
J1020-05 - 1	PL-01-012418-C	914	0.1900	-7471B	PPB		25 Jan 2018 18:15:16	U	SMPL
J1020-07 - 1	PL-01-012418-D	2071	0.4625	-7471B	PPB		25 Jan 2018 18:17:32	U	SMPL
J1020-09 - 1	PL-02-012418-A	1424	0.3101	-7471B	PPB		25 Jan 2018 18:19:49	U	SMPL
J1020-11 - 1	PL-02-012418-B	2520	0.5683	-7471B	PPB		25 Jan 2018 18:22:05	U	SMPL
J1020-13 - 1	PL-02-012418-C	2287	0.5134	-7471B	PPB		25 Jan 2018 18:24:22	U	SMPL
J1020-15 - 1	PL-02-012418-D	2129	0.4762	-7471B	PPB		25 Jan 2018 18:26:39	U	SMPL
J1289-01 - 1	SB-08-9.5-10.0	1368	0.2969	-7471B	PPB		25 Jan 2018 18:28:56	U	SMPL
J1289-03 - 1	SB-09-9.5-10.0	187	0.0187	-7471B	PPB		25 Jan 2018 18:31:13	U	SMPL
J1289-05 - 1	SB-10-9.5-10.0	1011	0.2128	-7471B	PPB		25 Jan 2018 18:33:31	U	SMPL
J1289-07 - 1	SB-11-9.5-10.0	12772	2.9833	-7471B	PPB		25 Jan 2018 18:35:47	U	SMPL
CCV69 - 1	CCV69	22182	5.2000	-7471B	PPB		25 Jan 2018 18:38:04	U	SMPL
CCB69 - 1	CCB69	-103	-0.0496	-7471B	PPB		25 Jan 2018 18:40:19	U	SMPL
J1289-09 - 1	SB-60-9.5-10.0	1478	0.3229	-7471B	PPB		25 Jan 2018 18:42:35	U	SMPL
J1289-11 - 1	SB-56-13.5-14.0	1032	0.2178	-7471B	PPB		25 Jan 2018 18:44:52	U	SMPL
J1289-13 - 1	SB-53-13.5-14.0	914	0.1900	-7471B	PPB		25 Jan 2018 18:47:08	U	SMPL
J1289-15 - 1	SB-52-13.5-14.0	22399	5.2511	-7471B	PPB		25 Jan 2018 18:49:25	U	SMPL
J1289-17 - 1	SB-54-13.5-14.0	21662	5.0775	-7471B	PPB		25 Jan 2018 18:51:42	U	SMPL
J1289-19 - 1	SB-55-8.5-9.0	2828	0.6409	-7471B	PPB		25 Jan 2018 18:53:59	U	SMPL
J1300-02 - 1	TP-1-5	1168	0.2498	-7471B	PPB		25 Jan 2018 18:56:17	U	SMPL
PB105989BL - 1	PBS	383	0.0649	-7471B	PPB		25 Jan 2018 18:58:35	U	SMPL
PB105989BS - 1	LCSS	15152	3.5440	-7471B	PPB		25 Jan 2018 19:00:52	U	SMPL
J1290-01 - 1	SB-51-10.5-11.0	8633	2.0083	-7471B	PPB		25 Jan 2018 19:03:08	U	SMPL
CCV70 - 1	CCV70	22061	5.1715	-7471B	PPB		25 Jan 2018 19:05:25	U	SMPL
CCB70 - 1	CCB70	-106	-0.0503	-7471B	PPB		25 Jan 2018 19:07:41	U	SMPL
J1290-01DUP - 1	SB-51-10.5-11.0DUP	8680	2.0194	-7471B	PPB		25 Jan 2018 19:13:22	U	SMPL
J1290-01MS - 1	SB-51-10.5-11.0MS	24876	5.8346	-7471B	PPB		25 Jan 2018 19:19:21	U	SMPL
J1290-01MSD - 1	SB-51-10.5-11.0MSD	24510	5.7484	-7471B	PPB		25 Jan 2018 19:21:36	U	SMPL
J1290-03 - 1	SB-50-9.5-10.0	6475	1.5000	-7471B	PPB		25 Jan 2018 19:23:54	U	SMPL
J1290-05 - 1	SB-48-9.5-10.0	19611	4.5943	-7471B	PPB		25 Jan 2018 19:26:12	U	SMPL
J1290-07 - 1	SB-58-9.5-10.0	2765	0.6260	-7471B	PPB		25 Jan 2018 19:28:29	U	SMPL
J1290-09 - 1	SB-59-9.5-10.0	1076	0.2282	-7471B	PPB		25 Jan 2018 19:30:46	U	SMPL
J1290-11 - 1	SB-48-0-2.0	7028	1.6302	-7471B	PPB		25 Jan 2018 19:33:04	U	SMPL
J1290-12 - 1	SB-50-0-2.0	5924	1.3702	-7471B	PPB		25 Jan 2018 19:35:21	U	SMPL
J1290-13 - 1	SB-46-0-2.0	8185	1.9028	-7471B	PPB		25 Jan 2018 19:37:38	U	SMPL
CCV71 - 1	CCV71	20902	4.8984	-7471B	PPB		25 Jan 2018 19:39:54	U	SMPL
CCB71 - 1	CCB71	-103	-0.0496	-7471B	PPB		25 Jan 2018 19:42:10	U	SMPL
J1290-14 - 1	SB-45-0-2.0	7707	1.7902	-7471B	PPB		25 Jan 2018 19:44:27	U	SMPL
J1290-15 - 1	SB-43-0-2.0	2049	0.4574	-7471B	PPB		25 Jan 2018 19:46:43	U	SMPL
J1290-16 - 1	SB-49-9.5-10.0	5944	1.3749	-7471B	PPB		25 Jan 2018 19:48:59	U	SMPL
J1290-18 - 1	SB-46-9.5-10.0	17645	4.1312	-7471B	PPB		25 Jan 2018 19:51:15	U	SMPL
J1290-20 - 1	SB-45-9.5-10.0	11464	2.6752	-7471B	PPB		25 Jan 2018 19:53:32	U	SMPL
J1290-22 - 1	SB-43-6.5-7.0	717	0.1436	-7471B	PPB		25 Jan 2018 19:55:49	U	SMPL
J1299-01 - 1	IM-CF-HA-01	263	0.0366	-7471B	PPB		25 Jan 2018 19:58:07	U	SMPL
I1306-01 - 1	WC1	7409	1.7200	-7471B	PPB		25 Jan 2018 20:00:24	U	SMPL
I1306-05 - 1	WC2	29619	6.9519	-7471B	PPB		25 Jan 2018 20:02:42	U	SMPL
I1306-09 - 1	WC3	15898	3.7197	-7471B	PPB		25 Jan 2018 20:04:59	U	SMPL
CCV72 - 1	CCV72	20697	4.8502	-7471B	PPB		25 Jan 2018 20:07:16	U	SMPL
CCB72 - 1	CCB72	-105	-0.0500	-7471B	PPB		25 Jan 2018 20:09:33	U	SMPL
I1306-13 - 1	WC4	1935	0.4305	-7471B	PPB		25 Jan 2018 20:11:50	U	SMPL
I1306-17 - 1	WC5	10117	2.3579	-7471B	PPB		25 Jan 2018 20:14:07	U	SMPL
I1020-01LX5 - 1		215	0.0253	-7471B	PPB		25 Jan 2018 20:16:23	U	SMPL
I1020-01A - 1		18447	4.3201	-7471B	PPB		25 Jan 2018 20:18:39	U	SMPL
I1290-01LX5 - 1		1616	0.3554	-7471B	PPB		25 Jan 2018 20:20:56	U	SMPL
I1290-01A - 1		25477	5.9762	-7471B	PPB		25 Jan 2018 20:23:14	U	SMPL
CCV73 - 1	CCV73	21865	5.1253	-7471B	PPB		25 Jan 2018 20:25:31	U	SMPL
CCB73 - 1	CCB73	-131	-0.0562	-7471B	PPB		25 Jan 2018 20:27:49	U	SMPL