



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB92865

Temp.: 19.6 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
IODINE SOLUTION .025N 1L	W2261
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/24/2018	18:10
2	CAL2	1	Water			7.01	01/24/2018	18:11
3	CAL3	1	Water			10.03	01/24/2018	18:12
4	ICV	1	Water			6.99	01/24/2018	18:15
5	CCV1	1	Water			2.01	01/24/2018	18:15
6	J1270-01	1	Water			7.04	01/24/2018	18:18
7	J1270-01DUP	1	Water			7.05	01/24/2018	18:19
8	J1300-07	1	Water			7.38	01/24/2018	18:20
9	CCV2	1	Water			12.01	01/24/2018	18:30

WORKLIST(Hardcopy Internal Chain)

LB 02865

WorkList Name : CORROSIONITY J1300 WorkList ID : 107836 Date : 1/24/2018 5:52:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/22/2018	Water	J1270-01-17	Corrosivity	Cool 4 deg C	SEAC01	L63	GROUNDWATER	01/22/2018	9040C
01/24/2018	Water	J1300-07	Corrosivity	Cool 4 deg C	PSEG01	M52	MANHOLE	01/24/2018	9040C

Date/Time 12-04-18 6:00 PM
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 12-04-18 8:00 PM
 Received by: [Signature]
 Relinquished by: [Signature]