



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92870

Temp.: 18.4 °C

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.0 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/25/2018	13:35
2	CAL2	1	Water			6.99	01/25/2018	13:37
3	CAL3	1	Water			10.01	01/25/2018	13:39
4	ICV	1	Water			7.00	01/25/2018	13:41
5	CCV1	1	Water			2.00	01/25/2018	13:43
6	J1327-01	1	Water			6.19	01/25/2018	13:45
7	J1327-01DUP	1	Water			6.20	01/25/2018	13:47
8	CCV2	1	Water			12.0	01/25/2018	13:50

WORKLIST(Hardcopy Internal Chain)

LB 92870

WorkList Name : pH J1327

WorkList ID : 107897

Date : 1/25/2018 1:20:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/23/2018	Water	J1327-01	pH	Cool 4 deg C	STER06	N42	STERIGENICS-SALEM-CELL	01/23/2018	SM 4500-PH B

Date/Time 01/25/18 1:25 PM

Received by: SP

Relinquished by: CP

Date/Time 01/25/18 2:10 PM

Received by: CP

Relinquished by: SP